University of Groningen

Applications of various range shifters for proton pencil beam scanning radiotherapy

Lin, Haibo; Shi, Chengyu; Huang, Sheng; Shen, Jiajian; Kang, Minglei; Chen, Qing; Zhai, Huifang; McDonough, James; Tochner, Zelig; Deville, Curtiland

Published in:
Radiation oncology

DOI:
[10.1186/s13014-021-01873-8](https://doi.org/10.1186/s13014-021-01873-8)

IMPORTANT NOTE: You are advised to consult the publisher's version (publisher's PDF) if you wish to cite from it. Please check the document version below.

Document Version
Publisher's PDF, also known as Version of record

Publication date:
2021

[Link to publication in University of Groningen/UMCG research database](#)

Citation for published version (APA):

Lin, H., Shi, C., Huang, S., Shen, J., Kang, M., Chen, Q., Zhai, H., McDonough, J., Tochner, Z., Deville, C., Simone, C. B., & Both, S. (2021). Applications of various range shifters for proton pencil beam scanning radiotherapy. *Radiation oncology*, 16(1), [146]. <https://doi.org/10.1186/s13014-021-01873-8>

Copyright

Other than for strictly personal use, it is not permitted to download or to forward/distribute the text or part of it without the consent of the author(s) and/or copyright holder(s), unless the work is under an open content license (like Creative Commons).

The publication may also be distributed here under the terms of Article 25fa of the Dutch Copyright Act, indicated by the "Taverne" license. More information can be found on the University of Groningen website: <https://www.rug.nl/library/open-access/self-archiving-pure/taverne-amendment>.

Take-down policy

If you believe that this document breaches copyright please contact us providing details, and we will remove access to the work immediately and investigate your claim.

Downloaded from the University of Groningen/UMCG research database (Pure): <http://www.rug.nl/research/portal>. For technical reasons the number of authors shown on this cover page is limited to 10 maximum.

RESEARCH

Open Access

Applications of various range shifters for proton pencil beam scanning radiotherapy

Haibo Lin^{1*} , Chengyu Shi¹, Sheng Huang¹, Jiajian Shen², Minglei Kang¹, Qing Chen¹, Huifang Zhai¹, James McDonough³, Zelig Tochner³, Curtiland Deville⁴, Charles B. Simone II¹ and Stefan Both⁵

Abstract

Background: A range pull-back device, such as a machine-related range shifter (MRS) or a universal patient-related range shifter (UPRS), is needed in pencil beam scanning technique to treat shallow tumors.

Methods: Three UPRS made by QFix (Avondale, PA, USA) allow treating targets across the body: U-shaped bolus (UB), anterior lateral bolus (ALB), and couch top bolus. Head-and-neck (HN) patients who used the UPRS were tested. The in-air spot sizes were measured and compared in this study at air gaps: 6 cm, 16 cm, and 26 cm. Measurements were performed in a solid water phantom using a single-field optimization pencil beam scanning field with the ALB placed at 0, 10, and 20 cm air gaps. The two-dimensional dose maps at the middle of the spread-out Bragg peak were measured using ion chamber array MatriXX PT (IBA-Dosimetry, Schwarzenbruck, Germany) located at isocenter and compared with the treatment planning system.

Results: A UPRS can be consistently placed close to the patient and maintains a relatively small spot size resulting in improved dose distributions. However, when a UPRS is non-removable (e.g. thick couch top), the quality of volumetric imaging is degraded due to their high Z material construction, hindering the value of Image-Guided Radiation Therapy (IGRT). Limitations of using UPRS with small air gaps include reduced couch weight limit, potential collision with patient or immobilization devices, and challenges using non-coplanar fields with certain UPRS. Our experience showed the combination of a U-shaped bolus exclusively for an HN target and an MRS as the complimentary device for head-and-neck targets as well as for all other treatment sites may be ideal to preserve the dosimetric advantages of pencil beam scanning proton treatments across the body.

Conclusion: We have described how to implement UPRS and MRS for various clinical indications using the PBS technique, and comprehensively reviewed the advantage and disadvantages of UPRS and MRS. We recommend the removable UB only to be employed for the brain and HN treatments while an automated MRS is used for all proton beams that require RS but not convenient or feasible to use UB.

Keywords: Proton therapy, Pencil beam scanning (PBS), Range shifter, Bolus

Background

Proton therapy has developed rapidly in recent years and the pencil beam scanning (PBS) delivery technique has been introduced to many new proton therapy centers [1]. It is well known that the PBS technique can provide more

conformal dose distribution to the target volume than the traditional proton double scattering and uniform scanning techniques [2]. The potential of the proton PBS technique to generate highly conformal dose distributions relies heavily on the ability to maintain small beam spot sizes [3, 4]. Many factors can affect the spot sizes such as the proton energy, the treatment depth, and the air gap between a beam modifier and the patient [5, 6].

*Correspondence: hlin@nyproton.com

¹ New York Proton Center, New York 10035, USA

Full list of author information is available at the end of the article

© The Author(s) 2021. **Open Access** This article is licensed under a Creative Commons Attribution 4.0 International License, which permits use, sharing, adaptation, distribution and reproduction in any medium or format, as long as you give appropriate credit to the original author(s) and the source, provide a link to the Creative Commons licence, and indicate if changes were made. The images or other third party material in this article are included in the article's Creative Commons licence, unless indicated otherwise in a credit line to the material. If material is not included in the article's Creative Commons licence and your intended use is not permitted by statutory regulation or exceeds the permitted use, you will need to obtain permission directly from the copyright holder. To view a copy of this licence, visit <http://creativecommons.org/licenses/by/4.0/>. The Creative Commons Public Domain Dedication waiver (<http://creativecommons.org/publicdomain/zero/1.0/>) applies to the data made available in this article, unless otherwise stated in a credit line to the data.

Due to the extraction and transport efficiency, all current PBS systems have limitations in their minimum deliverable energy. For example, the IBA system (Ion Beam Applications, Louvain-La-Neuve, Belgium) provides minimum proton energy of 60–100 MeV [3]. The traditional way to treat targets shallower than the minimum range of protons is to use machine-mounted range shifters (MRS). The range shifters (RS) are a slab of uniform material, usually made of plastic, to pull back the range of protons to the desired treatment depth. The MRS has limitations in clinical applications. Firstly, the air gap between the MRS and the isocenter plane for the non-movable nozzle is usually considerable (e.g. >40 cm). The air gap to the isocenter plane can be reduced to about 25 cm for the movable nozzle before RS hits the couch. It is well known that the air gap distance has a strong impact on the spot size [7]. Hence, to maintain small spot sizes, it is essential to keep the air gap as small as possible. Secondly, it is very challenging for the existing proton analytical algorithms to accurately model this large range of air gaps when an MRS is used. For example, the proton convolution superposition (PCS) in Eclipse TPS (Varian Medical System, Palo Alto, CA, USA) underestimates the spot size distal to an RS because the spot geometric broadening along the air gap between the RS and the patient is not fully considered [7]. This introduces considerable inaccuracies, namely up to 30% errors on spot sizes when the position of RS differs more than 15 cm from the commissioning position [8]. Ding et al. [9] reported their commissioning results of RS for PBS with RayStation TPS (RaySearch Laboratories, Stockholm, Sweden) and recommend using air gaps from 5–10 cm to maintain accurate dose calculations.

Both et al. [3] proposed using a universal bolus that can be placed very close to the patient as an alternative to the MRS. When a patient bolus is employed with air gaps from 2 to 10 cm, the spot size remains nearly identical to the open field. This universal bolus has been successfully implemented for the treatment of brain and HN patients.

Inspired by the success of the aforementioned work by Both et al. [3], a new universal bolus (U-shaped bolus) with an increased dimension in the patient's anterior–posterior direction, an indexed attachable design to allow flexible lateral beam angles and an improved reproducibility on the placement of the bolus was developed. Furthermore, two customized universal patient-related range shifters (UPRS) were designed and implemented to treat different disease sites beyond the brain and HN. In this work, we present the development and clinical implementations of the new U-Shaped bolus and the two new UPRS. We comprehensively reviewed the advantages and disadvantages between UPRS and MRS. Although this is a single institute experience, we hope our experience will

provide useful guidance for other proton institutes on choosing the appropriate RS technique.

Materials and methods

The Roberts Proton Therapy Center at the University of Pennsylvania currently has five PBS treatment rooms, including four gantries and one fixed beam room (IBA, Louvain-La-Neuve, Belgium). Due to low beam efficiency (<1%), the proton PBS system runs with minimum proton energy of 100 MeV, which corresponds to a proton range of 7.71 cm water equivalent thickness (WET). Other proton PBS delivery systems also have a similar limitation, but the threshold energy varies. To overcome this limitation, different designs of range shifters are used to fulfill all clinical needs.

As shown in Fig. 1a, three UPRS made by QFix (Avondale, PA, USA) allow treating targets across the body: U-shaped bolus (UB), anterior lateral bolus (ALB), and couch top bolus (CTB), while the MRS (from IBA) is mounted either on movable (Fig. 1a) or fixed nozzle (Fig. 1b). Homogeneous water equivalent materials of uniform thickness with different densities were employed for UPRS. The materials and components of MRS and UPRS are summarized in Table 1. To verify the uniformity of the slab, the RSs were scanned using computed tomography (CT) scanner (Siemens Somatom Sensation, Munich, Germany) during the acceptance and commissioning. No noticeable structure heterogeneity such as holes or gaps was found by visual inspection of the 3D CT images of the RSs. Their WETs are measured by Zebra (IBA-Dosimetry, Schwarzenbruck, Germany) at multiple locations with a given RS and derived from the measured range differences of the same proton beam with and without RS present in the beam path.

Figure 1c presents an example of an HN case planned with all three UPRS. The UB (Fig. 1f) is the only one that is always included in the CT simulation. When the proton beam is not designed to pass through the UB, the UB may be excluded from body contour and not calculated in dose calculation. Typically, the UB is placed so that its inferior ends are close to the patient's shoulders while remaining comfortable. This is to ensure the lateral field can cover sufficient low neck targets during proton planning without approaching the UB corner. Generally, ALB (Fig. 1e) and CTB are manually added to the plan. The Hounsfield units (HU) of UPRS are overridden with proper values that provide the same range pullback to what is measured. In this example, although there is no beam passing CTB, it is also added and included in the dose calculation volume to increase the freedom in choosing proton beam orientation. All three UPRS are placed at the closest position to the patient and maintain minimal air gaps. UPRS and MRS placement can be

Fig. 1 **a** Patient-related range shifters: U-shaped bolus (UB) indexed on couch top (two tracks on the couch top to let the UB slide in and index its position along the cranial-caudal direction), anterior-lateral bridge bolus (ALB) on a movable cart with indexed height (can move up and down), couch top bolus (CTB). The UB is applied in the clinic to treat brain, head, and neck targets. It allows anterior and left lateral beam in **a** and the ALB on the left side can be manually moved to the right side when it is needed (**d**). ALB is used across the patient body when anterior and lateral fields are planned. CTB is used for all treatments whenever posterior fields are used. **b** MRS mounted on the nozzle on a fixed beam nozzle. **c** An example of a head and neck patient planned with UB, ALB, and CTB in the planning images, but only UB and ALB were in the beam paths. **d** the same setup to **a** but the left ALB is moved to the right side. **e** ALB and CTB. **f** UB only

Table 1 Materials and components of range shifter

Range shifter	Materials	Density (g/cc)	Thickness (cm)	Components
USB	Polyurethane	1.2	6.5	9.4% H, 64.2% C, 3.1% N, 23.3% O
ALB	Polyurethane	1.2	7.4	9.4% H, 64.2% C, 3.1% N, 23.3% O
CTB	G-10	1.85	3.5	54% O, 38% Si, 8% Al, and 1 mm C on top and bottom
IBA MRS	Lexan	1.2	6.48	5.5% H, 75.6% C, 18.9% O

H: hydrogen; C: carbon; N: nitrogen; O: oxygen; Si: silicon; Al: aluminum

checked by using in-house software for potential collisions [10].

Three sets of data for in-air spot size were measured and compared in this study. The air gap and snout position for our system is described in Fig. 2. Firstly, the spot profiles for open field and with MRS at different air gaps (6.0 cm, 16.0 cm, and 26.0 cm) from the isocenter were measured using a Lynx (IBA-Dosimetry, Schwarzenbruck, Germany). Then a simulation study was carried out using an analytic approximation method (“differential Molière” formula) described by Shen et al., [11] for a few air gaps including the ones described above (air gaps of 2.0 cm, 6.0 cm, 10.0 cm, 16.0 cm, and 26.0 cm). To explore the limitation of TPS, a single proton spot with a nominal energy of 110 MeV and 200 MeV was calculated

by the analytic dose calculation algorithm in Eclipse (Pencil Beam Convolution Superposition, version 13.7, Varian Medical Systems) within a phantom for a UPRS with WET of 7.35 cm placed at various locations with air gaps of 6 cm, 16 cm, 26 cm, and 33 cm. For a given energy, the same planning parameters such as the beam geometry, monitor units, and dose calculation volumes were used. Comparisons of lateral profiles at the phantom surface and the percent depth dose curves were generated and analyzed.

To explore the impact of various air gaps on the delivered PBS dose, a single-field optimized PBS field (range 28 cm (~210 MeV), modulation 13 cm) was delivered three times on solid water phantom with the ALB placed at 0, 10, and 20 cm air gaps to the phantom surface. The

two-dimensional dose maps at the middle of the spread-out Bragg peak were measured using the ion chamber array MatriXX PT (IBA-Dosimetry, Schwarzenbruck, Germany) located at isocenter. Measured dose maps were compared to the dose map calculated from Eclipse. Evaluation of gamma passing rates (criteria of 3%, 3 mm with a threshold of 10%) was also carried out for all three measurements.

The 6.20 cm WET UB (Fig. 1f) was designed specifically for HN and brain targets. It allows lateral to vertex treatment fields with small air gaps and has no clearance issue as shown in Fig. 1c. Generally, the patient is immobilized with a thermoplastic mask and the UB can be inserted in two tracks on the treatment couch top with designed indexes along the cranial-caudal direction. During daily treatment, once patient alignment is completed with orthogonal kV images, the UB is inserted to the indexed locations which were decided during the CT simulation. Due to the irregular shape of the UB, the relative position between the UB and the patient is confirmed with orthogonal kilovolts (kV) images by comparing it with the UB contour in the treatment plan, noting especially the edge of the UB. If an oblique field is used (e.g. the right-posterior oblique field in Fig. 1c) the field edge to the end of the UB should be carefully checked. We recommend at least a 1 cm buffer distance to be maintained from the field edge to the bolus edges during the planning stage.

The 7.35 cm WET CTB permits a posterior or posterior oblique proton field to treat shallow targets. The couch top (QFix, Avondale, PA, USA) is capable of sliding longitudinally. With four indexed sliding positions, the couch top permits an extra 63 cm longitudinal extension. This is extremely important for treating long and posterior shallow target volumes such as craniospinal treatments with PBS technology [12]. To reduce the potential uncertainties, a large oblique angle is prohibited to prevent the beam from passing through the couch corner. If a large oblique beam is required, the distance from isocenter to couch edge must be compared to the distance in the treatment plan, typically provided by the planner through a setup note.

The 7.16 cm WET ALB allows proton beams to be delivered from anterior or lateral directions for shallow targets. ALB includes two rectangular slabs (anterior: 40 cm × 60 cm; lateral: 40 cm × 40 cm) that are attached to a movable cart (Fig. 1a). ALB is the only UPRS that has limited ability to adjust the air gap to the patient surface for anterior beams. Three indexed vertical locations 10 cm apart are available to maintain a small air gap and avoid a collision. The distance between the bottom of the bolus and the surface of the couch top is used for every anterior proton beam to verify the consistency between planned and treatment positions of ALB. To minimize the total weight of ALB itself, only one lateral RS is inserted at any time and it can be manually switched to the opposite side if needed. To reduce their impact on

imaging quality, the UB and ALB are removed during the patient initial alignment using orthogonal kV images, while the CTB always remains and the optimized imaging parameters are required to maintain a reasonable anterior–posterior imaging quality. Before the delivery of a field, a therapist reads through all the planned devices and their associated parameters and another therapist confirms all these values.

Results

Different types of UPRS materials may have a different impact on the verification imaging quality, in particular on volumetric imaging such as cone beam computed tomography (CBCT). As the thick couch top is always present, high Z materials [for example G-10 fiberglass (Table 1)] in the couch can cause artifacts degrading the image quality. This kind of degradation could make it a challenge to identify the implanted fiducial markers or bony structures such as a vertebral body during the patient alignment with kilovoltage orthogonal images. Figure 3 presents the comparison of the CBCT images of a pelvis phantom placed on solid water phantom and CTB with the same WET. Both images were collected using the pelvis scanning protocol and presented in the

same window/level settings. Image degradation can be seen especially in the axial view.

As shown in Fig. 4, both measurements and Moliere simulation show that the spot sizes increase with larger air gaps and the increases are much more significant for low energy protons. The UPRS, especially UB and CTB, can always be placed close to the patient; therefore, spot sizes don't change significantly relative to the open beam (less than 2 mm change for air gaps of 2 to 10 cm). MRS would have the same effect for small air gaps. Small air gaps are achievable with MRS with no couch rotations and the snout can be brought close to the patient without causing collisions. However, it may not be always achievable for fields with large oblique angles or associated with complex patient geometries. For example, the small air gap for the lateral field for brain and HN treatment with MRS may cause a collision with the treatment couch or patient shoulder. Therefore, the air gap may increase up to 15 cm (a maximal increase of 4.5 mm on spot size). Often the air gap predicted by the TPS may not reflect the true treatment air gap and collision can still occur between the nozzle and the patient's body or the immobilization device that is not included in the CT scan. To deliver the planned beam, the snout must be pulled back

Fig. 3 **a–c** CBCT images of the RANDO phantom on a 7.5 cm solid water phantom; **d–f** CBCT images of the RANDO phantom on a 3.5 cm G-10 CTB (7.35 cm WET). Artifacts are observed and the bony structures are blurred in the axial view. The green crosses in the images represent the imaging isocenters and the similar isocenter locations indicate the same CBCT scanning geometry

and the air gap must be increased. As a result, the delivered treatment may be suboptimal.

As shown in Fig. 5, the comparison studies in Eclipse show no change in spot size or percent depth dose (PDD) when the air gaps increase from 6 to 33 cm. The same PDDs are expected as shown in Fig. 5c since the same RS is placed in the beam path and should cause the same amount of pullback on the range of the protons. Thus, the protons should eventually stop at the same depth. However, the trend of an increase in spot size with larger air gaps presented in Fig. 5 is not found in the comparison study in Eclipse. According to Fig. 5d, Eclipse does not consider the effect of the various air gaps on the spot size when a single proton spot travels in between the RS and patient. Here PDDs and spot profiles for air gaps of 6 cm and 33 cm are presented, but results for the other two intermediary simulated air gaps of 16 cm and 26 cm are found to be consistent with that of the two extreme conditions.

Dose deviations between measured and planned dose maps for increased air gaps were found in iso-dose level comparisons and gamma evaluations (3%/3 mm), as shown in Fig. 6. When no air gaps exist between ALB and phantom, a great agreement was observed (Fig. 6a) and the gamma analysis shows 99.5% of the points of interest pass the comparison criteria. As the air gap increases to 10 cm, iso-dose lines start to deviate from the calculated

ones. As clearly shown in Fig. 6c, the 90% isodose line shrinks and the 10% isodose line expands, which becomes more significant when the air gap increases to 20 cm (Fig. 6e). The failure rate of gamma evaluation increases from 0.5% to 5.73% and 25.5% when the air gaps increase from 0 to 10 and 20 cm respectively. Most of the failures occur in the dose regions of 80% to 100% and 10% to 30% of the prescribed dose, where the dose gradient changes. As indicated in Fig. 4, the spot size increases with air gaps for all energy levels. Regardless of the size of the air gap, each spot carries the same integral dose, but the dose spreads laterally more with a larger air gap. As a result, the dose changes little at the center of a uniform dose region due to dose equilibrium as dose loss from a single spot is compensated by dose tails of the surrounding spots. The equilibrium is broken at the region where the dose gradient changes (80–100% dose) which can potentially cause under-coverage of the target volume. At the field edge, the dose spreads out further with increased air gaps due to larger spot sizes potentially resulting in a higher dose to surrounding OARs. In this current study, we did not model the air gaps for ALB in the treatment planning system, which contributed to the low passing gamma rate.

Discussion

A small spot is especially critical in maintaining a high quality of treatment for HN and brain targets where critical organs are often close to the target volume. Quan et al. evaluated the multi-field and single-field optimization for PBS treatment of HN cancer, which reported that with smaller spot sizes, better normal tissue sparing without sacrificing target coverage can be obtained [13]. Wang et al. studied the impact of spot size on plan quality of PBS radiosurgery for peripheral brain lesions and concluded that the spot size at the patient surface must be small to achieve comparable or smaller brain necrosis normal tissue complication probability (NTCP) relative to photon radiosurgery techniques [4]. Kang et al. presented a case using a patient-specific bolus to maintain a small spot size during the treatment of periorbital disease [14]. Van de Water et al. investigated salivary gland sparing in oropharyngeal cancer for PBS technique with a reduced spot size and concluded that reduced spot improved sparing of the salivary glands and reduced NTCP for xerostomia and parotid and submandibular salivary dysfunction [15]. Kralik et al. demonstrated that the smaller spot size, afforded by a PBS-dedicated nozzle relative to a universal nozzle, allows sharper dose distribution fall-off around the target and greater adjacent normal tissue sparing for pediatric brain tumors [16]. Therefore, the use of UB, which consistently reduces the air gap to the patient's surface and maintains the integrity

Fig. 5 Comparison study for 110 MeV single proton spot with UPRS at different locations (indicated by the green boxes) in Eclipse TPS. **a** Planning geometry with 6 cm air gaps; **b** planning geometry with 33 cm air gaps. **c** the percent depth dose curves for 6 and 33 cm air gaps; the depth 0 cm corresponds to the entrance of the beam, which is the top surface of the first rectangular contour along beam direction (yellow dash line). **d** spot profiles at the phantom surface for 6 and 33 cm air gaps

of the original spot, is beneficial for HN and brain proton therapy with the PBS technique.

The movable nozzle with MRS can bring the range shifters closer to the patient. In a clinic, an air gap of 10–20 cm is often employed. It is not rare to see during planning that a large air gap must be used to avoid collision between the bulky gantry and the patient, or the patient immobilization device or the couch. Especially, when the scanned length of a CT image is limited, a plan with a more conservative air gap is preferred to avoid complications on the treatment day. For example, a collision between the gantry and the patient's arms when the patient's arms are up above the head, or a breath-hold device can occur for a breast or thoracic patient when the patient's arm or breath-hold device is not included in the CT scan. Another common example is the collision between the nozzle and the couch at a lateral beam or posterior oblique beam angle for an HN treatment when the planer tries to achieve a small air gap during the planning stage. In contrast, when a UPRS, especially for UB and CTB is used, the air gap can be reduced to a few cm (very challenge for MRS) and there is no concern about

the potential collision. When a patient is treated with an oblique beam angle or a couch rotation, the reduction of the air gap will be even larger.

The application of the three-dimensional (3D) printing technique to radiation therapy is of increasing interest and application [17]. 3D printed devices have been used to modulate proton beam energy [18] and served as a customized bolus for passive-scattering proton beams. Because of the high precision of 3D printing, this technique can produce a device that best fits the shape of a patient's body such as the head and extremities. Besides acting as an immobilization device, the device can also serve as RS to allow proton treatment to shallow targets. The use of 3D-printed RS has its unique benefits including negligible air gap and patient or beam specific design. As a non-removable device, 3D-printed RS may cause a challenge in imaging the patient similar to CTB. However, multiple printing materials are available on the market, and the materials such as PLA and polyamide, without containing any high Z components, potentially have little impact on volumetric imaging quality. As described by Zou [19], the 3D printing technique

can typically achieve high precision during the manufacture of a device, which is critical for proton therapy. With lower cost and improved printing efficiency, this 3D printing application in PBS radiotherapy is promising.

All the UPRS devices are included in the patient body contour to be accounted for accurately in the proton dose calculation, as shown in Fig. 1c. Small air gaps (< 10 cm) could typically be maintained for UB and CTB, while the determination of the air gaps between the ALB and patient rely on our experience (such as treatment site and patient position) and collision may be predicted by in-house program [10]. Compared to the MRS, the UPRS devices are not interlocked with the beam delivery system. In some cases, multiple UPRS and MRS are used for a single treatment. Thus, effective procedures and clinic workflow along with enhanced communication within the radiation oncology team must be maintained to provide safe and accurate treatment.

The material, dimension, and weight of the UPRS should be considered when the UPRS is designed. Shen reported the impact of the range shifter material on PBS spot characteristics and show that Lexan

and Lucite are desirable range shifter materials as their scattering power is similar to water [11]. The minimal energy of a proton system determines the maximal thickness of an RS required to treat a surface target. Thinner RS CTB and UB that are attached to the couch system may reduce the total weight limit of the system. For example, the CTB reduces the weight limit of 440 lbs. for the regular QFix treatment couch to 350 lbs. When a UB (25 lbs.) is employed, the total weight limit is further reduced. To maintain a reasonable weight limit, the ALB is designed to be independent of the couch system. Besides, the use of UPRS often restricts proton beam angles. For example, CTB only allows a posterior field with a limited oblique angle. Due to the flat surface and the limited dimension of the RS, a large oblique angle or couch kick may not be feasible for ALB. In contrast, MRS provides more freedom on beam angle selection while the air gap may depend on an individual case. Some current proton systems allow minimal proton energy of 70 MeV, which corresponds to a WET of 4.1 cm. As a result, the scattering effect, potential collision, reduction of the weight limit of

the patient supporting system, and the degradation of image quality from a bulky RS are lower.

HN planning is always one of the challenges in radiotherapy due to the complex target shapes and the proximity of OARs to target volumes. To achieve the best target coverage and maximal sparing of surrounding OARs, the smallest proton beam size has to be maintained for PBS treatment. Thus, air gaps need to be kept as small as possible. To pursue this goal, we standardize the HN planning with two posterior oblique fields and one anterior field with a gradient dose match [20]. For unilateral and simple target volume (for example, in Fig. 1c) optimal field angles (short and simple beam path) other than standard field geometry along with proper RS may be selected based on the size and location of the target.

Recently, Lin et al. [21] evaluated the accuracy of the AcurosPT algorithm, a commercial Monte Carlo algorithm in the Eclipse 13.7 TPS. Although the new algorithm showed a faster increase of spot size along with depth than measurements due to overestimated scattering effect, high consistency between the calculated and measured spot size at the phantom surface was reported for different air gaps. These results indicated that the relationship between the air gap and spot size was accurately modeled by this new algorithm in the next version of Eclipse, which provided improved accuracy on proton dose calculation.

Since our group deployed the first in-room volumetric imaging in proton therapy [22], fixed UPRS devices may not be desirable as the interference to the image and clearance issue may decrease the utility of the device. The image quality of CBCT is limited by scattered radiation [23], reconstruction artifacts from high-density objects [24], and objects extending outside the scanning field of view [25]. The fixed UPRS affects the image quality in various ways as X-rays pass through the RS slab. Firstly, the x-ray beam gets attenuated; secondly, scattered radiation increases; thirdly, UPRS extended outside the field of view leads to reconstruction artifacts. Besides UPRS like CTB, containing high-density materials introduces more artifacts to the CBCT image and further degrade the image quality. However removable UPRS such as the Universal Bolus [3] may provide superior dose distribution while not affecting volumetric imaging, although the workflow would have to be modified.

Conclusion

We have described how to implement UPRS and MRS for various clinical indications using the PBS technique, and comprehensively reviewed the advantage and disadvantages of UPRS and MRS. We recommend the removable UB only to be employed for the brain and HN treatments while an automated MRS is used for

all proton beams that require RS but not convenient or feasible to use UB. During the development of UPRS, its impact on the imaging system should be investigated. While using MRS, it is essential to verify that the TPS accurately models the air gap.

Abbreviations

RS: Range shifter; MRS: Machine-related range shifter; UPRS: Universal patient-related range shifter; UB: U-shaped bolus; ALB: Anterior lateral bolus; CTB: Couch top bolus; IGRT: Image-guided radiation therapy; PBS: Pencil beam scanning; HN: Head and neck; TPS: Treatment planning system; PCS: Proton convolution superposition; WET: Water equivalent thickness; CT: Computed tomography; HU: Hounsfield units; CBCT: Cone-beam computed tomography; PDD: Percent depth dose; NTCP: Normal tissue complication probability; 3D: Three dimensional.

Acknowledgements

We would like to thank our radiation oncology colleagues at the Roberts Proton Therapy Center at the University of Pennsylvania for their continuous supports of this project. Also, we would like to acknowledge Michael Kuske from Purdue University for helping us to improve the quality of the manuscript.

Authors' contributions

Dr. Haibo Lin, Dr. Jiajian Shen, Dr. Minglei Kang, Dr. Stefan Both, Ms. Huifang Zhai, and Dr. James McDonough were involved in the development and clinical implantation of all the devices. Dr. Zelig Tochner, Dr. Curtland Deville, and Dr. Charels B. Simone II were the major team members providing insight from the physician's point of view; Dr. Haibo Lin, Dr. Chengyu Shi, Dr. Sheng Huang, Mr. Qing Chen and were responsible for drafting the manuscript as well as all the data collection and analysis. All authors read and approved the final manuscript.

Funding

The project was partially supported by the U.S. Department of Defense (DOD) Grant W81XWH-04-2-0022.

Availability of data and materials

Not applicable.

Declarations

Ethical approval and consent to participate

Not applicable.

Consent for publication

Not applicable.

Competing interests

The authors declare that they have no competing interests.

Author details

¹New York Proton Center, New York 10035, USA. ²Mayo Clinic, Phoenix, AZ 85054, USA. ³Department of Radiation Oncology, University of Pennsylvania, Philadelphia, PA 19104, USA. ⁴Department of Radiation Oncology and Molecular Radiation Sciences, Johns Hopkins University, Baltimore, MD 21231, USA. ⁵Department of Radiation Oncology, University Medical Center Groningen, Groningen 9713 GZ, Netherlands.

Received: 1 March 2021 Accepted: 29 July 2021

Published online: 06 August 2021

References

- Particle Therapy Co-Operative Group. Particle therapy facilities in operation: information about technical equipment and patient statistics. 2021. <http://ptcog.ch/index.php/facilities-in-operation>.

2. Ding XF, Dionisi F, Tang SK, Ingram M, Hung CY, Prionas E, Lichtenwalner P, Butterwick I, Zhai HF, Yin LS, et al. A comprehensive dosimetric study of pancreatic cancer treatment using three-dimensional conformal radiation therapy (3DCRT), intensity-modulated radiation therapy (IMRT), volumetric-modulated radiation therapy (VMAT), and passive-scattering and modulated-scanning proton therapy (PT). *Med Dosim.* 2014;39(2):139–45.
3. Both S, Shen JJ, Kirk M, Lin LY, Tang SK, Alonso-Basanta M, Lustig R, Lin HB, Deville C, Hill-Kayser C, et al. Development and clinical implementation of a universal bolus to maintain spot size during delivery of base of skull pencil beam scanning proton therapy. *Int J Radiat Oncol.* 2014;90(1):79–84.
4. Wang DX, Dirksen B, Hyer DE, Buatti JM, Sheybani A, Dinges E, Felderman N, TenNapel M, Bayouth JE, Flynn RT. Impact of spot size on plan quality of spot scanning proton radiosurgery for peripheral brain lesions. *Med Phys.* 2014;41(12):121705.
5. Hyer DE, Hill PM, Wang DX, Smith BR, Flynn RT. Effects of spot size and spot spacing on lateral penumbra reduction when using a dynamic collimation system for spot scanning proton therapy. *Phys Med Biol.* 2014;59(22):N187–96.
6. Titt U, Mirkovic D, Sawakuchi GO, Perles LA, Newhauser WD, Taddei PJ, Mohan R. Adjustment of the lateral and longitudinal size of scanned proton beam spots using a pre-absorber to optimize penumbræ and delivery efficiency. *Phys Med Biol.* 2010;55(23):7097–106.
7. Schaffner B. Proton dose calculation based on in-air fluence measurements. *Phys Med Biol.* 2008;53(6):1545–62.
8. Matysiak W, Yeung D, Slopsema R, Li Z. Evaluation of the range shifter model in a commercial proton therapy planning system. *Int J Radiat Oncol.* 2014;90:5917–5917.
9. Ding X, Rosen L, Wu H. Commission range shifter on a spot scanning proton system using raystation treatment planning system. *Med Phys.* 2015;42(6):3375–3375.
10. Zou W, Lin HB, Plastaras JP, Wang HS, Bui V, Vapiwala N, McDonough J, Tochner Z, Both S. A clinically feasible method for the detection of potential collision in proton therapy. *Med Phys.* 2012;39(11):7094–101.
11. Shen JJ, Liu W, Anand A, Stoker JB, Ding XN, Fatyga M, Herman MG, Bues M. Impact of range shifter material on proton pencil beam spot characteristics. *Med Phys.* 2015;42(3):1335–40.
12. Lin HB, Ding XF, Kirk M, Liu HY, Zhai HF, Hill-Kayser CE, Lustig RA, Tochner Z, Both S, McDonough J. supine craniospinal irradiation using a proton pencil beam scanning technique without match line changes for field junctions. *Int J Radiat Oncol.* 2014;90(1):71–8.
13. Quan EM, Liu W, Wu R, Li YP, Frank SJ, Zhang XD, Zhu XR, Mohan R. Preliminary evaluation of multifield and single-field optimization for the treatment planning of spot-scanning proton therapy of head and neck cancer. *Med Phys.* 2013;40(8):081709.
14. Kang ML, Hasan S, Press RH, Yu F, Abdo M, Xiong WJ, Choi JJ, Simone CB, Lin HB. Using patient-specific bolus for pencil beam scanning proton treatment of periorbital disease. *J Appl Clin Med Phys.* 2021;22(1):203–9.
15. Van de Water TA, Lomax AJ, Bijl HP, Schilstra C, Hug EB, Langendijk JA. Using a reduced spot size for intensity-modulated proton therapy potentially improves salivary gland-sparing in oropharyngeal cancer. *Int J Radiat Oncol Biol Phys.* 2012;82(2):e313–319.
16. Kralik JC, Xi L, Solberg TD, Simone CB 2nd, Lin L. Comparing proton treatment plans of pediatric brain tumors in two pencil beam scanning nozzles with different spot sizes. *J Appl Clin Med Phys.* 2015;16(6):41–50.
17. Mayer R, Liacouras P, Thomas A, Kang M, Lin L, Simone CB 2nd. 3D printer generated thorax phantom with mobile tumor for radiation dosimetry. *Rev Sci Instrum.* 2015;86(7):074301.
18. Lindsay C, Kumlin J, Jirasek A, Lee R, Martinez DM, Schaffer P, Hoehr C. 3D printed plastics for beam modulation in proton therapy. *Phys Med Biol.* 2015;60(11):N231–240.
19. Zou W, Fisher T, Zhang M, Kim L, Chen T, Narra V, Swann B, Singh R, Siderit R, Yin LS, et al. Potential of 3D printing technologies for fabrication of electron bolus and proton compensators. *J Appl Clin Med Phys.* 2015;16(3):90–8.
20. Apinorasetkul O, Kirk M, Teo K, Swisher-McClure S, Lukens JN, Lin A. Pencil beam scanning proton therapy vs rotational arc radiation therapy: A treatment planning comparison for postoperative oropharyngeal cancer. *Med Dosim.* 2017;42(1):7–11.
21. Lin L, Huang S, Kang M, Hiltunen P, Vanderstraeten R, Lindberg J, Siljamäki S, Wareing T, Davis I, Barnett A, et al. A benchmarking method to evaluate the accuracy of a commercial proton monte carlo pencil beam scanning treatment planning system. *J Appl Clin Med Phys.* 2017;18(2):44–9.
22. Veiga C, Janssens G, Teng CL, et al. First clinical investigation of cone beam computed tomography and deformable registration for adaptive proton therapy for lung cancer. *Int J Radiat Oncol Biol Phys.* 2016;95(1):549–59.
23. Endo M, Tsunoo T, Nakamori N, Yoshida K. Effect of scattered radiation on image noise in cone beam CT. *Med Phys.* 2001;28(4):469–74.
24. Williamson JF, Whiting BR, Benac J, Murphy RJ, Blaine GJ, O'Sullivan JA, Polite DG, Snyder DL. Prospects for quantitative computed tomography imaging in the presence of foreign metal bodies using statistical image reconstruction. *Med Phys.* 2002;29(10):2404–18.
25. Ohnesorge B, Flohr T, Schwarz K, Heiken JP, Bae KT. Efficient correction for CT image artifacts caused by objects extending outside the scan field of view. *Med Phys.* 2000;27(1):39–46.

Publisher's Note

Springer Nature remains neutral with regard to jurisdictional claims in published maps and institutional affiliations.

Ready to submit your research? Choose BMC and benefit from:

- fast, convenient online submission
- thorough peer review by experienced researchers in your field
- rapid publication on acceptance
- support for research data, including large and complex data types
- gold Open Access which fosters wider collaboration and increased citations
- maximum visibility for your research: over 100M website views per year

At BMC, research is always in progress.

Learn more biomedcentral.com/submissions