University of Groningen

Comparison of the X-ray tube spectrum measurement using BGO, NaI, LYSO, and HPGe detectors in a preclinical mini-CT scanner

Lohrabian, Vahid; Kamali-Asl, Alireza; Harvani, Hossein Ghadiri; Aghdam, Seyed Rashid Hosseini; Arabi, Hossein; Zaidi, Habib

Published in:
 RADIATION PHYSICS AND CHEMISTRY

DOI:
[10.1016/j.radphyschem.2021.109666](https://doi.org/10.1016/j.radphyschem.2021.109666)

IMPORTANT NOTE: You are advised to consult the publisher's version (publisher's PDF) if you wish to cite from it. Please check the document version below.

Document Version
 Publisher's PDF, also known as Version of record

Publication date:
 2021

[Link to publication in University of Groningen/UMCG research database](#)

Citation for published version (APA):

Lohrabian, V., Kamali-Asl, A., Harvani, H. G., Aghdam, S. R. H., Arabi, H., & Zaidi, H. (2021). Comparison of the X-ray tube spectrum measurement using BGO, NaI, LYSO, and HPGe detectors in a preclinical mini-CT scanner: Monte Carlo simulation and practical experiment. *RADIATION PHYSICS AND CHEMISTRY*, 189, [109666]. <https://doi.org/10.1016/j.radphyschem.2021.109666>

Copyright

Other than for strictly personal use, it is not permitted to download or to forward/distribute the text or part of it without the consent of the author(s) and/or copyright holder(s), unless the work is under an open content license (like Creative Commons).

The publication may also be distributed here under the terms of Article 25fa of the Dutch Copyright Act, indicated by the "Taverne" license. More information can be found on the University of Groningen website: <https://www.rug.nl/library/open-access/self-archiving-pure/taverne-amendment>.

Take-down policy

If you believe that this document breaches copyright please contact us providing details, and we will remove access to the work immediately and investigate your claim.

Downloaded from the University of Groningen/UMCG research database (Pure): <http://www.rug.nl/research/portal>. For technical reasons the number of authors shown on this cover page is limited to 10 maximum.

Comparison of the X-ray tube spectrum measurement using BGO, NaI, LYSO, and HPGe detectors in a preclinical mini-CT scanner: Monte Carlo simulation and practical experiment

Vahid Lohrabian^a, Alireza Kamali-Asl^{a,*}, Hossein Ghadiri Harvani^b,
Seyed Rashid Hosseini Aghdam^a, Hossein Arabi^c, Habib Zaidi^{c,d,e,f}

^a Department of Medical Radiation Engineering, University of Shahid Beheshti, Tehran, Iran

^b Medical Physics and Biomedical Engineering Department, and Research Center for Molecular and Cellular Imaging, Tehran University of Medical Sciences, Iran

^c Division of Nuclear Medicine and Molecular Imaging, Department of Medical Imaging, Geneva University Hospital, CH-1211, Geneva 4, Switzerland

^d Geneva University Neurocenter, Geneva University, CH-1205, Geneva, Switzerland

^e Department of Nuclear Medicine and Molecular Imaging, University of Groningen, University Medical Center Groningen, 9700, RB, Groningen, Netherlands

^f Department of Nuclear Medicine, University of Southern Denmark, DK-500, Odense, Denmark

ARTICLE INFO

Keywords:

Computed tomography
Monte Carlo simulation
Energy resolution
Solid-state detectors

ABSTRACT

Background: In diagnostic X-ray computed tomography (CT) imaging, some applications, such as dose measurement using the Monte Carlo method and material decomposition using dual/multi-energy approaches, rely on accurate knowledge of the energy spectrum of the X-ray beam. In this regard, X-ray detectors providing an accurate estimation of the X-ray spectrum could greatly impact the quality of dual/multi-energy CT imaging and patient-specific dosimetry.

Purpose: The aim of this study is to estimate the intrinsic efficiency and energy resolution of different types of solid-state gamma-ray detectors in order to generate a precise dual-energy X-ray beam from the conventional x-ray tube using external X-ray filters.

Materials and methods: The X-ray spectrum of a clinical X-ray tube was experimentally measured using a high purity Germanium detector (HPGe) and the obtained spectrum validated by Monte Carlo (MC) simulations. The obtained X-ray spectrum from the experiment was employed to assess the energy resolution and detection efficiency of different inorganic scintillators and semiconductor-based solid-state detectors, namely HPGe, BGO, NaI, and LYSO, using MC simulations. The best performing detector was employed to experimentally create and measure a dual-energy X-ray spectrum through applying attenuating filters to the original X-ray beam.

Results: The simulation results indicated 9.16% energy resolution for the HPGe detector wherein the full width-at-half-maximum (FWHM) of the energy resolution for the HPGe detector was about 1/3rd of the other inorganic detectors. The X-ray spectra estimated from the various source energies exhibited a good agreement between experimental and simulation results with a maximum difference of 6%. Owing to the high-energy discrimination power of the HPGe detector, a dual-energy X-ray spectrum was created and measured from the original X-ray spectrum using 0.5 and 4.5 mm Aluminum external filters, which involves 70 and 140 keV energy peaks with 8% overlap.

Conclusion: The experimental measurements and MC simulations of the HPGe detector exhibited close agreement in high-energy resolution estimation of the X-ray spectrum. Given the accurate measurement of the X-ray spectrum with the HPGe detector, a dual-energy X-ray spectrum was generated with minimal energy overlap using external X-ray filters.

* Corresponding author.

E-mail address: a_r_kamali@yahoo.com (A. Kamali-Asl).

<https://doi.org/10.1016/j.radphyschem.2021.109666>

Received 31 October 2020; Received in revised form 28 May 2021; Accepted 21 June 2021

Available online 24 June 2021

0969-806X/© 2021 Elsevier Ltd. All rights reserved.

1. Introduction

Advances in detector technologies and accurate extraction of the energy spectrum of radioactive sources or X-ray generators have improved the quality and accuracy of X-ray imaging. In this regard, due to the low detection efficiency of gamma-ray detectors, the calibration of organic detection devices is still considered a significant challenge for gamma-rays with high energies. The reason for this is that the scintillation detectors do not generate any photopeak in their spectral response, especially for gamma rays with energies above 100 keV (Peyvandi et al., 2015). Precise knowledge about the X-ray energy spectrum of a diagnostic X-ray beam could aid to enhance the image quality and reliably estimate the absorbed dose to the patient (Arabi and Asl, 2010). In preclinical imaging researches, varieties of X-ray spectra are employed depending on the subjects under study and the end-point of the tasks which may be far different from the ones used in clinical practices (Jia et al., 2009).

Sodium Iodide activated with Thallium, NaI(Tl), is the most widely used scintillation material. The NaI(Tl) and high pure germanium (HPGe) detectors, in addition to their applications in radiation and dosimetry systems, are considered ideal gamma-ray detectors owing to their good energy resolution and detection efficiency. HPGe detectors are made of extremely pure p-type or n-type germanium, wherein *the active volume of these detectors is large enough for efficient and high-energy resolution gamma-ray spectroscopy. HPGe detectors have two advantages over germanium crystals that were doped with lithium ions (GeLi) detectors: 1. they are able to function properly in the ambient temperature and 2. The production cost and procedure are much cheaper and easier. The energy resolution of an HPGe detector is usually expressed as the full width at half maximum (FWHM) for the 1332.5 keV energy peak of the Co-60 source. However, the energy resolution for a scintillation detector is usually expressed as the percentage of the relative efficiency for the energy peak of a Cs-137 source (Hossain et al., 2012; Ješkovský et al., 2019). This parameter defines the capability of the detector to distinguish the different energy peaks in a multi-energy photon beam. In general, the energy resolution is commonly determined/affected by the factors such as; 1) fluctuation in the light generation for the same mono-energy irradiation, 2) the non-proportionality of light response, 3) statistical fluctuation in charge collection in the anode of the photomultiplier tube and 4) electronics noise. In Monte Carlo simulation (particularly MCNP platform), Gaussian energy broadening (GEB) is a special order for tallies to better simulate a radiation detector whose energy peaks exhibit Gaussian energy broadening. Estimation of the Gaussian energy broadening parameters of the gamma-ray detectors plays a critical role in accurate modeling and interpretation of the resulting energy spectra (Ješkovský et al., 2019; Taheri et al., 2016).*

In this regard, an energy broadening model for Bismuth germanium oxide (BGO) detector was developed by Askari et al. (2018a) wherein the GEB parameters for a 10×10 BGO array were calculated for the task of spectral response analysis in the presence of the wide energy range of gamma-rays in a simulation setting. Bugby et al. (2016) conducted a comparison study between Thallium activated Cesium Iodide (CsI:TI) and Gadolinium oxysulfide (GOS) scintillators for the use in a portable gamma camera using the GEB parameters of the Taheri et al. study (Taheri et al., 2016). This study demonstrated that the major difference between these two types of scintillator lies in their intrinsic spatial resolution, which was calculated to be 230 μm for the CsI:TI scintillator and 1090 μm for the GOS scintillator. Goodell et al. (Goodell et al., 2017) validated a Monte Carlo model for HPGe detector in order to predict the product ratios (chain) of activities in Fe, Cr, Ni, and stainless-steel foils. Ordóñez et al. (Ordóñez et al., 2019) investigated the full energy peak efficiency curves for an HPGe detector using MCNP6 and GEANT4. A Monte Carlo simulation of the efficiency of the HPGe detector was conducted by Azbouche et al. (2015) to measure a large volume of environmental samples. The GEB parameters of an HPGe Detector were

estimated by Eftekhari Zadeh et al. (EftekhariZadeh et al., 2014) to measure the energy resolution of GMX (GAMMA-X) series of coaxial detector systems.

Since the GEB parameters and detector energy resolution are the determining factors for accurate modeling of the detector response, the precise estimation of these parameters is the key for reliable evaluation of these detectors (Ordóñez et al., 2019; Askari et al., 2018b). There have been always interests in designing detectors with the highest possible sensitivity (to detect very low activity concentration) and high spatial resolution (to localize the spatial distribution of the activity concentration). Normally, there is a fundamental tradeoff between improvements in sensitivity and spatial resolution of a detector, wherein strategies such as reducing the field of view (FOV) or modifying the distances between detector, object, and X-ray source (variable magnification) have been adopted to find an optimal compromise between these two key parameters (Lohrabian et al., 2019; Arabi et al., 2015; Arabi, KAMALIASL, Aghamiri; Kamali-Asl et al., 2009; Arabi et al., 2011; Sanaat et al., 2020).

HPGe is one of the highly accurate detectors for reliable measurement of the gamma-ray radionuclides and X-ray energy spectra (Boson et al., 2008; Olivares et al., 2017). Monte Carlo simulation is normally employed for precise modeling of the HPGe-based detection systems and estimation of the outcome spectra typically using either the Monte Carlo N-Particle radiation transport code (MCNP) or the Geant4 toolkit (Goorley et al., 2012). Determination of the X-ray spectrum plays a key role in Monte Carlo-based dose calculations, dual- and/or multi-energy CT imaging, and material decomposition. Currently, there is a growing interest in patient-specific dose estimation from CT scans as well as patient-specific material/tissue decomposition using dual- or multi-energy CT imaging in clinical practices. Therefore, accurate X-ray spectroscopy approaches are highly required for such applications. However, due to the high photon flux, normally used in X-ray imaging, it is challenging to directly measure the spectrum of a CT scanner X-ray source. Different methods have been developed to estimate CT spectra which could be classified into two different groups, namely model-based (generating spectra from empirical or semi-empirical physical models), measurement-based (reconstructing/estimating spectra from measured data) (Duan et al., 2011).

One of the key components of CT scanners is the detection system which greatly contributes to the quality of the resulting images. Cadmium zinc telluride-based (CZT) semiconductor detectors are widely used in medical imaging systems owing to their excellent energy resolution, the capability of providing high spatial resolution, and proper detection efficiency for a wide range of gamma-ray energies (Bushberg and Boone, 2011). Other types of gamma-ray detectors introduced in CT scanners are inorganic scintillators crystals NaI (Tl) and Bismuth Germinate (BGO).

The attenuation coefficients, estimated from CT imaging, depend on the energy of the X-ray and the physical properties of the materials/tissues (mass density and atomic number). However, different materials can have the same attenuation coefficient at a certain X-ray energy range while they have distinctive physical properties. For this reason, dual-energy techniques have been introduced in industry and clinical settings to accurately identify the underlying physical properties of the materials/tissues. The application of dual-energy CT imaging would be material differentiation (Johnson et al., 2007; Gholamiankhah et al., 2104) such as bone removal in angiography and bone marrow edema, material optimization such as metal artifact reduction and detection of urinary stones (Arabi and Zaidi, 2021; Ghorbanzadeh et al., 2105), material decomposition such as iodine maps generation and kidney lesion characterization, and generation of atomic number or stopping power map for proton (heavy ions) radiation therapy in clinical practice (Sanghavi and Jankharia, 2019; Arabi and Zaidi, 2020). Moreover, dual-energy CT imaging is used for material decomposition, detection, and characterization (such as luggage/cargo control) in the industry (Fuchs et al., 2013). The generated dual-energy X-ray spectrum in this

study is intended to be used for clinical applications, in particular, stopping power generation for proton radiation therapy as well as material decomposition.

Today, dual- and/or multi-energy CT imaging systems have become a popular approach to discriminating biological tissues in clinical and research settings. In this framework, the image is formed by two different energies, wherein an algorithm of image processing is employed to extract/detect specific tissue types separately. The underlying idea behind this approach is to use two specific energy ranges with the maximum separation between their spectra. Hence, the accurate extraction of the X-ray tube spectra plays a key role in the overall performance of this approach (Clark et al., 2018). To achieve dual-energy CT imaging, two X-ray source/detector systems rotate simultaneously to acquire images at different X-ray energy. Each set of source/detector system is intended to generate a specific energy range and capture the corresponding data. Alternatively, a single set of source/detector could be used, where the X-ray source would switch between two defined energy levels to enable dual-energy CT imaging. Moreover, external filters could be used with the conventional X-ray sources to generate favorable X-ray spectrums. By changing the filter thickness, layers, or materials, different X-ray spectrums could be generated (Goo and Goo, 2017; Rebuffel and Dinten, 2007). To acquire images at two different X-ray energy levels, gas detectors, solid-state X-ray detectors (amorphous silicon and germanium), scintillator-based detectors are used for measurement/detection of X-ray intensity (Pfeiffer et al., 2020; Åslund et al., 2010; Alvarez et al., 2004). Owing to the superior energy resolution of the high-purity germanium detector, this detector could be an excellent option for dual-energy CT imaging (Panth et al., 2020).

In this research, we set out to measure the X-ray spectrum of a mini-CT scanner using several detector models in order to generate a dual-energy X-ray source through applying attenuating filters to the X-ray beam. To this end, Monte Carlo simulation was employed to evaluate different detector types and then the detector with the best/satisfactory performance was chosen for the practical experiment. The focuses will be on the evaluation of the key detection characteristics such as intrinsic efficiency and energy resolution. The comparison of the different detector types was carried out in Monte Carlo simulation as well as practical experiments.

2. Materials and methods

The major goal of this study is to extract the energy spectrum of an X-ray tube in order to generate a dual-energy X-ray beam with two separate energy peaks. To this end, four different radiation detectors were examined in the MC simulation environment to choose/select the detector with the highest energy resolution and detection efficiency. The HPGe exhibited superior performance and hence was selected to experimentally measure the energy spectrum of the X-ray tube. To validate the experimentally obtained results, the X-ray tube was simulated in the MC simulation environment to create a baseline for comparison. Given the experimentally obtained X-ray spectrum, external attenuating filters were designed and applied to create a dual-energy X-ray beam with separate energy peaks of 70 and 140 keV. In the following, we elaborate on each of the above-mentioned steps to generate a dual-energy X-ray spectrum.

2.1. Gamma-ray detectors

In order to detect and monitor ionizing radiation at higher energies, bulky scintillator materials were introduced in the form of a single crystal such as Tl-doped NaI and CsI single crystals (Van Sciver and Hofstadter, 1951). One of the important parameters in the simulation of the scintillators with the MCNPX (MCNP is an open-source Monte Carlo N-Particle simulation code) is the Gaussian energy broadening (GEB) (Taheri et al., 2016; Ordóñez et al., 2019). The estimation of this parameter would aid precise simulation of the spectral response of the

Table 1

Specification of the γ & X-ray crystals (Bugby et al., 2016; Askari et al., 2018b; Perez-Andujar and Pibida, 2004; Ryzhikov et al., 2005). The emission peak of a detector refers to the maximum wavelength of the photons that the detector emits.

Detector	BGO	LYSO	NaI	CWO	GOS
Effective atomic number	75	66	56	65	61.1
Density (g/cm ³)	7.13	7.1	3.67	7.9	7.3
Emission Peak (nm)	480	428	415	475	545
Chemical formula	Bi ₄ Ge ₃ O ₁₂	Lu ₂ (1-x)Y _{2x} SiO ₅	NaI (TI)	CdWO ₄	Gd ₂ O ₂ S

Table 2

Characteristics of the HPGe detector.

device type	model	component	Dimension (mm)
detector model number	GMX40P4-83	Ge crystal diameter	69.8
preamplifier model	A257N	Ge crystal length	89.5
cryostat configuration	CFG-PG4-1.2	Core hole diameter	11.6
HV filter model	138 EMI	Aluminum end cap	1.5

scintillator using the Monte Carlo framework. The focus of this study is on the inorganic as well as HPGe radiation detectors. They are several solid-state detectors such as Bismuth Germinate (BGO), sodium iodide (NaI), and Lutetium-yttrium oxyorthosilicate (LYSO) which are mostly used in SPECT, PET, CT, bone densitometers, and medical probes (Perez-Andujar and Pibida, 2004).

The X-ray spectrum of a mini-CT scanner for the energy range of 20 keV–160 keV was measured in a laboratory setup using several scintillator gamma-ray detectors listed in Table 1 with their major specifications. Moreover, the energy spectrum of a portable cone-beam X-ray tube (Medex instrument, model GemX-160) with a focal spot area of $0.5 \times 0.7 \text{ mm}^2$ (500 μm nominal focal spot) was estimated in the MC simulation environment (MCNPX platform). The X-ray tube with a LiPo battery power supply of 28–35 Vdc was employed which is capable of generating X-ray beams with energies ranging from 20 to 160 kVp (typically 30–140 kVp), with a maximum beam current of 2 mA (typically 0.1–2 mA). The spectrum of the X-ray tube was also estimated using the gamma- or X-ray detectors listed in Table 1 using the Monte Carlo simulation. The HPGe detector exhibited the highest energy resolution compared to other semiconductors or scintillation detectors.

2.2. Measurement of HPGe detector characteristics

HPGe is the only radiation detection medium that provides sufficient information to accurately and reliably identify radionuclides from their mono-energy gamma-rays. HPGe detectors have almost 20 to 30 times more energy resolution than sodium iodide detectors. The detector used in this study was a GMX series HPGe coaxial detector system (GAMMA-X) which is a coaxial Germanium (Ge) detector with an ultra-thin entrance window. While most coaxial detectors have entrance windows from 500- to 1000- μm thick, the entrance window of the GAMMA-X detector is a 0.3- μm -thick and ion-implanted contact which extends the lower range of useful energies to around 3 keV. The characteristics of the HPGe detector are presented in Table 2.

One of the limitations of the HPGe detectors is the loss of sensitivity or efficiency when they are exposed to high flux radiation. To address this issue in the experimental setting, a collimator is used to reduce the photon flux, which is normally made from lead or tungsten. The diameter of the hole created on the lead surface should meet two specific criteria: (1) the diameter aperture is sufficiently large to supply the detector with an acceptable flux of photon. (2) The diameter of the aperture is small enough that the photon flux does not saturate the

Fig. 1. A) The X-ray tube and the lead block to render a point source. B) HPGe coaxial detector model number GMX40P4-83. C) X-ray tube, lead block, and HPGe detector under test condition.

Fig. 2. A) Absolute efficiency curves as a function of photon energy and B) the escape fraction of X-rays.

detector. One of the practical issues when working with HPGe is that the radiation source should be in the form of a point source in order to avoid the early saturation of the HPGe detector. In this regard, a lead block with a thickness of 7 cm and a hole with a diameter of 1.5 mm was placed against the x-ray tube beam to render a parallel and/or needle-shaped beam as illustrated in Fig. 1.

To obtain the final X-ray tube spectrum based on the signals acquired from the HPGe detector, certain corrections/modifications pertinent to the material and geometry of the detector should be made. These modifications include I. correction for the counts related to the characteristic X generated by the photoelectric interactions, II. Correction for the absolute efficiency of the detector depending on various factors such as the size, material, and radiation energy, and III. Correction for the

background radiation. Fig. 2 shows the escape fraction curve at different energies and the absolute efficiency of the HPGe detector obtained from the algorithm proposed in (EftekhariZadeh et al., 2014).

2.3. X-ray spectral analysis using IPeM78

A wide range of computational software has been developed for extraction and estimation of the X-ray spectrum through the selection of different filters as well as source and beam properties such as the half-value layer, energy, mAs, and exposure levels. In this study, we utilized the models of Birch and Marshall (IPeM Report 78) (Tucker et al., 1991) which employs an XCOM database to estimate linear attenuation coefficients for numerous materials as well as wide ranges of radiology and mammography X-ray spectra (Chen et al., 2012). The spectra for tungsten target at the tube voltages varying from 30 kV to 150 kV and target angles from 6° to 22° were generated for this study. The entire spectra were generated at an energy step/bin of 0.5 keV.

2.4. X-ray spectral analysis using Spektr

The Spektr toolset, based on the TASMIP algorithm of Boone and Seibert (Punnoose et al., 2016), provides realistic X-ray spectra across a wide range of kVp and beam filtration, wherein a databank for atomic number elements ($Z = 1-92$) and combinations of materials from the National Institute of Standards and Technology (NIST) is exploited. This toolset was used to generate X-ray beams in the energy range of 1–150

Table 3

Gaussian energy broadening of some radiation detector.

Crystal	Energy range (MeV)	Fitting parameter (a)	Fitting parameter (b)	Fitting parameter (c)
BGO	0.059 keV-1.33 MeV	1.56E-03	1.216E-01	2.455
LYSO	20 keV- 1.43 MeV	-6.387E-03	1.097E-01	6.574E-01
NaI	0.062keV-1.35 MeV	-2.4E-03	5.165E-02	2.858
HPGe	40 keV-1.46 MeV	5.868E-04	3.951E-04	7.467

Fig. 3. Schematic geometry of the simulated HPGe detector.

kV with 1 kV energy steps.

2.5. Monte Carlo simulations

The modeling of various inorganic scintillators and semiconductor detectors was performed using the MCNP software. There are three different methods for evaluation of the detector efficiency, namely the semi-empirical, Monte Carlo simulation, and the direct mathematical methods. In this study, the Monte Carlo simulation was employed to evaluate the response of the different detectors wherein the same simulation settings such as positioning, distances, and sources were utilized. Gaussian Energy Broadening (GEB) of all radiation detectors according to Table 3, were incorporated in the MC simulation (Taheri et al., 2016; Askari et al., 2018a; Ordóñez et al., 2019; EftekhariZadeh et al., 2014). The detector efficiency was determined using F8 tally, which is specific for pulse height analysis without any variance reduction. The geometry sought to simulate the detectors was a cylinder with a size of 7.62 cm (height) $\times$ 7.62 cm (diameter). The absolute detector efficiency was calculated by the pulse-height distribution per photon emitted from an X-ray source with kVp = 140 (typical CT scan X-ray spectrum).

The statistical uncertainty was less than 1% in the entire MC simulations. The energy cutoff was set to 0.5 and 0.01 MeV for electrons and photons, respectively. In order to achieve sufficient statistical accuracy, single energy source simulations were performed using ¹³⁷Cs and ²⁴¹Am sources. Other sources, namely ⁶⁰Co and ¹⁵²Eu, were employed for multi-energy simulations. Total history numbers of 108 and 109 were considered for single and multi-energy simulations, respectively, to minimize the statistical uncertainties. No variance reduction method was used and the relative uncertainty did not exceed 0.6%. Since the objective is to detect the peak energies of 57.5 and 69 keV (K-edge related to the Anode (Tungsten)), the X-ray cut-off energy was set to 0.01 MeV within the simulations. This cut-off energy has no significant impact on the resulting spectra and improved the statistical uncertainties. Fig. 3 illustrates the schematic sketch of the simulated HPGe detector.

2.6. Detector efficiency

Detection efficiency is one of the key parameters in spectrometry and X-ray imaging systems which is specified by the intrinsic and geometry efficiency. Intrinsic efficiency is the probability that an incident photon in the detector will produce a meaningful pulse/signal. Geometry efficiency is the fraction of arrived photons on the detector surface per number of emitted photons from a source. The product of the intrinsic and geometry efficiencies gives absolute efficiency which depends on

photon energy and source to detector distance. The detector efficiency can be determined via the count rate measurement of a calibrated source (with known energy spectrum) or Monte Carlo (MC) simulations.

Normally the entire emitted radiations would not arrive at the detector's surface and all of the incident radiations would not interact in the sensitive volume depending on the detector sensitivity and distance from the source. Detector efficiency is divided into two intrinsic and geometric components. Detector efficiency usually depends on material density, sensitive volume size, radiation type, radiation energy, and connected electronic systems (Boson et al., 2008; Olivares et al., 2017).

The intrinsic efficiency (Eq. (1)) is defined as the number of recorded photons in the detector per number of incident photons to the surface of the detector.

$$\epsilon_{\text{int}} = \frac{\text{number of recorded on detector}}{\text{number of incident on detector}} \quad (1)$$

The geometry efficiency (ϵ_g) only depends on the detector size and the distance from the source. This also depends on the solid angle of the detector seen from the actual source position which is obtained from Eq. (2) (Knoll, 2010):

$$\epsilon_g = \frac{1}{2} \left(1 - \frac{d}{\sqrt{d^2 + R^2}} \right) \quad (2)$$

Where d is the distance between source and detector and R is the detector radius. If the number of recorded events on the active volume is calculated per number of photons emitted by the source, a new parameter could be defined as absolute efficiencies (ϵ_{abs}) formulated in Eq. (3) (Knoll, 2010).

$$\epsilon_{\text{abs}} = \epsilon_{\text{int}} \times \epsilon_g \quad (3)$$

2.7. Energy resolution

Radiation spectrometry can also be examined by recording the response to a monoenergetic source of radiation. This procedure can be repeated for different levels of energy. In this regard, the energy resolution is defined as the ability of the detector to accurately determine the energy of the incident radiation (Knoll, 2010). Solid-state semiconductor detectors can have an energy resolution of less than 1%, while scintillation detectors used in gamma-ray spectroscopy normally exhibit an energy resolution of 3–10% (Moszyński et al., 2016).

The energy resolution of a detection system is determined by a combination of noise contribution from electronic devices, variation in the efficiency of charge collection, and the inherent stochastic behavior of charge carriers. In Monte Carlo MCNP simulation code, to better model a radiation detector and energy distribution, Gaussian energy broadening (GEB) (which is a special treatment for tallies) is employed. This tool broadens the distribution of energy through sampling from a Gaussian function (Eq. (4)). The Gaussian width is defined by full width at half maximum (FWHM) in Eq. (5), which is determined by Eq. (6) using a set of user-defined constants (a, b, and c in Eq. (6)). The constant parameters used for simulation of the different detectors were obtained from (EftekhariZadeh et al., 2014).

The response function of a detector in a laboratory setting is normally expressed in the form of a Gaussian function. The Gaussian energy broadening, which is the key parameter to accurately simulate and assess the different detector performance, is formulated in Eq. (4).

$$f(E) = C e^{-\left(\frac{E-E_0}{A}\right)^2} \quad (4)$$

Here, E is the broadened energy, E_0 is the energy of the source, C is a normalization constant, and A is a factor related to the full width at half maximum (FWHM) of the Gaussian function calculated by Eq. (5).

$$A = \frac{FWHM}{2\sqrt{\ln 2}} \quad (5)$$

Fig. 4. The spectrum of the X-ray tube at the energy of 140 keV recorded by the HPGe detector.

The FWHM in Eq. (5) is calculated from Eq. (6) wherein a , b , and c are the specific parameters of the GEB, and E is the energy recorded in the detector. These parameters are shown in Table 3 for some well-known detector examples.

$$FWHM = a + b\sqrt{E + cE^2} \quad (6)$$

The energy resolution of the detector is defined as the FWHM divided by the photo-peak centroid H_{max} . The energy resolution (R) is thus calculated from Eq. 7

$$R\% = \frac{FWHM}{H_{max}} \quad (7)$$

2.8. Dual-energy CT

Dual-energy CT (DECT) is one of the commonly used techniques in the medicine and industry for the separation/decomposition of the underlying elements of the material. The primary goal of the DECT scan is the decomposition and/or discrimination of the different elements or biological tissues. To this end, it is necessary to create two spectra with different energies that are able to render maximum differentiation between the different materials or tissues. In DECT imaging, two images are obtained with two different X-ray spectra, which can be generated by modifying the X-ray tube voltages or altering the X-ray spectrum using external filters. In this research, different spectra have been created by applying external filters to the X-ray beam.

The X-ray source (GMX-160) used in this experiment was able to produce X-ray spectra ranging from 20 kVp to 160 kVp. To create the maximum separation between the two X-ray spectra for the purpose of dual-CT imaging, aluminum filters with optimal thicknesses of 0.5–4.5 mm were employed to attenuate the original X-ray beam. These filters (thicknesses of the aluminum filter) have been optimized through Monte Carlo simulation using the NIST (National Institute of Standard and Technology) library.

3. Results

In order to record the spectrum of the X-ray with an HPGe detector, the different channels connected to the detector must first be calibrated

Fig. 5. The pulse height distribution of different solid-state radiation detectors for 140 kVp X-ray radiation obtained from MC simulation.

with known energies. Regarding the energy range generated by the X-ray source, an Americium-241 source was employed with γ -ray energies of 59.5 (35.9%) keV for the primary radiation and 13.9 keV (13%), 17.6 keV (20.2%), and 26.4 keV (5.2%) for the secondary's. The distance between the source and the detector was set about 2 m to avoid saturation of the detector due to high photon flux and maintaining good statistics at the same time. After the calibration of the HPGe energy channels, the X-ray tube was set at an energy of 140 kVp, a current of 0.1 mA, and an exposure time of 10 s. The spectrum of the resulted X-ray beam was recorded by the calibrated HPGe detector. The calculated pulse height distribution (the spectrum of the tube at specific energy) for a 140 kVp X-ray source is depicted in Fig. 4.

The Monte Carlo simulation results exhibited superior energy resolution of the HPGe detector over the other detectors as shown in Table 4. Thus, the HPGe detector was selected to experimentally measure the spectrum of the X-ray tube. To validate the spectra obtained from the practical experiment, the measured GEB parameters were utilized in the Monte Carlo simulation (using the MCNPX package). As Fig. 5 shows, the HPGe detector resulted in remarkably higher energy resolution compared to the other detectors. In Fig. 4 two energy peaks of 57.6 and 69.2 are clearly visible/distinguishable in the energy spectrum obtained from the HPGe detector. The simulation results confirmed the experimental measurements where the maximum difference of 5% in energy resolution was observed for the HPGe detector.

Table 4 summarizes the energy resolution and FWHM obtained from the simulation of the different detectors. It should be mentioned that the reported results are corresponding to the pulse height distribution of the detector response function for 140 and 70 kVp X-ray energies calculated by Eq. (8) (H_{max} is the pulse height distribution). Moreover, Supplemental Table 1 presents the results obtained from 120 to 80 kVp X-ray energies.

$$Resolution = \frac{FWHM}{H_{max}} \times 100 \quad (8)$$

Table 4

FWHM and energy resolution for the different detectors measured for an X-ray beam with energies of 140 and 70 kVp.

Detector	FWHM _{140keV} (%)	FWHM _{70keV} (%)	Resolution _{140keV} (%)	Resolution _{70keV} (%)	H _{max}
HPGe	0.55	0.49	9.16	8.16	0.060
NaI	1.4	1.49	23.33	24.83	0.060
LYSO	2.2	2.36	36.66	39.33	0.060
BGO	3.7	3.9	61.66	65	0.060

Fig. 6. Comparison of x-spectrum estimated by MC simulation, experimental measurement, Spektr, and IPEM 78 tools for the X-ray energies of A) 60 kVp B) 80 kVp C) 100 kVp D) 140 kVp.

Fig. 7. The dual-energy X-ray spectra obtained from applying 0.5 mm and 4.5 mm external filters (Al) to the original X-ray beam with the energy of 140 kVp (measured by HPGe detector).

In the field of computed tomography imaging, one of the important applications of X-ray spectroscopy is to perform dosimetry using the Monte Carlo method as well as the analysis of the materials by dual-energy and multi-energy scanning approaches. In this study, HPGe detector was employed to carry out the X-ray spectroscopy experimentally as well as using Monte Carlo, IPEM78, and Spektr simulation platforms. Fig. 6 illustrates the comparison between the spectra measured experimentally by the HPGe detector and by the Monte Carlo, IPEM78, and Spektr modeling software's for X-ray energies of 60, 80, 100, and 140 kVp.

For the task of dual-energy CT scanning, the key issue is to generate two X-ray spectra with a minimum overlap which leads to maximal

discrimination of the underlying component of the subjects under study. Fig. 7 shows the generated dual-energy X-ray spectrum obtained from applying the external filters (aluminum filters with thicknesses of 0.5–4.5 mm) to the original X-ray beam with an energy of 140 kVp. This spectrum has been measured by the HPGe detector which exhibits two separable peaks with minimal tails overlap. Regarding Fig. 6, some differences between the results of experimental and simulation measurements were observed. These discrepancies could be attributed to the differences between nominal and actual properties/characteristics of the sources and detectors such as impurity in the filters or uncertainty in the calibration of the sources. The observed differences between the Monte Carlo simulation and the experimental measurement (reported in Fig. 6) are within the normal ranges reported in the literature. Similar differences between experimental and Monte Carlo-based measurement of the CT spectra have been reported in previous studies (Skrzyński, 2014).

4. Discussion

The Monte Carlo method is a powerful and precise tool for the calculation and analysis of the X- and gamma-ray spectra. These spectra would be used for the tasks of dosimetry, image formation/correction, and material decomposition using dual- or multi-energy CT scanning. This study focused on MC simulation and comparison of the several detectors for the energy ranges commonly used in clinical CT scanning. The pulse high distribution increased when the energy of the X-ray beam increased for the entire detectors up to the energy of 60 keV and then dropped to zero at 140 keV energy. The reason for this observation might be due to the incorporation of the Gaussian Energy Broadening (GEB) into the MC simulation. Regarding the characteristics of the detectors such as density, effective atomic number, and different GEB parameters, there was a particular pulse high spectrum for each detector. However, there would be the same pulse high spectra for all

detectors, if we do not consider the GEB parameters in the MC simulation.

Crystal characteristics, impurity, traps, degradation due to long time usage, and instability of the temperature could be the contributory factors to determine the energy resolution of a detector quantified by FWHM. The results suggest that high purity germanium crystal (HPGe) has a narrow energy FWHM compared to the other crystals measured in the same situations, owing to the physical characteristics of the HPGe detector which is made of semiconductor materials. These types of detectors allow for fast, high energy, and spatial resolution image acquisition due to the small and efficient detection medium. On the other hand, the MC simulation results showed a poor energy resolution (large FWHM) for the Bismuth Germanate (BGO) detector, though this detector has a high density as well as the atomic number which render this detector highly sensitive to the incident radiation. The low light gain of the BGO detector is its major disadvantage which results in poor energy discrimination power. Nevertheless, this detector is widely used in nuclear medicine imaging and dosimetry. The Sodium Iodide (NaI) crystal has relatively good energy resolution (exhibiting relatively small energy FWHM) when using a 140 keV X-ray beam. The NaI crystal is transparent with a relatively high light gain which makes it a scintillator detector with good energy resolution. However, NaI crystal is not suitable for CT scanner because this crystal should be coupled with PMTs.

A key factor in measuring X-ray energy spectra is to establish stable and reproducible experimental conditions such as source to object and object to detector distances, temperature, ambient noise, exposure time, acquisition parameters (mAs and kV), and signal processing. Regarding the separate measurement of the 70 and 140 kV spectra, the pileup effect may be significant when the measurement is performed in a real condition (e.g. dual-energy CT imaging) on the combined spectra. Previous studies showed about 12% differences between separate and combined measurements of the 70 and 140 kV spectra, where mostly occurred within a small overlapped energy range between these two spectra (McCollough et al., 2015). Our observation in this study demonstrated negligible pileup effects on the precise detection of the energy peaks when the two spectra were combined. Though up to 13% differences were observed when the energy spectra were combined, this alteration in energy detection only occurred in a small energy range where the two spectra overlapped. Owing to the high energy resolution of the pure germanium detectors, the peak energy detection did not change significantly when the two energy spectra were combined.

In dual-energy CT imaging, selection of the energy peaks plays a key role in the identification of the underlying components of mediums/tissues. The selection of the energy peaks depends on 1) type of study/application such as human, animal, or industry 2) underlying components of the medium/tissue which determine the cross-section or probability of photoelectric or Compton interactions 3) k-edge of the components, and 4) differences between effective atomic number and electron density of tissues/materials at different energies (or different CT numbers). Energy peaks of 70 and 140 kV are of special interest in clinical application since they are commonly used in mammography and whole-body CT imaging. These energy peaks enable maximum differentiation of the underlying component of the biological tissues for a wide range of clinical studies. In the future study, dual-energy CT imaging would be employed for the determination of the proton stopping power for biological tissues. Dual-energy CT imaging is regarded as a reliable technique for estimation attenuation coefficient map in proton radiation therapy. In this regard, energy peaks of 70 and 140 kV were selected to enable accurate estimation of the proton stopping power (providing maximum differentiation of the underlying components) for biological tissues (Mohan and Grosshans, 2017; Royce and Efstathiou, 2019).

The simulated X-ray spectra were compared with spectra generated by Spektr, IPeM78, and practical measurement at the same kVp. The experimental energy spectrum measurement exhibited good agreement with the Monte Carlo simulation results which demonstrates the

Fig. 8. Comparison of different X-ray spectrum obtained from the different models for an X-ray beam with energy of 100 kVp.

superior capability of the HPGe detector to resolve the significant properties of the X-ray energy spectrum. Fig. 8 shows the comparison between the X-ray spectrum obtained from different experimental measurements and simulation models including the simulation and experimental estimation conducted by Ay et al. (2004), Expectation-Maximization modeling by Duan et al. (2011), PENELOPE simulation by Jia et al. (2009), practical measurement by Primak et al. (2010), the model developed by Punnoose et al. (2016) based on Spektr version 3.0 platform. The graphs in Fig. 8 exhibited good agreement between the X-ray spectra observed in this study and the different analytical models and experimental measurements reported in the literature. The small differences in these graphs were due to the different types of filters used for the X-ray beam and the different detector materials.

5. Conclusion

In the present work, Monte Carlo simulations of some commonly used detectors, namely HPGe, BGO, LYSO, and NaI were carried out to calculate the detector efficiency and energy resolution of these detectors for the photon energy ranging from 20 keV to 140 keV. Given the superior energy resolution of the HPGe detector, this detector was selected to practically measure the energy spectrum of a mini-CT scanner in order to generate a dual-energy X-ray beam. The practical spectrum obtained from the HPGe detector was in close agreement with the Monte Carlo simulation. External attenuating filters (0.5 and 4.5 mm Al) were applied to the X-ray beam which resulted in a dual-energy X-ray spectrum with minimal overlap and two distinct energy peaks.

Declaration of competing interest

The authors declare that they have no known competing financial interests or personal relationships that could have appeared to influence the work reported in this paper.

Appendix A. Supplementary data

Supplementary data to this article can be found online at <https://doi.org/10.1016/j.radphyschem.2021.109666>.

Author statement

Study concept and design: **Vahid Lohrabian, Hossein Arabi and Alireza Kamali-Asl**; analysis and interpretation of data: **Vahid Lohrabian, Hossein Ghadiri Harvani and Hossein Arabi**; drafting of the manuscript: **Alireza Kamali-Asl and Habib Zaidi**; critical revision of

the manuscript for important intellectual content: **Habib Zaidi, Vahid Lohrabian, Seyed Rashid Hosseini Aghdam and Hossein Arabi**; statistical analysis: **Vahid Lohrabian and Hossein Arabi**.

References

- Alvarez, R.E., Seibert, J.A., Thompson, S.K., 2004 Mar. Comparison of dual energy detector system performance. *Med. Phys.* 31 (3), 556–565.
- Arabi, H., Asl, A.R., 2010. Feasibility study of a new approach for reducing of partial volume averaging artifact in CT scanner. In: 2010 17th Iranian Conference of Biomedical Engineering (ICBME). IEEE, pp. 1–4.
- Arabi, H., Zaidi, H., 2020 Dec. Applications of artificial intelligence and deep learning in molecular imaging and radiotherapy. *Eur. J. Hybrid Img.* 4 (1), 1–23.
- Arabi, H., Zaidi, H., 2021 Feb 10. Deep learning-based metal artefact reduction in PET/CT imaging. *Eur. Radiol.* 1–3.
- Arabi, H., Asl, A.R., Ay, M.R., Zaidi, H., 2011 Mar. Novel detector design for reducing intercell x-ray cross-talk in the variable resolution x-ray CT scanner: a Monte Carlo study. *Med. Phys.* 38 (3), 1389–1396.
- Arabi, H., Asl, A.R., Ay, M.R., Zaidi, H., 2015 Jul 1. Monte Carlo-based assessment of the trade-off between spatial resolution, field-of-view and scattered radiation in the variable resolution X-ray CT scanner. *Phys. Med. Biol.* 31 (5), 510–516.
- Arabi H, Kamaliasl A, Aghamiri SM. The Effect of Focal Spot Size on the Spatial Resolution of Variable Resolution X-Ray CT Scanner.
- Askari, M., Taheri, A., Sasanpour, M.T., 2018 Jan. Energy broadening model for BGO array crystal from 0.059 to 1.332 MeV. *Instrum. Exp. Tech.* 61 (1), 9–12.
- Askari, M., Taheri, A., Sasanpour, M.T., 2018 Jan. Energy broadening model for BGO array crystal from 0.059 to 1.332 MeV. *Instrum. Exp. Tech.* 61 (1), 9–12.
- Åslund, M., Fredenberg, E., Telman, M., Danielsson, M., 2010 Apr 1. Detectors for the future of X-ray imaging. *Radiat. Protect. Dosim.* 139 (1–3), 327–333.
- Ay, M.R., Shahriari, M., Sarkar, S., Adib, M., Zaidi, H., 2004 Oct 8. Monte Carlo simulation of x-ray spectra in diagnostic radiology and mammography using MCNP4C. *Phys. Med. Biol.* 49 (21), 4897.
- Azbouche, A., Belgaid, M., Mazrou, H., 2015 Aug 1. Monte Carlo calculations of the HPGe detector efficiency for radioactivity measurement of large volume environmental samples. *J. Environ. Radioact.* 146, 119–124.
- Boson, J., Ågren, G., Johansson, L., 2008 Mar 21. A detailed investigation of HPGe detector response for improved Monte Carlo efficiency calculations. *Nucl. Instrum. Methods Phys. Res. Sect. A Accel. Spectrom. Detect. Assoc. Equip.* 587 (2–3), 304–314.
- Bugby, S.L., Jambi, L.K., Lees, J.E., 2016 Sep 22. A comparison of CsI: Tl and GOS in a scintillator-CCD detector for nuclear medicine imaging. *J. Instrum.* 11 (9), P09009.
- Bushberg, J.T., Boone, J.M., 2011 Dec 20. The Essential Physics of Medical Imaging. Lippincott Williams & Wilkins.
- Chen, S.C., Jong, W.L., Harun, A.Z., 2012 Jul. Evaluation of x-ray beam quality based on measurements and estimations using SpekCalc and IPEM78 models. *Malays. J. Med. Sci.* 19 (3), 22.
- Clark, D.P., Holbrook, M., Badea, C.T., 2018 Mar 9. Multi-energy CT decomposition using convolutional neural networks. In: *Medical Imaging 2018: Physics of Medical Imaging*, 10573, 105731O. International Society for Optics and Photonics.
- Duan, X., Wang, J., Yu, L., Leng, S., McCollough, C.H., 2011 Feb. CT scanner x-ray spectrum estimation from transmission measurements. *Med. Phys.* 38 (2), 993–997.
- Eftekhari Zadeh, E., Feghhi, S.A., Bayat, E., Roshani, G.H., 2014. Gaussian energy broadening function of an HPGe detector in the range of 40 keV to 1.46 MeV. *Journal of Experimental Physics* 2014.
- Fuchs, T., Keßling, P., Firsching, M., Nachtrab, F., Scholz, G., 2013. Industrial applications of dual X-ray energy computed tomography (2X-CT). In: *Nondestructive Testing of Materials and Structures*. Springer, Dordrecht, pp. 97–103.
- Gholamiankhah, F., Mostafapour, S., Goushbolagh, N.A., Shojaerazavi, S., Layegh, P., Tabatabaei, S.M., Arabi, H., 2021. Automated Lung Segmentation from CT Images of Normal and COVID-19 Pneumonia Patients arXiv preprint arXiv:2104.02042, Apr 5.
- Ghorbanzadeh, M., Hosseini, S.A., Vahdat, B.V., Mirzaei, H., Akhavanalaf, A., Arabi, H., 2021. Quantitative and Qualitative Performance Evaluation of Commercial Metal Artifact Reduction Methods: Dosimetric Effects on the Treatment Planning arXiv preprint arXiv:2105.09845, May 20.
- Goo, H.W., Goo, J.M., 2017 Jul. Dual-energy CT: new horizon in medical imaging. *Korean J. Radiol.* 18 (4), 555.
- Goodell, J.J., Egnatuk, C.M., Padgett, S.W., Bandong, B.B., Roberts, K.E., Mignerey, A.C., 2017 Dec. Validation of a Monte Carlo HPGe detector model against irradiated foil gamma-ray spectroscopy measurements. *J. Radioanal. Nucl. Chem.* 314 (3), 1793–1802.
- Goorley, T., James, M., Booth, T., Brown, F., Bull, J., Cox, L.J., Durkee, J., Elson, J., Fensin, M., Forster, R.A., Hendricks, J., 2012 Dec 1. Initial MCNP6 release overview. *Nucl. Technol.* 180 (3), 298–315.
- Hossain, I., Sharip, N., Viswanathan, K.K., 2012 Jan 9. Efficiency and resolution of HPGe and NaI (TI) detectors using gamma-ray spectroscopy. *Sci. Res. Essays* 7 (1), 86–89.
- Jeřkovský, M., Javorník, A., Breier, R., Slučiak, J., Povinec, P.P., 2019 Dec. Experimental and Monte Carlo determination of HPGe detector efficiency. *J. Radioanal. Nucl. Chem.* 322 (3), 1863–1869.
- Jia, P., Xie, Y., Bao, S., 2009. Monte Carlo simulation of x-ray tube spectra with PENELOPE. September 7–12 InWorld Congress on Medical Physics and Biomedical Engineering. Springer, Berlin, Heidelberg, pp. 503–506. Munich, Germany 2009.
- Johnson, T.R., Krauss, B., Sedlmair, M., Grasruck, M., Bruder, H., Morhard, D., Fink, C., Weckbach, S., Lenhard, M., Schmidt, B., Flohr, T., 2007 Jun. Material differentiation by dual energy CT: initial experience. *Eur. Radiol.* 17 (6), 1510–1517.
- Kamali-Asl, A., Arabi, H., Tamhidi, S., 2009. Optimization of magnification in a VRX CT scanner. In: *InWorld Congress on Medical Physics and Biomedical Engineering*, September 7–12, 2009, Munich, Germany. Springer, Berlin, Heidelberg, pp. 266–269.
- Knoll, G.F., 2010 Aug 16. *Radiation Detection and Measurement*. John Wiley & Sons.
- Lohrabian, V., Kamali-Asl, A., Arabi, H., Mamashi, F., Hemmati, H.R., Zaidi, H., 2019 Sep 1. Design and construction of a variable resolution cone-beam small animal mini-CT prototype for in vivo studies. *Radiat. Phys. Chem.* 162, 199–207.
- McCollough, C.H., Leng, S., Yu, L., Fletcher, J.G., 2015 Sep. Dual-and multi-energy CT: principles, technical approaches, and clinical applications. *Radiology* 276 (3), 637–653.
- Mohan, R., Grosshans, D., 2017 Jan 15. Proton therapy—present and future. *Adv. Drug Deliv. Rev.* 109, 26–44.
- Moszyński, M., Syntfeld-Kazuch, A., Swiderski, L., Grodzicka, M., Iwanowska, J., Słobczyński, P., Szcześniak, T., 2016 Jan 1. Energy resolution of scintillation detectors. *Nucl. Instrum. Methods Phys. Res. Sect. A Accel. Spectrom. Detect. Assoc. Equip.* 805, 25–35.
- Olivares, D.M., Guevara, M.M., Velasco, F.G., 2017 Dec 1. Determination of the HPGe detector efficiency in measurements of radioactivity in extended environmental samples. *Appl. Radiat. Isot.* 130, 34–42.
- Ordóñez, J., Gallardo, S., Ortiz, J., Sáez-Muñoz, M., Martorell, S., 2019 Feb 1. Intercomparison of full energy peak efficiency curves for an HPGe detector using MCNP6 and GEANT4. *Radiat. Phys. Chem.* 155, 248–251.
- Panth, R., Liu, J., Abt, I., Liu, X., Schulz, O., Wei, W.Z., Mei, H., Mei, D.M., Wang, G.J., 2020 Jul. Characterization of high-purity germanium detectors with amorphous germanium contacts in cryogenic liquids. *Eur. Phys. J. C* 80 (7), 1–1.
- Perez-Andujar, A., Pibida, L., 2004 Jan 1. Performance of CdTe, HPGe and NaI (TI) detectors for radioactivity measurements. *Appl. Radiat. Isot.* 60 (1), 41–47.
- Peyvandi, R.G., Taheri, A., Lehdarboni, M.A., 2015 Jun 11. Evaluation of a new position sensitive detector based on the plastic rod scintillators. *J. Instrum.* 10 (6), T06004.
- Pfeiffer, D., Pfeiffer, F., Rummeny, E., 2020. Advanced X-ray imaging Technology. In: *Molecular Imaging in Oncology*. Springer, Cham, pp. 3–30.
- Primak, A.N., Giraldo, J.C., Eusemann, C.D., Schmidt, B., Kantor, B., Fletcher, J.G., McCollough, C.H., 2010 Nov. Dual-source dual-energy CT with additional tin filtration: dose and image quality evaluation in phantoms and in vivo. *Am. J. Roentgenol.* 195 (5), 1164–1174.
- Punnonen, J., Xu, J., Sienega, A., Zbijewski, W., Siewerdsen, J.H., 2016 Aug. Spektr 3.0—a computational tool for x-ray spectrum modeling and analysis. *Med. Phys.* 43 (8Part1), 4711–4717.
- Rebuffel, V., Dinten, J.M., 2007 Oct 1. Dual-energy X-ray imaging: benefits and limits. *Insight-non-destructive testing and condition monitoring*, 49, pp. 589–594, 10.
- Royce, T.J., Efsthathiou, J.A., 2019 Sep 1. Proton therapy for prostate cancer: a review of the rationale, evidence, and current state. In: *Urologic Oncology: Seminars and Original Investigations*, 37. Elsevier, pp. 628–636, 9.
- Ryzhikov, V.D., Opolonin, A.D., Pashko, P.V., Svishch, V.M., Volkov, V.G., Lysetskaya, E. K., Kozin, D.N., Smith, C., 2005. Instruments and detectors on the base of scintillator crystals ZnSe (Te), CWO, CsI (TI) for systems of security and customs inspection systems. *Nucl. Instrum. Methods Phys. Res. Sect. A Accel. Spectrom. Detect. Assoc. Equip.* 537 (1–2), 424–430.
- Sanaat, A., Arabi, H., Ay, M.R., Zaidi, H., 2020 Feb 13. Novel preclinical PET geometrical concept using a monolithic scintillator crystal offering concurrent enhancement in spatial resolution and detection sensitivity: a simulation study. *Phys. Med. Biol.* 65 (4), 045013.
- Sanghavi, P.S., Jankharia, B.G., 2019 Jul. Applications of dual energy CT in clinical practice: a pictorial essay. *Indian J. Radiol. Imag.* 29 (3), 289.
- Skrzyński, W., 2014 Nov 1. Measurement-based model of a wide-bore CT scanner for Monte Carlo dosimetric calculations with GMCTdospp software. *Phys. Med. Biol.* 30 (7), 816–821.
- Taheri, A., Lehdarboni, M.A., Gholipour, R., 2016 May 25. Determination of Gaussian energy broadening parameters for organic scintillators. *J. Instrum.* 11 (5), P05020.
- Tucker, D.M., Barnes, G.T., Chakraborty, D.P., 1991 Mar. Semiempirical model for generating tungsten target x-ray spectra. *Med. Phys.* 18 (2), 211–218.
- Van Sciver, W., Hofstadter, R., 1951 Dec 1. Scintillations in thallium-activated Ca I 2 and CsI. *Phys. Rev.* 84 (5), 1062.