University of Groningen

Comparison of Gas Chromatography Quadrupole Time-Of-Flight and Quadrupole Orbitrap Mass Spectrometry in Anti-doping Analysis

Abushareeda, Wadha; Tienstra, Marc; Lommen, Arjen; Blokland, Marco; Sterk, Saskia; Kraiem, Suhail; Horvatovich, Peter; Nielen, Michel; Al-Maadheed, Muhammad; Georgakopoulos, Costas

Published in:

Rapid Communications in Mass Spectrometry

DOI:

[10.1002/rcm.8281](https://doi.org/10.1002/rcm.8281)

IMPORTANT NOTE: You are advised to consult the publisher's version (publisher's PDF) if you wish to cite from it. Please check the document version below.

Document Version

Final author's version (accepted by publisher, after peer review)

Publication date:

2018

[Link to publication in University of Groningen/UMCG research database](#)

Citation for published version (APA):

Abushareeda, W., Tienstra, M., Lommen, A., Blokland, M., Sterk, S., Kraiem, S., Horvatovich, P., Nielen, M., Al-Maadheed, M., & Georgakopoulos, C. (2018). Comparison of Gas Chromatography Quadrupole Time-Of-Flight and Quadrupole Orbitrap Mass Spectrometry in Anti-doping Analysis: I. Detection of Anabolic-androgenic Steroids. *Rapid Communications in Mass Spectrometry*, 32(23), 2055-2064. <https://doi.org/10.1002/rcm.8281>

Copyright

Other than for strictly personal use, it is not permitted to download or to forward/distribute the text or part of it without the consent of the author(s) and/or copyright holder(s), unless the work is under an open content license (like Creative Commons).

The publication may also be distributed here under the terms of Article 25fa of the Dutch Copyright Act, indicated by the "Taverne" license. More information can be found on the University of Groningen website: <https://www.rug.nl/library/open-access/self-archiving-pure/taverne-amendment>.

Take-down policy

If you believe that this document breaches copyright please contact us providing details, and we will remove access to the work immediately and investigate your claim.

Comparison of Gas Chromatography Quadrupole Time-Of-Flight and Quadrupole Orbitrap Mass Spectrometry in Anti-doping Analysis: I. Detection of Anabolic-androgenic Steroids.

Wadha Abushareeda ¹, Marc Tienstra ², Arjen Lommen ², Marco Blokland ², Saskia Sterk ², Suhail Kraiem ¹, Peter Horvatovich ³, Michel Nielen ², Muhammad Al-Maadheed ¹, Costas Georgakopoulos ¹

1 Anti-Doping Lab Qatar, Sports City Road, P.O. Box 27775, Sports City, Doha, Qatar

2 RIKILT, Wageningen University, P.O. Box 230, 6700 AE Wageningen, Netherlands

3 University of Groningen, P.O. Box 196, 9700 AD Groningen, Netherlands

ABSTRACT

RATIONALE: World Anti-doping Agency (WADA) encourages drug testing laboratories to develop screening methods that can detect as many doping substances as possible in urine. The use of full scan high-resolution acquisition (FS/HR) with GC/MS for the detection of known and unknown TMS derivatives of AAS provides anti-doping testing bodies with a new analytical tool.

METHODS: The AAS extracted from urine samples by generic liquid-liquid extraction, after enzymatic hydrolysis, and trimethylsilyl (TMS) derivatization step. The extracted urine analysed by GC/Q-TOF and GC/Q-Orbitrap to compare the performance of both instruments for the detection of 46 AAS in human urine. The quantitation of endogenous anabolic steroids and the ability of the two analytical platforms to comply with the requirements for testing as part of the WADA Athlete Biological Passport (ABP) was also assessed.

RESULTS: Data presented shows that the analytical performance is in compliance with the WADA specifications for both instruments. The LOD(s) for both instruments are well below the 50% MRPL sensitivity level. The mass errors in the current study for the GC/Q-Orbitrap platform are lower compared to the respective of the GC/Q-TOF.

CONCLUSIONS: The data presented herein proved that both molecular profiling platforms can be used for antidoping screening. The mass accuracies are excellent in both instruments, however the GC/Q-Orbitrap shows superior due to higher resolution compared to GC/Q-TOF.

INTRODUCTION

The World Anti-doping Agency (WADA) Prohibited List [1] comprises prohibited classes of drugs with pharmacological activity. The small molecules included on the list are detected using mass spectrometry (MS) coupled with either liquid chromatography (LC) or gas chromatography (GC). The criteria used to select the method of analysis are determined mainly by the sensitivity, specificity and matrix effects that fulfil the specifications of the WADA International Standard for Laboratories (ISL) [2] and the WADA Minimum Required Performance Levels (MRPL) technical document TD MRPL [3]. The molecules analysed by GC/MS are usually those with chemical structures that result in low ionization efficiency with LC/MS [4].

In the sports anti-doping field, GC/MS screening typically utilizes low-resolution GC triple quadrupole technology (GC/QQQ) [5]; however, a few methods have been published that utilize full-scan (FS) and high-resolution (HR) acquisition modes. In 2007, a time-of-flight (TOF) screening method was published [6] that could be used for the analysis of synthetic anabolic-androgenic steroids (AAS). However, the WADA specifications in 2007 were less demanding than the contemporary regulations. In another study, a limited screening method based on two-dimensional (2D) GC coupled to time-of-flight mass spectrometry was published that could be used for selected anabolic agents [7]. In another, more comprehensive study from the same research group, a screening method based on quadrupole-Orbitrap GC/MS was published for use in the detection of a large number of substances in urine, such as synthetic AAS, β -agonists, stimulants, narcotics, metabolic modulators and diuretics, that fulfilled the WADA sensitivity specifications [8]. Finally, in another recent study [9], a routine screening method that used GC/MS with a triple quadrupole instrument was adapted for use with a QTOF GC/MS instrument and subjected to full validation, in which the FS/HR MS acquisition mode was applied to detect and quantify exogenous and endogenous steroids [10].

However, FS/HR GC/MS technologies are more widely in other analytical fields than in anti-doping analysis. A recent review that focused on the use of GC/MS with TOF mass analysers describes their use for the analysis of a large number of organic contaminants and residues present at trace levels for environmental, food safety and biological applications [11]. A recent example of the use of FS/HR GC/MS in the food analysis field combined atmospheric pressure chemical ionization (APCI) with the use of a QTOF mass analyser for the analysis of the volatile components of olive oil [12].

The use of GC/MS coupled with TOF and QTOF mass analyser technologies and two-dimensional GC (GC×GC) has been recently reviewed [13]. The author observed that the popularity of GC×GC is increasing and the number of components that can be simultaneously detected is limited by characteristics of the MS system such as the dynamic range, resolution and scanning speed, and eluting GC×GC peaks. Moreover, several recent applications of this method to toxicology, clinical chemistry, food and environmental analysis has been recently reviewed [13]. In toxicology, the volatile organic compounds (VOCs) produced during the early stages of bodily decomposition can be detected using thermal desorption coupled to GC/GC/TOF MS, which could be used for a wide variety of forensic or epidemiologic purposes, such as during natural catastrophes with large numbers of collapsed buildings [14]. In clinical chemistry, the quantitative analysis of organic acids in urine has been performed using GC/GC/TOF MS to determine abnormal patterns that can indicate the presence of inherited disorders of organic acid metabolism [15]. In food analysis, GC/GC TOF MS has been applied to the investigation of dioxin-like micropollutants in animal-derived food matrices [16]. In environmental analysis, the profiling of short- and medium-chain chlorinated paraffins in sediment and fish samples using GC/GC/TOF MS with negative ionization has been demonstrated [17].

Along with molecular profiling using TOF mass analysers, the use of the GC/Q-Orbitrap has been recently developed. The use of HR/FS GC/Orbitrap MS with electron ionization was evaluated for use in pesticide residue analysis in fruit and vegetables [18]. In food analysis, GC and LC coupled to Orbitrap MS has been applied to the identification of substances that have migrated from nano-films in food packaging [19].

The advantages of the HR/FS-MS method for anti-doping screening analysis have been described by Friedmann *et al* [20]. The HR mass filter results in enhanced specificity and sensitivity and the reduction of background noise, and full scan acquisition allows for the detection of a virtually unlimited number of analytes, which can be identified by reprocessing of the acquired data files. Using FS/HR-GC/MS for the identification of designer drugs [21,22], and as well as novel metabolites that allow for the prolonged detection of substances with “zero tolerance,” such as AAS [23,24] has been demonstrated.

Synthetic AAS are prohibited [1] and present a challenge [3] for anti-doping screening laboratories, since their presence should still be detected at low concentrations for a long period of time after administration of banned substances. Moreover, endogenous AAS [10],

such as testosterone and prohormones, are administered exogenously to enhance endogenous AAS levels. Because there is no difference in the mass spectra produced by exogenously administered AAS and their endogenously produced counterparts, an indirect methodology based on the ratio of AAS metabolites is currently used to identify possible abuse of substances. To alleviate these challenges, the WADA introduced the steroid profile and WADA Steroidal Athlete Biological Passport (ABP), which contains the quantified results of endogenous AAS screening analyses conducted during the career of the professional athlete for the following parameters: Testosterone (T), Epitestosterone (E), Androsterone (A), Etiocholanolone (Etio), 5 α -androstan-3 α ,17 β -diol (5 α adiol), 5 β -androstan-3 α ,17 β -diol (5 β adiol) and the following ratios: T/E, A/Etio, 5 α adiol/5 β adiol, A/T, 5 α adiol /E, for each analyzed sample [10]. The concentrations of the markers, in addition to the ratios of AAS metabolites, are also used in the identification of possible abuse [25].

The goal of this study is the comparison of two HR/FS-GC/MS molecular screening technologies, which are based on the Agilent GC/Q-TOF and the Thermo GC/Q-Orbitrap, using the same sample aliquots and the same GC parameters and conditions for both instruments. Both the QTOF and Orbitrap mass analysers allow for high mass resolution, but the Orbitrap mass analyser generally produces higher-resolution spectra at a lower scanning speed and higher dynamic range, depending on the applied acquisition parameters, due its increased ion trapping ability. We present an analytical platform comparison study that uses these two MS analysers for the detection of anabolic steroids and that includes the qualitative screening of synthetic AAS, the quantitative profiling of endogenous AAS and the reprocessing of the electronic data files for preventive analysis. In this article, the performance of the GC/QTOF and GC/Q-Orbitrap in the detection and quantitation of exogenous and endogenous AAS in the same urine samples is compared. The analysis comprises a limited number of analytes, including representative exogenous AAS and all of the endogenous AAS that are present in the steroidal ABP.

EXPERIMENTAL

Chemicals and Reagents

Sodium hydrogen carbonate and diethyl ether were supplied by Merck (Darmstadt, Germany). Methanol (HPLC grade), 2-propanethiol, di-potassium hydrogen phosphate, potassium dihydrogen phosphate, ammonium iodide and sodium bicarbonate were supplied

by Sigma Aldrich (Darmstadt, Germany). β -Glucuronidase derived from *Escherichia coli* (*E. coli*) was supplied by Roche (Mannheim, Germany). MSTFA (N-methyl-N-(trimethylsilyl) trifluoroacetamide) was supplied by Carl Bucher (Waldstetten, Germany).

Reference Materials

The following internal standards (ISTD) were purchased from LGC (Wesel, Germany): D5-etiocholanolone (d5 Etio), D4-androsterone glucuronide (d4A Glu), D3-testosterone (d3T), D3-epitestosterone (d3E), and D5-5 β -androstane-3 α -17 β -diol (D5-5 β Diol). The remaining reference standard materials were purchased from LGC (Wesel, Germany), TRC (Toronto, Canada), Sigma Aldrich (Darmstadt, Germany), Steraloids (Newport, USA), and Cerilliant (Round Rock, USA). Standard stock solutions of the analytes were individually prepared in methanol. For validation purposes, standard working solutions containing the analytes were prepared in methanol by subsequent dilution of the stock solutions. The analytes from the endogenous steroid profile were diluted in a different working solution. All of the solutions were stored at -20°C in amber vials.

Sample preparation

The sample preparation was described in detail for a previous study [9]. In brief, it included a liquid-liquid extraction with diethyl ether at pH 9-10 and a desalting step, which is an approach that is commonly used to extract doping substances from the urine matrix. A clean extract was obtained after the separation of the organic layer from the aqueous phase following centrifugation and the cooling of the extraction tubes in an ethanol basin at -80°C. Prior to the extraction step, deconjugation of the Phase II steroid glucuronide conjugates was performed by enzymatic hydrolysis using the *E. coli*-derived β -glucuronidase as indicated in [10]. The final step of the sample preparation was the derivatization of the extracts with trimethyl silane (TMS), which was performed in a mixture of MSTFA, ammonium iodide and propanethiol; in these conditions, both the hydroxyl and the keto steroidal groups were derivatized [26] by the TMS group.

Instruments and analytical conditions

GC/EI/quadrupole time-of-flight analysis conditions

The GC/MS system used in the current study was an GC 7890 coupled with a 7200 Q-TOF mass spectrometer (G3850-64101; Agilent, Delaware, USA) equipped with a BPX5 5%

phenyl polysilphenylene-siloxane capillary column (30 m length, 0.25 mm ID, 0.1 μ m film thickness; SGE, Victoria, Australia) and a back-flush system. The combination of the quadrupole device with the TOF MS analyser provides the capability to conduct MS/MS experiments. Helium was used as a carrier gas at a constant flow of 1.1 mL/min. The injection port temperature was set to 280 °C, and the injection volume was set to 2 μ L with a split ratio of 20:1. The oven temperature was initially set at 160 °C, increased at 10 °C/min to 200 °C, then increased at 2 °C/min to 220 °C, followed by an increase at 6 °C/min to 292 °C and 50 °C/min to 310 °C, where it was held for 3 min. The analysis run time was 29.36 minutes. The interface temperature was set to 280 °C and the ion source was set to 250 °C. Electron ionization (EI) of the compound ionization was performed using 70 eV of electron energy. A 2 GHz extended dynamic range (EDR) acquisition mode was used for the TOF data acquisition. The acquisition rate was 5 spectra per sec (200 msec per spectrum), and the number of transients per spectrum was set to 2718. GC/Q-TOF has the capacity to acquire MS data in FS/HR mode with a mass accuracy of <5 ppm mass error in EI mode, depending on the concentration of the analytes. The applied MS range (m/z 80-670) allows for the measurement of small molecules typically analysed with GC/QTOF. Automated mass calibration was performed after every three injections using perfluorotributylamine (PFTBA, Agilent).

GC-EI-quadrupole Orbitrap analysis conditions

The second GC/MS system used in the current study was a GC/Q-Orbitrap (Q Exactive GC, Thermo Scientific, Bremen, Germany), equipped with an SGE BPX5 column (30 m length, 0.25 mm ID, 0.1 μ m film thickness). This system consisted of a TriPlus RSH autosampler, a TRACE 1310 GC with a hot split/splitless injector, an EI source, and a hybrid quadrupole Orbitrap (Q Exactive) mass spectrometer with HCD (higher energy collision-induced dissociation). The sample introduction was performed using the TriPlus RSH autosampler. Helium was used as carrier gas with a constant flow set at 1.1 mL/min. The same analysis conditions for the GC parameters described in the previous section (GC/EI/Quadrupole time-of-flight analysis conditions) was used. EI at 70 eV was used for the compound ionization with a source temperature set at 250 °C. Full scan acquisition mode was applied with a mass range of m/z 80-670 with 1 μ sec scans, a resolving power of 60,000 at m/z 200 and an AGC target set at $3 \cdot 10^6$. The mass calibration procedure was performed before each acquisition batch using PFTBA. The internal mass calibration during the measurement was conducted

using three different background ions that originated from the column bleed (m/z 207.03236/ $C_2H_{15}O_3Si_3^+$, 281.05115/ $C_7H_{21}O_4Si_4^+$ and 355.06994/ $C_9H_{27}O_5Si_5^+$) with a search window of ± 2 ppm.

Qualitative analysis

The validation of the screening method on both instruments was performed according to ISL guidelines [2]. In this procedure the following parameters were evaluated and validated: the identification capability, the specificity and the limit of detection. To assess the retention time for each compound, a mixture of standards at a high concentration (10-fold of the MRPL) was injected into both instruments. The evaluation of the compound identification capability based on the retention time was performed using 10 different blank urine samples that were spiked with a standard mixture of 46 synthetic anabolic steroids at concentrations of 50% of the MRPL [3]. To evaluate the specificity of the developed methods, 10 blank urine samples were analysed in order to demonstrate the absence of any interference. The limit of detection (LOD) was determined by spiking the urine samples with a mixture of standards at 10%, 25%, 50% and 100% of the MRPL. The LOD is the lowest concentration at which a compound can be detected with sufficient data points and the elimination of background interference. The validation results showed that the analytical methods that were used in this study to detect AAS in urine are in compliance with the criteria laid out in the WADA ISL guidelines.

Quantitative analysis

The method used for quantitative compound profiling was validated on both instruments. The validation process was carried out over a period of three days. The linearity, precision, accuracy and combined measurement uncertainty were evaluated. The calibration curves for quantification purposes were generated from stripped urine samples, which were blank urine samples collected from female children and depleted of endogenous steroids using C18 SPE extraction and collection of the eluent to avoid any interference with the endogenous steroids in the urine matrix, that were spiked with the standards. The calibration curves were obtained by measuring the levels of endogenous AAS included in the steroid profile of the ABP at seven different concentrations of standards; concentrations of 2-400 ng/mL were used for T and E, 100-8000 ng/mL for A and Etio, and 4-800 ng/mL for 5α adiol and 5β adiol. The ratio

of the peak height (m/z) that was obtained for the analyte to the peak height of the internal standard was calculated and plotted against the concentration of the added analyte. Linearity was determined by using the weighed linear regression model ($W=1/X$, where X is the concentration of the analyte). The precision and accuracy of the method were determined using a level of spiked quality control (QC) samples that corresponded to level 5 of the calibration curve, which were prepared in four different aliquots. The analysis was performed using 4 replicates for each level for each day ($n=4$) over a period of three days ($n=12$). The intermediate precision was determined from the data obtained from the QC samples collected during the 3 days of the experiments. Both the intermediate precision and accuracy (% bias) that were calculated based on the QC samples were used to estimate the combined measurement uncertainty (MU) for each steroid according to the following equation:

$$\text{Combined } MU = \sqrt{\text{intermediate precision}^2 + \text{Bias}^2}$$

RESULTS AND DISCUSSION

The current study was conducted using the same samples and chromatographic conditions in two different laboratories: Anti-Doping Lab Qatar (ADLQ) in Qatar (GC/QTOF) and RIKILT-Institute of Food Safety (GC/Q-Orbitrap) in the Netherlands. The same sample set was analysed on both analytical platforms within two weeks period in order to reduce possible variability due to sample stability. The methods applied herein were based on that of a previous study [9], with the exception of the GC/Q-Orbitrap MS method, which is described in the Materials and Methods section. The entire sample preparation procedure was conducted at ADLQ, where the aliquots were subsequently analysed using QTOF.

Afterwards, the aliquots were safely sent to RIKILT at room temperature and reanalysed using the GC/Q-Orbitrap instrument with the same chromatographic conditions, including identical injection port liner, chromatographic column and temperature parameters. The only difference in the two chromatographic systems was the inclusion of a back-flush device installed on the QTOF instrument. In this study, representative AAS were selected for analysis. AAS are considered to be drugs with anabolic pharmacological activity that is highly beneficial for the enhancement of athletic performance, even long after administration. Consequently, AAS are required to be detected by anti-doping laboratories at the lowest possible concentrations, even below the values specified by WADA [3]. To this end, AAS were analysed at concentrations close to their limit of detection (LOD). The measured AAS

analytes included endogenous AAS that are part of the ABP SP, AAS analytes that are easily detected by GC analysis, as well as analytes that are problematic for GC analysis due to matrix interferences.

The MS tuning and calibration procedures were applied to both instruments according to the manufacturers' specifications. The full width at half maximum (FWHM) resolution was set to 60,000 (specified at m/z 200) during the tuning procedure for the Q/Orbitrap and to 12,000 (specified at m/z 272) for the QTOF. These conditions, in addition to the conditions described in the instruments and analytical conditions section, provided sufficient data points to identify and quantify the analyte peaks in the FS-HR-GC/MS data; there were 20-30 and 30-50 data points across the chromatographic peaks for the Q/Orbitrap and the QTOF data, respectively. For the GC/Q-Orbitrap, the mass accuracy reached 2 ppm error due to the use of lock mass recalibration with ions (m/z 207.03236, 281.05115, and 355.06994) that originated from the column bleed. The GC/QTOF method did not include a lock mass procedure during acquisition; however, the mass calibration procedure was repeated after three sample injections.

Qualitative analysis results

Table 1 summarizes the qualitative parameters used to validate the AAS screening conducted using both instruments. The GC/MS method provided sufficient data quality to discriminate between the MS signals derived from each of the investigated analytes using both instruments. Figure 1 shows the comparison of the mass accuracies for the molecular ions and other fragment ions produced by the analysed AAS from Table 1, which were spiked at 2.5 ng/mL as a function of the m/z . Table 2 summarizes the reproducibility of the mass accuracy for the GC/QTOF and GC/Q-Orbitrap analyses of 10 different urine samples based on the base peaks of the selected AAS. Overall, the mass errors for the GC/Q-Orbitrap platform in the current study are lower than those for the GC/QTOF system.

The identification capability parameters [2], which are based on the detectability of the AAS within the different urine matrices at 50% of the MRPL concentration [3], are shown in Table 1. The LOD for each AAS is also shown in Table 1.

Figure 2 shows the extracted ion chromatograms for the main fragment ions derived from the following AAS compounds: 1-testosterone, 5 β -androst-1-ene-17 β -ol-3-one (boldenone m1),

3'OH stanozolol, epimetendiol, bolasterone, 9 α -fluoro-18-nor-17,17-dimethyl-4,13-diene-11 β -ol-3-one (flouxymesterone-18nor), 13 β ,17 α -diethyl-5 β -gonane-3 α ,17 β -diol (norbolethone m2), 19-norandrosterone (nandrolone m1), oxymesterone and clenbuterol. An improvement in the detection of 3'OH stanozolol using GC/Q-Orbitrap was noted, which was due to the lower level of matrix interference it had compared to that of GC/QTOF. Moreover, clenbuterol was detected unambiguously using GC/Q-Orbitrap with the FS-HR mode at 0.1 ng/ml; it was only possible to detect this compound in MS/MS mode using GC/QTOF [9]. The detection of other compounds, such as boldenone, bolasterone and 13 β ,17 α -diethyl-5 α -gonane-3 α ,17 β -diol, was subject to interference using either instrument.

Generally, the chromatographic data obtained from routine screening using GC and LC/MS is independently evaluated by two analysts in order to identify suspicious ions that may be the result of a prohibited substances abuse. These suspicious signals activate specific confirmatory procedures that include testing of a new aliquot of the suspicious urine sample. A routine batch of samples analysed by GC/MS or LC/MS methods may generate tens of thousands of ion chromatograms that have to be reviewed one by one. Consequently, the quality and reproducibility of the ion chromatograms to be reviewed is important in reducing errors during the evaluation of doping tests. In an HR instrument, the mass accuracy of the ion chromatograms influences the quality of their evaluation. If the mass extraction window is too wide, then the matrix interference is enhanced; if it is too narrow, then there is an increased risk of not detecting a prohibited substance due to mass error. In the current study, the mass extraction windows were optimized and set to 20 ppm for GC/QTOF and 5 ppm for GC/Q-Orbitrap, and the same m/z values were used for the detection of the analytes. The Orbitrap method can be further optimized via the use of the chromatographic back-flush device and the inclusion of MS/MS acquisition, both of which were applied only to the GC/QTOF analysis in the current study. Both the GC/QTOF and the GC/Q-Orbitrap instruments were equipped with a quadrupole device that allowed for the acquisition of MS/MS.

Quantitative analysis results

The development of a new GC/MS anti-doping screening method that cannot measure endogenous AAS has no practical utility [10, 25]. Therefore, the current study included screening of the six endogenous AAS that are listed in Table 3 for both instruments. The data

presented in Table 3 demonstrates that the analytical performance of both instruments is in compliance with the WADA specifications [10]. It can be seen that the MU is less than 20% for A and Etio, 25% for 5 α adiol and 5 β adiol and less than 20% for T and E. The endogenous concentrations of androsterone and etiocholanolone may be high relative to those of the other AAS (e.g., 8 μ g/mL), so to avoid detector saturation, the full scan MS mode was not used for QTOF, but instead the transition of m/z from 434.3031 to 419.2796 was retraced at lower abundances using the MS/MS data. The GC/Q-Orbitrap method was conducted in FS mode during testing of all of the steroids. It should be noted, however, that the Orbitrap platform is less prone to detector saturation than QTOF because the Orbitrap is an ion trapping instrument that can adapt to higher analyte concentrations by reducing the ion acquisition time via automatic gain control and, thereby, allow for the detection of a larger dynamic concentration range than QTOF. Overall, the data acquired from both instruments was in compliance with the specifications given in the technical document TD2016EAAS [10]. In another study, a comparison of steroid profile data obtained using GC/QTOF and GC/QQQ was previously made [9].

Analysis of a proficiency test sample

The analysis of a proficiency testing sample was also successfully conducted using both instruments. The sample was obtained from an individual to whom nandrolone was administered; the two main metabolites of nandrolone, 19-norandrosterone and 19-noretiocholanolone, were detected. Figure 3 presents the FS-HR spectrum of 19-norandrosterone obtained at a concentration of 6.2 ng/mL using both the GC/QTOF and GC/Q-Orbitrap platforms.

CONCLUSIONS

The goal of the current study was the comparison of the performance of the GC/QTOF and GC/Q-Orbitrap molecular profiling technologies during routine anti-doping screening and the assessment of the qualitative and quantitative analyses resulting from the use of both platforms. The data presented herein demonstrate that both molecular profiling platforms can be used for anti-doping screening. It was also shown that it is possible to combine FS-HR acquisition with MS/MS to enhance specificity in order to facilitate screening for AAS. The

LODs for both instruments were well below 50% of the MRPL sensitivity level. The identification criteria data shows that a limited number of compounds were not able to reach 100% detectability in the various urine matrices. The mass accuracies were excellent for both instruments; however, the GC/Q-Orbitrap had superior mass accuracy due to its higher resolution. The superiority of the GC/Q-Orbitrap was also demonstrated by the lower matrix effects found in the urine samples in relation to the mass accuracies. Overall, both instruments proved to be sufficiently robust for routine anti-doping testing and analysis. The results of the reprocessing of the existing data files to identify substances that previously escaped detection in FS mode will be presented in a future report.

Acknowledgements

The authors are very grateful to the Qatar National Research Fund for funding this project (contract NPRP 6-334-3-087). Mrs. Noor Al-Motawa, Director of the ADLQ Education and Research Office, is sincerely acknowledged for her constant support. Dominic Roberts and Paul Silcock (TFS, Runcorn, UK) are also acknowledged for their technical support, stimulating discussions and facilitation of the use of the Q-Exactive GC instrument at RIKILT.

REFERENCES

1. World Anti-doping Agency (WADA) Prohibited List (2018). https://www.wada-ama.org/sites/default/files/prohibited_list_2018_en.pdf. Accessed April 25, 2018.
2. World Anti-Doping Agency(WADA) International Standard for Laboratories(ISL 2016). https://www.wada-ama.org/sites/default/files/resources/files/isl_june_2016.pdf. Accessed October 14, 2017.
3. World Anti-doping Agency (WADA) TD2018MRPL. https://www.wada-ama.org/sites/default/files/resources/files/td2018mrpl_v1_finaleng.pdf. Accessed April 25, 2018.
4. Thevis M and Schanzer W. Mass spectrometry in sports drug testing: structure characterization and analytical assays. *Mass Spectrometry Reviews*. 2007;26:79-107.
5. De Brabanter N, Van Gansbeke W, Geldof L, Van Eenoo P. An improved gas chromatography screening method for doping substances using triple quadrupole

- mass spectrometry, with an emphasis on quality assurance. *Biom. Chrom.* 2012; 26: 1416–1435
6. Georgakopoulos C, Vonaparti A, Stamou M, Kiouisi P, Lyris E, Angelis YS, Tsoupras G, Wuest B, Nielen MWF, Panderi I, Koupparis M. Preventive doping control analysis: liquid and gas chromatography time-of-flight mass spectrometry for detection of designer steroids. *Rapid Commun. Mass Spectrom.* 2007; 21: 2439–2446.
 7. Silva AI, Pereira HMG, Casilli A, Conceição FC, Aquino Neto FR. Analytical challenges in doping control: Comprehensive two-dimensional gas chromatography with time of flight mass spectrometry, a promising option. *Journal of Chromatography A.* 2009; 1216(14): 2913-2922.
 8. Cavalcanti GA, Rodrigues LM, Santos L, Zheng X, Gujar A, Cole J, Padilha MC, Aquino Neto FR. Non-targeted acquisition strategy for screening doping compounds based on GC-EI-hybrid quadrupole-Orbitrap mass spectrometry: A focus on exogenous anabolic steroids. *Drug Test Anal.* 2017; 1–11.
 9. Abushareeda W, Lyris E, Kraiem S, Wahaibi A, Alyazidi S, Dbes N, Lommen A, Nielen M, Horvatovich PL, Alsayrafi M, Georgakopoulos C. Gas chromatographic quadrupole time-of-flight full scan high resolution mass spectrometric screening of human urine in antidoping analysis. *Journal of Chromatography B.* 2017; 1063: 74–83.
 10. World Anti-Doping Agency. Endogenous Anabolic Androgenic Steroids: Measurement and Reporting. TD2016EAAS, ver. 1.0. <https://www.wada-ama.org/en/resources/science-medicine/td2016-eaas> . Accessed October 14, 2017.
 11. Hernández F, T. Portolés T, Pitarch E, López FJ. Gas chromatography coupled to high-resolution time-of-flight mass spectrometry to analyze trace-level organic compounds in the environment, food safety and toxicology. *Trends in Analytical Chemistry.* 2011; 30: 388-400.
 12. Sales C, Cervera MI, Gil R, Portolés T, Pitarch E, Beltran J. Quality classification of Spanish olive oils by untargeted gas chromatography coupled to hybrid quadrupole-time of flight mass spectrometry with atmospheric pressure chemical ionization and metabolomics-based statistical approach. *Food Chemistry.* 2017; 216: 365-373.
 13. Tranchida PQ, Franchina FA, Dugo P, Mondello L. Comprehensive two-dimensional gas chromatography-mass spectrometry: recent evolution and current trends. *Mass Spectrometry Reviews.* 2016; 35: 524-534.

14. Agapiou A , Zorba E , Mikić K, McGregor L, Spiliopoulou C, Statheropoulos M. Analysis of volatile organic compounds released from the decay of surrogate human models simulating victims of collapsed buildings by thermal desorption–comprehensive two-dimensional gas chromatography–time of flight mass spectrometry. *Analytica Chimica Acta*. 2015; 883: 99-108.
15. Sweetman L, Ashcraft P, Bennett-Firmin J. Quantitative Organic Acids in Urine by Two Dimensional Gas Chromatography-Time of Flight Mass Spectrometry (GCxGC-TOFMS). Clinical Applications of Mass Spectrometry in Biomolecular Analysis. *Methods in Molecular Biology*. 2016; 1378: 183-197.
16. Planche C, Ratel J, Mercier F, Blinet P, Debrauwer L, Engel E. Assessment of comprehensive two-dimensional gas chromatography-time-of-flight mass spectrometry based methods for investigating 206 dioxin-like micropollutants in animal-derived food matrices. *Journal of Chromatography A*. 2015; 1392: 74-81.
17. Xia D, Gao L, Zheng M, Tian Q, Huang H, Qiao L. A Novel Method for Profiling and Quantifying Short- and Medium-Chain Chlorinated Paraffins in Environmental Samples Using Comprehensive Two-Dimensional Gas Chromatography–Electron Capture Negative Ionization High-Resolution Time-of-Flight Mass Spectrometry. *Environmental Science Technology*. 2016; 50: 7601-7609.
18. Mol HGJ, Tienstra M, Zomer P. Evaluation of gas chromatography – electron ionization – full scan high resolution Orbitrap mass spectrometry for pesticide residue analysis. *Analytica Chimica Acta*. 2016; 935: 161-172.
19. Martínez-Bueno MJ, Hernando MD, Uclés S, Rajski L, Cimmino S, Fernández-Alba AR. Identification of non-intentionally added substances in food packaging nano films by gas and liquid chromatography coupled to orbitrap mass spectrometry. *Talanta*. 2017; 172: 68-77.
20. Friedmann T, Flenker U, Georgakopoulos C, Alsayrafi M, Sottas P, Williams S, Gill R. Evolving concepts and techniques for anti-doping. *Bioanalysis* . 2012 ; 4(13) : 1667–1680.
21. Abushareeda W, Fragkaki A, Vonaparti A, Angelis Y, Tsivou M, Saad K, Kraiem S, Lyrís E, Alsayrafi M, Georgakopoulos C. Advances in the detection of designer steroids in anti-doping. *Bioanalysis*. 2014 ; 6 (6) : 881-896
22. Gomez C, Fabregat A, Pozo ÓJ, Marcos J, Segura J, Ventura R. Analytical strategies based on mass spectrometric techniques for the study of steroid metabolism. *Trends in Analytical Chemistry*. 2014; 53: 106–116.

23. Sobolevsky T, Rodchenkov G, Detection and mass spectrometric characterization of novel long-term dehydrochloromethyltestosterone metabolites in human urine. *J. Steroid Biochem. Mol. Biol.* 2012; 128: 121– 127
24. Guddat S, Fußhöller G, Beuck S, et *al.* Synthesis, characterization, and detection of new oxandrolone metabolites as long-term markers in sports drug testing. *Anal Bioanal Chem.* 2013; 405(25): 8285–8294.
25. Kuuranne T, Saugy M, Baume N. Confounding factors and genetic polymorphism in the evaluation of individual steroid profiling. *Br J Sports Med.* 2014; 48(10): 848–855.
26. Donike M, Zimmermann J. Preparation of trimethylsilyl, triethylsilyl and tert-butyldimethylsilyl enol ethers from ketosteroids for investigations by gas chromatography and mass spectrometry. *J. Chromatogr.* 1980, 202, 483-486.

Table 1: Summary of the results of the qualitative analysis.

	Analytes (TMS derivatives)	Chemical formula	RT (min)	Ions (<i>m/z</i>)			Spiked conc* (ng/mL)	LOD (ng/mL)		Detection in 10 urine	
								GC/Q-TOF	GC/Q-Orbitrap	GC/Q-TOF	GC/Q-Orbitrap
1	18-Normetenol	C ₂₃ H ₃₈ O ₁ Si ₁	8.7	358.2686	253.1951	216.1873	2.5	1.25	1.25	10	10
2	1-Androstene 3 β ,17 β diol	C ₂₅ H ₄₆ O ₂ Si ₂	14.49	405.2637			2.5	1.25	0.63	9	10
3	1-Testosterone	C ₂₅ H ₄₄ O ₂ Si ₂	15.45	206.1121	194.1121		2.5	1.25	1.25	9	10
4	3 α OH-Tibolone	C ₂₇ H ₄₆ O ₂ Si ₂	16.31	443.2796	353.2295		2.5	0.63	0.63	10	10
5	3 β OH-Tibolone	C ₂₇ H ₄₆ O ₂ Si ₂	15.19	443.2796	353.2295		2.5	0.63	0.63	10	10
6	3'OH Stanozolol m1	C ₃₀ H ₅₆ N ₂ O ₂ Si ₃	24.17	560.3644	545.3409		1	1.00	0.50	7	10
7	17 α -Methyl-5 α -androstane-3 α ,17 β -diol	C ₂₆ H ₅₀ O ₂ Si ₂	15.69	450.3344	270.2342	255.2107	1	1.00	1.00	8	10
8	17 α -Methyl-5 β -androstane-3 α ,17 β -diol	C ₂₆ H ₅₀ O ₂ Si ₂	15.87	450.3344	270.2342	255.2107	1	1.00	1.00	8	10
9	6-Oxo-androstenedine	C ₂₈ H ₄₈ O ₃ Si ₃	18.34	516.2906	501.2671		2.5	0.63	0.63	10	10
10	Bolasterone met (7 α ,17 α -dimethyl-5 β -androstane-3 α ,17 β -diol)	C ₂₇ H ₅₂ O ₂ Si ₂	17.11	374.2999	284.2499	269.2264	2.5	1.25	1.25	10	10
11	Boldenone met (5 β -androst-1-ene-17 β -ol-3-one)	C ₂₅ H ₄₄ O ₂ Si ₂	12.43	194.1121	206.1121		2.5	1.25	0.30	10	10
12	Calusterone met (7 β ,17 α -dimethyl-5 β -androstane-3 α ,17 β -diol)	C ₂₇ H ₅₂ O ₂ Si ₂	16.8	374.2999	284.2499	269.2264	2.5	2.50	2.50	10	9
13	Calusterone	C ₂₇ H ₄₈ O ₂ Si ₂	18.56	460.3187	445.2953		2.5			10	10
14	Clenbuterol	C ₁₈ H ₃₄ N ₂ Cl ₂ O Si ₂	6.15	300.1007	335.06890		0.1	0.10	0.05	7	10

15	Clostebol met (4-chloroandrost-4-en-3 α -ol-17-one)	C ₂₅ H ₄₃ O ₂ ClSi ₂	17.49	466.2485	468.2587		2.5	0.63	0.63	10	10
16	Danazol met(Ethisterone)	C ₂₇ H ₄₄ O ₂ Si ₂	18.64	456.2874	441.2640		2.5	1.25	0.63	9	10
17	Drostanolone PC	C ₂₆ H ₄₈ O ₂ Si ₂	16.75	448.3187			2.5	0.63	0.63	10	10
18	Drostanolone met (Drostanolone 3ol17one)	C ₂₆ H ₄₈ O ₂ Si ₂	14.26	448.3187			2.5	0.63	0.63	9	10
19	Epimetendiol	C ₂₆ H ₄₈ O ₂ Si ₂	12.55	358.2686	343.2452	448.3187	1	1.00	0.25	9	10
20	Fluxymesterone met(9 α -fluoro-18-nor-17,17-dimethyl-4,13-diene-11 β -ol-3-one)	C ₂₆ H ₄₃ O ₂ F ₁ Si ₂	15.35	462.2780	447.2545	357.2076	2.5	1.25	1.25	10	10
21	Formebolone met(Dealdehyde-formebolone)	C ₂₉ H ₅₄ O ₃ Si ₃	19.14	534.3375	339.2170		2.5	0.63	0.63	10	10
22	Formestane		19.25	518.3062	503.2828		20			10	10
23	Furazabol	C ₂₃ H ₃₈ N ₂ O ₂ Si ₁	21.54	402.2697	387.2462		2.5	2.50	0.63	9	10
24	Furazanol met(16 β -hydroxyfurazabol)	C ₂₆ H ₄₆ N ₂ O ₃ Si ₂	24.35	490.3042	218.1153		2.5	1.25	0.63	10	10
25	Mesterolone met(1 α -methyl-5 α -androstane-3 α -ol-17-one)	C ₂₆ H ₄₈ O ₂ Si ₂	15.47	448.3187	433.2953		2.5	0.63	0.63	10	10
26	Mesterolone PC	C ₂₆ H ₄₈ O ₂ Si ₂	16.3	141.0730	433.2953		2.5	1.25	1.25	10	10
27	Methasterone met(2 α ,17 α -dimethyl-5 α -androstane-3 α ,17 β -diol)	C ₂₇ H ₅₂ O ₂ Si ₂	16.24	449.3266	374.2999		2.5	2.50	2.50	10	10
28	Methasterone PC	C ₂₇ H ₅₀ O ₂ Si ₂	18.3	462.3343	419.2796	332.2530	2.5		0.63	9	9
29	Methenolone met (3 α -hydroxy-1-methylen-5 α -androstane-17-one)	C ₂₆ H ₄₆ O ₂ Si ₂	15.06	446.3031	431.2796	251.1826	2.5	1.25	0.63	8	9

30	Mibolerone	C ₂₆ H ₄₆ O ₂ Si ₂	17.78	446.3031	431.2796	301.1982	2.5	1.25	0.63	10	10
31	19-Norandrosterone	C ₂₄ H ₄₄ O ₂ Si ₂	11.73	405.2640	315.2139		1	0.25	0.25	9	10
32	19-Noretiocholanolone	C ₂₄ H ₄₄ O ₂ Si ₂	13.13	405.2640	315.2139		2.5	0.63	0.63	10	10
33	Norbolethone m1(13β,17α-diethyl-5α- gonane-3α,17β-diol)	C ₂₇ H ₅₂ O ₂ Si ₂	18.28	144.0965	157.1034		2.5	1.25		8	10
34	Norbolethone m2(13β,17α-diethyl-5β- gonane-3α,17β-diol)	C ₂₇ H ₅₂ O ₂ Si ₂	19	144.0965	157.1034		2.5	1.25		10	10
35	Norclostebol	C ₁₈ H ₂₅ ClO ₂	20.15	452.2328	417.2640		2.5	0.63	0.63	10	10
36	Norethandrolone m2(17α-Ethyl-5β- estrane-3α,17β-diol)	C ₂₆ H ₅₀ O ₂ Si ₂	17.52	241.1951	331.2452		2.5		0.63	10	10
37	Oxabolone PC	C ₂₇ H ₄₈ O ₃ Si ₃	18.67	506.3062	507.31405		2.5	0.63	0.63	9	10
38	Oxymesterone	C ₂₉ H ₅₄ O ₃ Si ₃	20.81	534.3375	519.31405	389.2327	2.5	0.63	0.63	10	10
39	Stenbolone	C ₂₆ H ₄₆ O ₂ Si ₂	16.23	220.1278	208.1278		2.5	1.25	0.63	9	10
40	Testosterone	C ₂₅ H ₄₄ O ₂ Si ₂	16.83	432.2874	209.1357	417.2640	EAAS				
41	5α-Androstan-3α,17β- diol	C ₂₅ H ₄₈ O ₂ Si ₂	14.21	241.1951	256.2186		EAAS				
42	5B-Androstan-3α,17β- diol	C ₂₅ H ₄₈ O ₂ Si ₂	14.55	241.1951	256.2186		EAAS				
43	5B-Androstan-3α,17β- diol d5, ISTD	C ₂₅ H ₄₃ D ₅ O ₂ Si ₂	14.15	246.2265	261.2499		ISTD				
44	Androsterone	C ₂₅ H ₄₆ O ₂ Si ₂	13.8	434.3031	329.2295	419.2796	EAAS				
45	Androsterone d4, ISTD	C ₂₅ H ₄₂ D ₄ O ₂ Si ₂	13.5	438.3282	333.25463	423.3047	ISTD				
46	Epitestosterone	C ₂₅ H ₄₄ O ₂ Si ₂	15.99	432.2874	209.1356	417.2640	EAAS				
47	Epitestosterone- D3, ISTD	C ₂₅ H ₄₁ D ₃ O ₂ Si ₂	15.86	435.3063	420.2828		ISTD				

48	Etiocholanolone	C ₂₅ H ₄₆ O ₂ Si ₂	14.11	434.3031	329.2295	419.2796	EAAS
49	Etiocholanolone d5, ISTD	C ₂₅ H ₄₁ D ₅ O ₂ Si ₂	13.78	439.3345	334.2609	424.3109	ISTD

- Spiked concentration at 50% of the MRPL.
EAAS: endogenous androgenic-anabolic steroids
ISTD: internal standard

Table 2: Reproducibility of the mass accuracy in different urine matrices for both instruments.

Instrument	Analytes	exact mass (<i>m/z</i>)	Average (ppm)	STD	max	min
Q-ToF	5 β -androst-1-ene-17 β -ol-3-one	194.1121	3.15	2.96	9.27	0.00
	6-oxo- androstenedine	516.2906	7.62	3.73	13.17	2.32
	1 α -methyl-5 α -androstane-3 α -ol-17-one	433.2953	2.67	1.73	5.77	0.69
	17 α -Ethyl-5 β -estrane-3 α ,17 β -diol	331.2458	5.30	5.72	15.09	0.30
Q-Orbitrap	5 β -androst-1-ene-17 β -ol-3-one	194.1121	0.40	0.23	0.52	0.00
	6-oxo- androstenedine	516.2906	1.01	0.09	1.16	0.97
	1 α -methyl-5 α -androstane-3 α -ol-17-one	433.2953	0.18	0.22	0.46	0.00
	17 α -Ethyl-5 β -estrane-3 α ,17 β -diol	331.2458	1.81	0.48	2.72	1.21

Table 3: Summary of the results of the quantitative data analysis.

Compound	Calibration range (ng/mL)	Bias (%)		Intermediate precision (%)		MU (%)	
		GC/Q- TOF	GC/Q- Orbitrap	GC/Q- TOF	GC/Q- Orbitrap	GC/Q- TOF	GC/Q- Orbitrap
Androsterone (A)	100-8000	15.1	10.9	4.3	4.02	15.7	11.6
Etiocholanolone (ETIO)	100-8000	9.1	5.7	6.2	4.3	11	7.2
5 α -Androstandiol (5 α diol)	4-800	5.8	6.5	0.8	4.7	5.9	7.7
5 β -Androstandiol (5 β diol)	4-800	5.1	6.1	1.9	5.3	5.5	8.1
Testosterone (T)	2-400	9.1	17.2	1.2	1.1	9.2	17.2
Epitestosterone (E)	2-400	8.4	2.1	9	1.5	12.3	2.6

Figure1: Mass errors of representative AAS' diagnostic ions over the entire mass range in both instrument.

Figure 2.

Figure 3.