University of Groningen

New guidelines for delta C-13 measurements

Coplen, TB; Brand, WA; Gehre, M; Groning, M; Meijer, HAJ; Toman, B; Verkouteren, RM;
Coplen, Tyler B.; Brand, Willi A.; Gröning, Manfred

Published in:
Analytical Chemistry

DOI:
[10.1021/ac052027c](https://doi.org/10.1021/ac052027c)

IMPORTANT NOTE: You are advised to consult the publisher's version (publisher's PDF) if you wish to cite from it. Please check the document version below.

Document Version
Publisher's PDF, also known as Version of record

Publication date:
2006

[Link to publication in University of Groningen/UMCG research database](#)

Citation for published version (APA):

Coplen, TB., Brand, WA., Gehre, M., Groning, M., Meijer, HAJ., Toman, B., Verkouteren, RM., Coplen, T. B., Brand, W. A., & Gröning, M. (2006). New guidelines for delta C-13 measurements. *Analytical Chemistry*, 78(7), 2439-2441. <https://doi.org/10.1021/ac052027c>

Copyright

Other than for strictly personal use, it is not permitted to download or to forward/distribute the text or part of it without the consent of the author(s) and/or copyright holder(s), unless the work is under an open content license (like Creative Commons).

The publication may also be distributed here under the terms of Article 25fa of the Dutch Copyright Act, indicated by the "Taverne" license. More information can be found on the University of Groningen website: <https://www.rug.nl/library/open-access/self-archiving-pure/taverne-amendment>.

Take-down policy

If you believe that this document breaches copyright please contact us providing details, and we will remove access to the work immediately and investigate your claim.

Downloaded from the University of Groningen/UMCG research database (Pure): <http://www.rug.nl/research/portal>. For technical reasons the number of authors shown on this cover page is limited to 10 maximum.

New Guidelines for $\delta^{13}\text{C}$ Measurements

Tyler B. Coplen,^{*,†} Willi A. Brand,[‡] Matthias Gehre,[§] Manfred Gröning,[⊥] Harro A. J. Meijer,^{||} Blaza Toman,[#] and R. Michael Verkouteren⁺

U.S. Geological Survey, 431 National Center, Reston, Virginia 20192, Max-Planck-Institute for Biogeochemistry, Beutenberg Campus, P.O. Box 100164, 07701 Jena, Germany, UFZ Umweltforschungszentrum Leipzig-Halle GmbH, Labor für Stabile Isotope, Permoserstrasse 15, 04318 Leipzig, Germany, Isotope Hydrology Laboratory, International Atomic Energy Agency, P.O. Box 100, A-1400 Vienna, Austria, Centrum voor Isotopen Onderzoek (CIO), Rijksuniversiteit Groningen, Nijenborgh 4, 9747 AG Groningen, Netherlands, Information Technology Laboratory, National Institute of Standards and Technology, Mail Stop 8980, Gaithersburg, Maryland 20899, Chemical Science and Technology Laboratory, National Institute of Standards and Technology, Mail Stop 8371, Gaithersburg, Maryland 20899

Consistency of $\delta^{13}\text{C}$ measurements can be improved 39–47% by anchoring the $\delta^{13}\text{C}$ scale with two isotopic reference materials differing substantially in $^{13}\text{C}/^{12}\text{C}$. It is recommended that $\delta^{13}\text{C}$ values of both organic and inorganic materials be measured and expressed relative to VPDB (Vienna Pee Dee belemnite) on a scale normalized by assigning consensus values of -46.6‰ to L-SVEC lithium carbonate and $+1.95\text{‰}$ to NBS 19 calcium carbonate. Uncertainties of other reference material values on this scale are improved by factors up to two or more, and the values of some have been notably shifted: the $\delta^{13}\text{C}$ of NBS 22 oil is -30.03‰ .

Differences in measured isotope amount ratios of stable carbon isotopes ($^{13}\text{C}/^{12}\text{C}$), commonly called $\delta^{13}\text{C}$ values, are used in oceanography, atmospheric sciences, biology, paleoclimatology, geology, environmental sciences, food and drug authentication, and forensic applications. Progress in these fields requires smaller measurement uncertainties to be achieved. Advances require more accurate and intercomparable measurements; for example, a difference of 0.01‰ in $\delta^{13}\text{C}$ of global atmospheric CO_2 can equate to a global carbon balance source/sink discrepancy of $\sim 400\text{ Tg}$ among international monitoring networks.¹ Although advances in instrumentation enable increasingly precise measurements, laboratories measuring the same specimen often disagree by 10 times their reported “uncertainty” of measurement,^{2,3} and agreement has not improved appreciably in the last two decades, except in special cases.⁴

Recognizing that two-point calibrations of the $\delta^2\text{H}$ and $\delta^{18}\text{O}$ scales substantially improved the agreement among laboratories,⁵ the International Atomic Energy Agency (IAEA) convened a consultants meeting in 2004 to review stable carbon isotopic reference materials and to recommend a second reference material for two-point normalization of the $\delta^{13}\text{C}$ scale. Four laboratories (Centrum voor Isotopen Onderzoek, Groningen, Netherlands; Max-Planck-Institute for Biogeochemistry, Jena, Germany; UFZ Leipzig-Halle, Leipzig, Germany; U.S. Geological Survey, Reston, Virginia) performed analytical measurements. Participants at the U.S. National Institute of Science and Technology (NIST) headed the task to estimate consensus means and uncertainties.

EXPERIMENTAL METHODS

Mass Spectrometry. In this study, 1055 state-of-the-art continuous flow elemental-analyzer (EA) mass spectrometry measurements using the general method of Qi et al.⁷ were performed on selected organic and inorganic carbon isotopic reference materials. NBS 19 calcium carbonate was adopted for anchoring at high ^{13}C amount and was assigned the value $+1.95\text{‰}$ relative to VPDB following recommendations of the IAEA and the International Union of Pure and Applied Chemistry (IUPAC).^{8,9}

Estimating Consensus Means Using the Bayesian Method.

In prior intercomparison exercises, consensus—mean estimation was done for each reference material independently using a statistical model that assumes that, for a reference material i and laboratory j , the mean value of the measurement can be decomposed into two components, one specific to the reference material

* Corresponding author. E-mail: tbcoplen@usgs.gov.

[†] U.S. Geological Survey.

[‡] Max-Planck-Institute for Biogeochemistry.

[§] UFZ Umweltforschungszentrum Leipzig-Halle GmbH.

[⊥] International Atomic Energy Agency.

^{||} Rijksuniversiteit Groningen.

[#] Information Technology Laboratory, National Institute of Standards and Technology.

⁺ Chemical Science and Technology Laboratory, National Institute of Standards and Technology.

(1) *Isotope aided studies of atmospheric carbon dioxide and other greenhouse gases, Phase II*; IAEA-TECDOC-1269; International Atomic Energy Agency: Vienna, 2002.

(2) Gonfiantini, R. *Advisory group meeting on stable isotope reference samples for geochemical and hydrological investigations*, Sept. 19–21, 1983, Report to the Director General; International Atomic Energy Agency: Vienna, 1984.

(3) Stichler, W. In *Reference and intercomparison materials for stable isotopes of light elements*; IAEA-TECDOC-825; International Atomic Energy Agency: Vienna, 1995; pp 67–74.

(4) Verkouteren, R. M.; Klindinst, D. B. *Value Assignment and Uncertainty Estimation of Selected Light Stable Isotope Reference Materials: RMs 8543–8545, RMs 8562–8564, and RM 8566*; NIST Special Publication 260-149; National Institute of Standards and Technology: Gaithersburg, Maryland, 2004.

(5) Gonfiantini, R. *Nature* **1978**, 271, 534–536.

(6) Böhlke, J. K.; Coplen, T. B. In *Reference and intercomparison materials for stable isotopes of light elements*; IAEA-TECDOC-825; International Atomic Energy Agency: Vienna, 1995; pp 51–66.

(7) Qi, H.; Coplen, T. B.; Geilmann, H.; Brand, W. A.; Böhlke, J. K. *Rapid Commun. Mass Spectrom.* **2003**, 17, 2483–2487.

(μ_i) and one specific to the laboratory (α_j). Methods such as those described by Rukhin and Vangel¹⁰ were then used to estimate the consensus mean, μ_i , and compute the type A uncertainty¹¹ associated with it. The α_j are called *random effects*, which are given a probability distribution with mean 0 and some variance, σ_L , and they are used to represent added variability in the data due to random differences between the laboratories. The type A uncertainty contains a component associated with the estimation of μ_i and a component that is an estimate of σ_L .

In this study, advantage has been taken of the fact that in all of the data sets, each laboratory has provided measurements on multiple reference materials. Such data makes it possible to deconvolute the mean value of a measurement into the component μ_i and a component α_{ij} . The α_{ij} values are given a distribution with mean λ_j and variance τ_j . Now λ_j represents a systematic laboratory effect, one present across all reference materials. Larger λ_j values represent larger departures of laboratory j from the consensus, whereas smaller λ_j values signify more “universal” departures that occur across most of the reference materials (i.e., α_{ij} values would be similar across materials). The μ_i are again estimated to provide the consensus mean for each reference material. These estimates are more accurate than the “one reference material at a time” estimates because they have been adjusted for systematic laboratory effects. Furthermore, the type A uncertainty associated with these consensus means is smaller than that of the corresponding “one reference material at a time” estimates. In this study, the estimation of the μ_i and their uncertainties—the λ_j and the τ_j values—was done using WinBUGS,^{12,13} a program using Markov Chain Monte Carlo computation with a Bayesian hierarchical model having noninformative priors on all of the hyperparameters.¹⁴

Evaluation of Uncertainty. The strategy for evaluating standard uncertainty was guided by ISO-GUM.¹¹ The multivariate Bayesian determination of consensus means (described above) calculates standard errors about those means. These standard errors are fairly consistent across all materials and compositions, and they may be considered to represent type A standard uncertainties (u_A). For reference materials with $\delta^{13}\text{C}$ values close to +1.95‰ (the realization point of the VPDB scale), the type A uncertainty is a good estimate of the total uncertainty because scale effects approach zero. However, it is well-known that measurement results become less reproducible as the $\delta^{13}\text{C}$ difference between the sample and the VPDB realization point increases, so although the normalization procedure described here substantially improves measurement reproducibility, analytical limitations still impose scale discrepancies. These discrepancies may be considered the result of type B uncertainty, which is

observed in the normalized data (Tables S-1 and S-2). A general and simple model for type B uncertainty was extracted from the normalized data (eq 1), where the proportionality constant k was estimated to be 0.001.

$$u_B = |0.00195 - \delta^{13}\text{C}| \cdot k \quad (1)$$

Combined standard uncertainties (u_C) were then determined through eq 2.

$$u_C = \sqrt{u_A^2 + u_B^2} \quad (2)$$

An expanded uncertainty ($U = 2u_C$) about the recommended value provides an interval that has about a 95% probability of encompassing the true value.

RESULTS

The measurements of $\delta^{13}\text{C}$ values are shown in Tables S-2 and S-3. To normalize $\delta^{13}\text{C}$ measurements, L-SVEC lithium carbonate (NIST RM 8545) was selected as the low- ^{13}C -content scale anchor because EA $\delta^{13}\text{C}$ values of amounts as small as 0.3 mg are statistically identical, and carbonates are easily prepared for analysis using H_3PO_4 . L-SVEC was assigned a $\delta^{13}\text{C}$ consensus value of −46.6‰ on the basis of high-accuracy dual-inlet mass spectrometry measurements (optimized to minimize memory and isotopic fractionation) by Ghosh et al.,¹⁵ who determined a value of $-46.607 \pm 0.057\text{‰}$ (u_C), and by Verkouteren and Klinedinst,⁴ who calculated a dual-inlet value of $-46.57 \pm 0.13\text{‰}$ (u_C) that was based on the relationship between $\delta^{17}\text{O}$ and $\delta^{18}\text{O}$ of 0.528¹⁶ and an $^{17}\text{O}/^{16}\text{O}$ ratio in VSMOW reference water of 0.000 386 913.¹⁷ Recommended $\delta^{13}\text{C}$ values (Table 1) were determined by multivariate Bayesian analysis. Recommended $\delta^{13}\text{C}$ values also were determined for three CO_2 gases (NIST RM 8562, RM 8563, and RM 8564) and three CaCO_3 reference materials (IAEA-CO-1, IAEA-CO-8, and NBS 18) because high-quality data were available.^{4,15}

DISCUSSION

Improvement in data consistency was determined by comparing the variations in newly normalized data to unnormalized data in several studies^{3,4} for all reference materials with values more negative than −25‰. The average variations (standard deviations) in results across laboratories were lowered 39–47%. As a result, uncertainties in value assignments were also improved, for which combined standard uncertainties were lowered by factors of 2–3 for specific materials (Figure 1).

In August 2005 at IUPAC’s 43rd General Assembly in Beijing, the Commission on Isotopic Abundances and Atomic Weights accepted the recommendations of this IAEA panel that $\delta^{13}\text{C}$ values of all carbon-bearing materials be measured and expressed relative

(8) Hut, G. *Consultants’ group meeting on stable isotope reference samples for geochemical and hydrological investigations*, Sept. 16–18, 1985, Report to the Director General; International Atomic Energy Agency: Vienna, 1987.

(9) Coplen, T. B. *Pure Appl. Chem.* **1994**, *66*, 273–276.

(10) Rukhin, A.; Vangel, M. J. *Am. Stat. Assoc.* **1998**, *93*, 303–308.

(11) *Guide to the Expression of Uncertainty in Measurement*; ISO: Geneva; 1993; ISBN: 92-67-10188-9.

(12) The BUGS Project, WinBUGS v. 1.4.1, Oct. 2004; <http://www.mrc-bsu.cam.ac.uk/bugs/welcome.shtml> (accessed Feb 2006); hosted by the MRC Biostatistics Unit, Cambridge, UK.

(13) Congdon, P. *Bayesian Statistical Modelling*; Wiley: New York, 2001; ISBN: 0-471-49600-6.

(14) Gelman, A.; Carlin, J.; Stern, H.; Rubin, D. *Bayesian Data Analysis*, 2nd ed.; Chapman and Hall/CRC: Boca Raton, FL, 2003; ISBN: 0-412-03991-5.

(15) Ghosh, P.; Patecki, M.; Rothe, M.; Brand, W. A. *Rapid Commun. Mass Spectrom.* **2005**, *19*, 1097–1119.

(16) Meijer, H. A. J.; Li, W. J. *Isotopes Environ. Health Stud.* **1998**, *34*, 349–369.

(17) Assonov, S. S.; Brenninkmeijer, C. A. M. *Rapid Commun. Mass Spectrom.* **2003**, *17*, 1017–1029.

(18) Gonfiantini, R.; Stichler, W.; Rozanski, K. In *Reference and intercomparison materials for stable isotopes of light elements*; IAEA-TECDOC-825; International Atomic Energy Agency: Vienna, 1995; pp 13–29.

(19) Stalker, L.; Bryce, A. J.; Andrew, A. S. *Org. Geochem.* **2005**, *36*, 827–837.

Table 1. Reference $\delta^{13}\text{C}$ Values of Stable Carbon Isotopic Reference Materials^{a,b}

name	description	$\delta^{13}\text{C} \times 10^3$
USGS41	L-glutamic acid	+37.63 ^c
IAEA-CO-1	calcium carbonate	+2.49 ^d
NBS 19	calcium carbonate	+1.95
RM 8562	carbon dioxide	-3.72
NBS 18	calcium carbonate	-5.01 ^e
IAEA-CO-8	calcium carbonate	-5.76 ^f
IAEA-CH-6	sucrose	-10.45
RM 8564	carbon dioxide	-10.45
USGS24	graphite	-16.05 ^g
IAEA-CH-3	cellulose	-24.72
USGS40	L-glutamic acid	-26.39 ^h
IAEA-600	caffeine	-27.77
IAEA-601	benzoic acid	-28.81
IAEA-602 ⁱ	benzoic acid	-28.85
NBS 22	oil	-30.03 ^k
IAEA-CH-7	polyethylene	-32.15
RM 8563	carbon dioxide	-41.59 ^j
L-SVEC	lithium carbonate	-46.6
IAEA-CO-9	barium carbonate	-47.32 ^m

^a Recommendations from a consultants meeting of the International Atomic Energy Agency. Analytical data in this study are shown in Tables S-2 and S-3, and estimates of uncertainty are provided in Table S-4 of the Supporting Information. ^b $\delta^{13}\text{C}$ values expressed relative to VPDB ($\delta^{13}\text{C}$ of NBS19 $\equiv +1.95\text{‰}$) and normalized to $\delta^{13}\text{C}$ of L-SVEC $\equiv -46.6\text{‰}$. ^c Qi et al. value⁷ (normalized to $\delta^{13}\text{C}$ of L-SVEC $\equiv -46.6\text{‰}$) of +37.77 ‰ ($u_c = 0.09\text{‰}$) is within 2 u_c of recommended value (Table S-5). ^d Qi et al. value⁷ (normalized to $\delta^{13}\text{C}$ of L-SVEC $\equiv -46.6\text{‰}$) of +2.48 ‰ ($u_c = 0.06\text{‰}$), Stichler Bayesian value³ (normalized to $\delta^{13}\text{C}$ of L-SVEC $\equiv -46.6\text{‰}$) of +2.46 ‰ ($u_c = 0.03\text{‰}$), and Ghosh et al. value¹⁵ (normalized to $\delta^{13}\text{C}$ of L-SVEC $\equiv -46.6\text{‰}$) of +2.49 ‰ ($u_c = 0.03\text{‰}$) are all within 1 u_c of recommended value (Tables S-5–7). ^e Stichler Bayesian value³ (normalized to $\delta^{13}\text{C}$ of L-SVEC $\equiv -46.6\text{‰}$) of -5.78 ‰ ($u_c = 0.03\text{‰}$) and Ghosh et al. value¹⁵ (normalized to $\delta^{13}\text{C}$ of L-SVEC $\equiv -46.6\text{‰}$) of -5.76 ‰ ($u_c = 0.03\text{‰}$) are within 1 u_c of recommended value (Tables S-6 and S-7). ^f Stichler Bayesian value³ (normalized to $\delta^{13}\text{C}$ of L-SVEC $\equiv -46.6\text{‰}$) of -16.06 ‰ ($u_c = 0.05\text{‰}$) is within 1 u_c of recommended value (Table S-6). ^g Qi et al. value⁷ (normalized to $\delta^{13}\text{C}$ of L-SVEC $\equiv -46.6\text{‰}$) of -26.31 ‰ ($u_c = 0.06\text{‰}$) is within 2 u_c of recommended value (Table S-5). ^h IAEA-602 is listed here for information purposes only. It is not recommended for routine carbon isotopic calibration because it is enriched in ^{17}O and ^{18}O , and its $^{17}\text{O}/^{18}\text{O}$ ratio no longer reflects that of natural terrestrial materials. ⁱ Qi et al. value⁷ (normalized to $\delta^{13}\text{C}$ of L-SVEC $\equiv -46.6\text{‰}$) of -29.99 $\pm$ 0.05 ‰ is within 1 u_c of recommended value (Table S-5) and is in accord with Stalker et al.¹⁹ observation that prior $\delta^{13}\text{C}$ values of NBS 22 and other organic reference materials are more positive by $\sim 0.25\text{‰}$. ^j Ghosh et al. value¹⁵ (normalized to $\delta^{13}\text{C}$ of L-SVEC $\equiv -46.6\text{‰}$) of -41.61 ‰ ($u_c = 0.05\text{‰}$) is within 1 u_c of recommended value (Table S-7). ^k Both Stichler Bayesian value³ (normalized to $\delta^{13}\text{C}$ of L-SVEC $\equiv -46.6\text{‰}$) of -47.31 ‰ ($u_c = 0.06\text{‰}$) and Verkouteren and Klinedinst [Table 10-D] Bayesian value⁴ of -47.38 ‰ ($u_c = 0.06\text{‰}$) are within 1 u_c of recommended value (Tables S-6 and S-8).

to VPDB on a scale normalized by assigning consensus values of -46.6 ‰ to L-SVEC lithium carbonate and +1.95 ‰ to NBS 19 calcium carbonate, and authors should clearly state so in their reports. Authors are encouraged to report their measurement results for $\delta^{13}\text{C}$ values of NBS 22 oil, USGS41 L-glutamic acid, IAEA-CH-6 sucrose, or other internationally distributed reference materials, as appropriate for the measurement method concerned.

CONCLUSIONS

Anchoring the $\delta^{13}\text{C}$ scale at a second point by assigning a consensus value of -46.6 ‰ to the reference material L-SVEC will

Figure 1. Improvement in combined standard uncertainty for $\delta^{13}\text{C}$ reference materials compared with previously assessed uncertainty. Sizes of solid points indicate newly estimated uncertainties (largest solid circle is equivalent to an uncertainty of 0.06 ‰), diameter of open circles (older materials only) indicate their previously estimated uncertainties (largest open circle is equivalent to an uncertainty of 0.15 ‰). By consensus, $\delta^{13}\text{C}$ values of NBS 19 and L-SVEC have no associated uncertainty on the normalized scale.

help establish a worldwide equivalence among laboratories, that is, the ability of independent laboratories to measure and to report intercomparable $\delta^{13}\text{C}$ values that, for common samples, agree with one another within measurement uncertainty. With these guidelines, the $\delta^{13}\text{C}$ scale can now be anchored at two points, as has been done for the $\delta^2\text{H}$ and $\delta^{18}\text{O}$ scales during the last 3 decades, to improve data reliability of a wide variety of programs, including authenticating the origin of pharmaceuticals to combat counterfeiting, addressing megaton discrepancies in global atmospheric carbon inventories, and harmonizing results from international laboratories that test for use of anabolic steroids at the Olympic Games.

ACKNOWLEDGMENT

This manuscript benefited from reviews by Dr. Linda Stalker (CSIRO Petroleum, North Ryde, Australia) and H. Qi (U.S. Geological Survey). The expert analytical work of H. Qi (U.S. Geological Survey), H. Geilmann (MPI-BGC Jena), U. Guenther (UFZ Leipzig-Halle), and Anita Aerts and other laboratory technicians (CIO Groningen) that made this work possible is greatly appreciated.

SUPPORTING INFORMATION AVAILABLE

Excel file having eight worksheets containing Tables S-1–8 as noted in text. This material is available free of charge via the Internet at <http://pubs.acs.org>.

Received for review November 15, 2005. Accepted January 23, 2006.

AC052027C