

University of Groningen

## Exploring next-generation sequencing in clinical microbiology

Peker, Nilay

DOI:  
[10.33612/diss.218469374](https://doi.org/10.33612/diss.218469374)

**IMPORTANT NOTE: You are advised to consult the publisher's version (publisher's PDF) if you wish to cite from it. Please check the document version below.**

*Document Version*  
Publisher's PDF, also known as Version of record

*Publication date:*  
2022

[Link to publication in University of Groningen/UMCG research database](#)

*Citation for published version (APA):*  
Peker, N. (2022). *Exploring next-generation sequencing in clinical microbiology: identification, characterization and typing of pathogens*. University of Groningen. <https://doi.org/10.33612/diss.218469374>

### Copyright

Other than for strictly personal use, it is not permitted to download or to forward/distribute the text or part of it without the consent of the author(s) and/or copyright holder(s), unless the work is under an open content license (like Creative Commons).

The publication may also be distributed here under the terms of Article 25fa of the Dutch Copyright Act, indicated by the "Taverne" license. More information can be found on the University of Groningen website: <https://www.rug.nl/library/open-access/self-archiving-pure/taverne-amendment>.

### Take-down policy

If you believe that this document breaches copyright please contact us providing details, and we will remove access to the work immediately and investigate your claim.

*Downloaded from the University of Groningen/UMCG research database (Pure): <http://www.rug.nl/research/portal>. For technical reasons the number of authors shown on this cover page is limited to 10 maximum.*

## *Chapter 2*

### Diagnosis of bloodstream infections from positive blood cultures: recent developments in molecular approaches

**Nilay Peker**, Natacha Couto, Bhanu Sinha<sup>†</sup>, John W. Rossen<sup>†</sup>

University of Groningen, University Medical Center Groningen, Department of Medical Microbiology, Groningen, The Netherlands

<sup>†</sup>Authors contributed equally.

*Clinical Microbiology and Infection* 24 (2018) 944-955.

## **Abstract**

### ***Background***

Bloodstream infections are a major cause of death with increasing incidence and severity. Blood cultures are still the gold standard for microbiological diagnosis, but are rather slow. Molecular methods can be used as add-on complementary assays. They can be useful to speed up microbial identification and to predict antimicrobial susceptibility, applied to direct blood samples or positive blood cultures.

### ***Aim***

To review recent developments in molecular based diagnostic platforms used for the identification of bloodstream infections, with a focus on assays performed directly on blood samples and positive blood cultures.

### ***Sources***

Peer reviewed articles, conference abstracts and manufacturers' websites.

### ***Content***

We give an update on recent developments of molecular methods in diagnosing BSIs. We first describe the currently available molecular methods to be used for positive blood cultures including: a) in situ hybridization-based methods; b) DNA-Microarray-based hybridization technology; c) nucleic acid amplification-based methods; and d) combined methods. Subsequently, molecular methods applied directly to whole blood samples are discussed, including the use of nucleic acid amplification-based methods, T<sub>2</sub> magnetic resonance-based methods and metagenomics for diagnosing BSIs.

### ***Implications***

Advances in molecular-based methods complementary to conventional blood culture diagnostics and antimicrobial stewardship programs may optimize infection management by allowing rapid identification of pathogens and relevant antimicrobial resistance genes. Rapid diagnosis of the causing microorganism and relevant resistance determinants is important for early administration and modification of appropriate antimicrobial therapy. Ultimately, this may lead to improved quality and cost-effectiveness of health care, as well as reduced antimicrobial resistance selection.

**Keywords:** blood culture; blood pathogens; microbial identification; antimicrobial susceptibility testing; molecular diagnostics; bloodstream infections; diagnostic stewardship

## Introduction

Infections caused by the presence of viable microorganisms in blood are described as bloodstream infections (BSIs). They are often associated with severe diseases with high morbidity and mortality [1–3]. The worldwide increase in incidence emerges as one of the leading causes of death [4], especially in critically ill patients [5]. Increased mortality of BSIs is often related to delayed, insufficient or inappropriate anti-infective treatment [2,3,6].

Currently, blood cultures (BCs) are considered to be the gold standard for diagnosing BSIs and are still indispensable for the diagnosis of BSIs. The single most important factor for achieving a sensitivity of >95% is sampling sufficient volumes (i.e., at least two to three sets, filled with 10ml per bottle), since >50% of the patients present with a bacteremia of <1 colony-forming unit (CFU)/mL [7]. BCs are rather straightforward to perform and display a good sensitivity for cultivable pathogens [8]. However, major drawbacks remain: i) a rather long turnaround time [9]; ii) failure to identify slow-growing or obligate intracellular microorganisms and pathogens other than bacteria or yeast; [8]; and iii) substantial delay or even failure to detect microorganisms involved in BSIs when the patient has previously received antimicrobials [9].


These drawbacks may lead to important consequences. First of all, starting appropriate antimicrobial therapy for sepsis as soon as possible, preferably within the first 1 hour after clinical recognition of sepsis, is crucial to reduce morbidity and mortality [10]. Secondly, since culture-based diagnosis is slow, physicians often have to start empiric, broad-spectrum antimicrobial treatment. Therefore, rapid detection and identification of pathogen(s) is required for optimizing the therapeutic effect of antimicrobial therapy and for de-escalation of the started therapy to limit collateral damage, such as toxicity, side effects, and selection of antimicrobial resistance [2]. Additionally, the long turnaround time may also result in delayed specific clinical diagnosis, increased length of stay, a higher complication rate, and thus ultimately increase medical costs [8].

Consequently, for more than a decade pathogen-specific assays aiming at early identification of pathogens and some resistance determinants have been developed and tested. These molecular tools have been tested as an alternative for or complementary to BCs [8]. Depending on the Gram-staining results of the positive BC, microbiology laboratories may decide on which specific molecular test panel meet their needs. Available approaches, methods and

commercialized products available before 2015 have been reviewed previously [2,8,11,12]. This review gives an update on recent developments of molecular methods in diagnosing BSIs, focusing on starting from positive BCs. The data presented here was obtained from independent studies, and only when these were not available, information acquired in the manufacturers' websites was introduced.

### ***Currently available molecular methods***

Currently available molecular methods can be divided into four groups: a) *in situ* hybridization-based methods; b) DNA-microarray-based hybridization technology; c) nucleic acid amplification-based methods; and d) combined methods. The different methods reviewed are summarized in Table 1 and Table 2. In addition, Fig. 1 shows an overview of the entire process from patient-to-patient. After incubation of BCs and Gram-staining examination, samples can be plated on agar plates or can undergo direct molecular detection (Fig. 1). We also discuss a recently developed highly promising culture-free metagenomics-based method together with other platforms based on molecular techniques used for BSIs diagnosis directly from blood. These methods could potentially reduce time to detection further by passing the first incubation step and Gram-staining (blood culture bottles).



**Fig. 1.** Methods to identify microorganisms from positive blood cultures or blood. After physical examination of the patient the physician may decide to take a blood sample. Subsequently, the sample in a blood culture flask has to be transported to the laboratory, where it will be incubated for 6-120 hours at 35-37°C. After incubation either molecular methods can be applied, a MALDI-TOF analysis may be performed to identify the microorganism and their antimicrobial resistance genes or sub-culturing may be started after which identification takes place by MALDI-TOF analyses and antibiotic susceptibility testing. Metagenomics can be applied directly to the blood sample without an incubation step.

## ***Molecular methods from positive blood culture bottles***

### ***In situ hybridization-based methods***

A hybridization assay involves specific binding of two complementary nucleic acid strands in which one, the probe, is a nucleic acid sequence of known identity and the other one, the target, is part of the genome of an unknown microorganism to be identified [13]. Fluorescent *in situ* hybridization (FISH) assays are based on this principle. In FISH, fluorescently labeled oligonucleotide probes are used to specifically bind to the ribosomal DNA (rDNA) region in the genome of targeted pathogens, i.e. to the 16S rDNA gene of bacteria or to the 18S rDNA gene of fungi. This specific binding between probe and complementary sequence is visualized by fluorescence microscopy [14]. Although FISH technology allows identification of most commonly found bacteria and yeast in blood, some bacteria can be identified only at the genus level due to the limited number of species-specific probes available in the assay [14].

### ***PNA-FISH® and QuickFISH®***

The first commercialized FISH based technology was PNA-FISH (OpGen®, Maryland, USA, previously AdvanDx), which uses peptide nucleic acid probes (PNA) to detect *Staphylococcus aureus*, coagulase-negative staphylococci (*S. aureus*/CNS PNA-FISH®), *Enterococcus faecalis*, *E. faecium* and other *Enterococcus* species (*E. faecalis*/OE PNA-FISH®), *Escherichia coli*, *Klebsiella pneumoniae*, *Pseudomonas aeruginosa* (Gram-Negative PNA-FISH®) and *Candida* species (*Candida* PNA-FISH® [*C. glabrata* and *C. albicans*]; Yeast Traffic Light® PNA-FISH [5 *Candida* species]). The limit of detection (LOD) of 10<sup>5</sup> CFU/mL, the limited number of PNA-FISH probes, as well as the lack of the antimicrobial susceptibility profiles of the identified microorganisms are the weaknesses of the PNA-FISH [12]. PNA-FISH provides results within 2.5 hours, which is reduced to less than 30 min by a faster and less labor intensive version called QuickFISH®, launched in 2013 [15]. The QuickFISH® has a self-reposting probe design and the following panels are available: *Staphylococcus* QuickFISH®, *Enterococcus* QuickFISH®, Gram-Negative QuickFISH® and *Candida* QuickFISH® (not FDA-approved, but CE marked). Results of the QuickFISH® can be obtained within the same timeframe as the results of Gram stain. However, to be cost-effective it has been suggested to perform QuickFISH® after initial Gram staining [16]. PNA-FISH® and QuickFISH® exhibit a sensitivity and specificity of 97-100% and 90-100%, respectively [15,17].

### ***AccuProbe®***

The AccuProbe® system (Hologic®, previously Gen-Probe, USA) is also based on hybridization of DNA probes. Previously, Gen-Probe commercialized tests to identify *S. pneumoniae*, *S. aureus*, *Enterococcus* spp. and group A and B streptococci [18]. These tests have been developed for the identification of pathogens from cultures, however Lindholm and colleagues modified the manufacturer's instructions for specimen preparation to use a pellet of bacteria made directly from positive blood culture broth [18]. Sensitivity and specificity was shown to be above 97.9% for the detection of most pathogens with a turnaround time of 1 hour. For *S. aureus* cut off values had been modified as to increase sensitivity from 72.4% to 80.8%, whilst only marginally decreasing specificity from 99.8% to 98.7% [18]. Currently, according to the Hologic® website, only the tests to identify *S. pneumoniae* and *S. aureus* are available.

### ***Accelerate Pheno™***

Recently, Accelerate Diagnostics™ (Arizona, USA) has developed a fully automated hybridization system called Accelerate Pheno™, which identifies pathogens and performs antimicrobial susceptibility testing (AST) in approximately seven hours [19]. This system integrates gel filtration for the sample preparation and FISH technology for identification, combined with digital microscopy for subsequent AST [20]. The system performed well for the identification of Gram-positive and Gram-negative organisms with an overall sensitivity and specificity of 95.6% and 99.5%, respectively [21]. AST results were analyzed for essential agreement, defined as minimum inhibitory concentrations within plus or minus 1 doubling dilution of the reference method, and for categorical agreement, defined as susceptible, intermediate, and resistant results matching to those in reference methods [21]. Accordingly, an essential agreement of 95.1% and a categorical agreement of 95.5% were reported compared to the routine methods, which included for identification the Vitek®2 (bioMérieux) or biochemical assays and BactiStaph (Thermo Fisher Scientific, Lenexa, USA) ; and for antimicrobial susceptibility testing the Vitek®2 (bioMérieux), ETEST® (bioMérieux) and disk diffusion (BD Microbiology Systems, Sparks, USA) [21].

### ***DNA-microarray-based hybridization technology***

DNA-microarrays consist of short oligonucleotides immobilized on a solid surface. They allow simultaneous, in parallel detection of many pathogens and their antimicrobial resistance genes. The species covered by DNA-microarrays are in general around 90-95% of all the pathogens known to cause BSIs with a sensitivity ranging from  $10^1$  to  $10^5$  cells/mL [2].

### ***Verigene®***

The Food and Drug Administration (FDA) approved the DNA-microarray based Verigene® system (Luminex® Corporation, previously Nanosphere, USA) that employs two distinct multiplex panels for rapid diagnosis: one for 12 Gram-positive bacteria (12 species) along with associated resistance genes (*mecA*, *vanA*, *vanB*), and another for 9 Gram-negative bacteria (1 species not FDA approved) and their resistance markers (*bla<sub>NDM</sub>*, *bla<sub>VIM</sub>*, *bla<sub>KPC</sub>*, *bla<sub>OXA</sub>* and *bla<sub>CTX-M</sub>*). The system can identify bacterial species in each panel along with associated resistance markers directly from positive BCs with a sensitivity of 81-100% and a specificity of higher than 98 % [22,23]. However, the Verigene® panels demonstrated discordant results in mixed BCs in which no association of antimicrobial resistance marker genes to a specific organism was observed [23]. The turnaround time of the test is reported to be 2.5 hours.

### ***Nucleic acid amplification-based methods***

The polymerase chain reaction (PCR) is the most commonly used nucleic acid amplification technique for the identification of pathogens directly from blood as well as from positive BCs. The PCR tests used in diagnostic laboratories are based on either broad range PCR or multiplex PCR. Broad range PCR uses universal primers targeting conserved rDNA regions of the bacterial (16S) or fungal (18S) genome and subsequent identification of the microorganism by sequencing or genus/species-specific real-time PCR assays [24,25]. Commercially available multiplex assays are designed to target different chromosomal regions of the pathogens in combination with assays to detect their antimicrobial resistance signatures.

The loop mediated isothermal amplification (LAMP) is a novel nucleic acid amplification method that amplifies DNA with high sensitivity and efficacy similar to PCR while possessing advantages over PCR [26,27], as amplification occurs under isothermal condition which does not require sophisticated equipment. Therefore, it is suitable for the detection of pathogens in field-based settings [28], or in smaller satellite laboratories.

### ***FilmArray®***

The FilmArray® (bioMérieux, France, previously Biofire Diagnostics, USA) is an automated multiplex PCR platform which combines sample preparation, PCR amplification, detection and analysis ([www.biomerieux.com](http://www.biomerieux.com)). The FilmArray® Blood Culture Identification Panel (BCID) can identify 19 bacterial species and 5 *Candida* species, as well as 4 resistance genes (*mecA*, *vanA*, *vanB*, *bla<sub>KPC</sub>*) from positive BCs in 1 h with 2 min hands-on time. It performs well in polymicrobial cultures [29]. In a randomized control trial, all microbial targets were identified with a sensitivity exceeding 96% and a specificity of 99%. The resistance genes *vanA/B*, *bla<sub>KPC</sub>* were identified with 100% sensitivity and

specificity [30]. However, the sensitivity and specificity observed for the *mecA* gene were reported to be 98.4% and 98.3%, respectively. In addition, false positive detection of the *mecA* gene in coagulase-negative staphylococci and *S. aureus* has been demonstrated previously [29,30]. Despite the short turnaround time of the FilmArray® system, only one sample at a time can be tested on the same instrument [31]. To overcome this latter limitation the company has now introduced the FilmArray® Torch that harbors 2-12 modules within one machine, thereby providing up to six times more sample throughput per square foot of bench top space ([www.biomerieux.com](http://www.biomerieux.com)).

### **Xpert® MRSA/SA BC**

The Xpert® MRSA/SA BC assay of Cepheid® (Carlsbad, USA) is based on real-time PCRs that can identify methicillin-resistant *S. aureus* (MRSA) and methicillin-susceptible *S. aureus* (MSSA) within approximately 1 hour ([www.cephid.com](http://www.cephid.com)). The assay detects sequences in the staphylococcal protein A (*spa*) gene, methicillin resistance (*mecA*) gene, and the junction between the staphylococcal chromosomal cassette that harbors the *mecA* gene (SCC*mec*) and the *S. aureus* genome (*orfX*) [32]. Amplification of these 3 targets is needed for the assay to identify MRSA strains [32]. The sensitivity and specificity for the identification was reported to be 99.6% (96.4%-100%) and 99.5% (98%-100%), respectively, for *S. aureus* and 98.1% and 99.6%, respectively, for MRSA [33].

### **BD Max™ StaphSR Assay**

The BD Max™ StaphSR Assay (BD Diagnostic Systems, Québec, Canada) is a CE marked and FDA approved, fully automated diagnostic test which was developed for the detection of *S. aureus* and MRSA in patients at risk for nasal colonization [34]. The test is based on a multiplex real-time PCR and differentiates *S. aureus* and MRSA by detecting the presence of the *nuc* and *mecA/C* genes, and the SCC*mec-orfX* junction within approximately 2.5 hours [34]. Although the BD Max™ StaphSR Assay was designed for rapid screening of nasal swabs, the performance of the test was also analyzed for detecting *S. aureus* and MRSA from positive BCs [34-36]. The test had a sensitivity and specificity for *S. aureus* of 99.5% (99.1%-100%) and 100%, respectively; for MRSA of 98.9% (97.9%-100%) and 98.1%, respectively; and for CoNS of 99% and 97.9%(95.8%-100%), respectively [35,36].

### **eazyplex®**

Amplex Diagnostics GmbH (Giessen, Germany) marketed the CE approved eazyplex® test system, which combines isothermal amplification of the target and real-time photometric detection of amplified material without prior DNA

extraction. Previously, one of the eazyplex<sup>®</sup> assays called eazyplex<sup>®</sup> SuperBug CRE was tested directly on positive BCs spiked with defined Gram-negative strains to detect the most common carbapenemases and ESBLs [37]. The eazyplex<sup>®</sup> system displayed a correct identification of all of the  $\beta$ -lactamase genes included in the assay in less than 15 min and with a simple sample processing [37]. In a recent study, investigators evaluated the potential use of LAMP methods in BC diagnosis [38]. They used the eazyplex<sup>®</sup> MRSA test, which detects *S. aureus*, *S. epidermidis*, *mecA*, and *mecC* within 30 min. The sensitivity and specificity for the identification reported was 100% and 98.2%, respectively, for *S. aureus*; and 92.3% and 98.4%, respectively, for *S. epidermidis* [38].

### **Combined platforms**

Recently, some platforms were developed that combined the advantages of different molecular methods.

#### **Sepsis Flow Chip**

Sepsis Flow Chip (Master Diagnóstica, Granada, Spain) is a novel CE labeled diagnostic assay, able to detect most of the pathogens causing BSIs and their resistance markers based on a multiplex PCR, followed by reverse dot blot hybridization [39]. Sepsis Flow Chip allows simultaneous detection of over 36 bacterial species, *Candida* spp. (*C. albicans* and other *Candida* spp.) and 20 antimicrobial resistance markers within 3 hours [40]. The clinical evaluation study of the Sepsis Flow Chip yielded promising results with an overall sensitivity and specificity of 93.3% and 100%, respectively, for bacterial identification, and 93.6% and 100%, respectively, for the detection of the antimicrobial resistance determinants [39]. In this validation study, a 96.2% and 89.1% concordance with standard microbiological diagnostic protocols was shown in monomicrobial and polymicrobial cultures, respectively [39].

#### **ePlex<sup>®</sup>**

The ePlex<sup>®</sup> system (GenMarkDx<sup>®</sup>, Carlsbad, USA) is a CE marked, automated and multiplex molecular diagnostic test system. It integrates a digital microfluidics-based sample preparation and an amplification assay together with the eSensor<sup>®</sup> detection technology (GenMarkDx<sup>®</sup>, USA) for the electrochemical detection of the target DNA within the same cartridge ([www.genmarkdx.com](http://www.genmarkdx.com)). Therefore, this fully automated system is expected to reduce sample contamination risk and it eliminates the time consuming washing and preparation steps required in earlier multiplexing technologies [41]. In contrary to fluorescence or optical detection of the first generation multiplex assays, the ePlex<sup>®</sup> detection system is highly specific and based on

competitive DNA hybridization and electrochemical detection [41]. It consists of different multiplex blood culture identification panels (BCID, Gram-positive [BCID-GP], Gram-negative [BCID-GN] and fungal pathogen [BCID-FP]), which are currently under development for the detection of common bloodstream infections. In a preliminary study, the ePlex® BCID panels were retrospectively tested with a turnaround time of 1.5 hours in clinical samples in which the discrepant results compared to routine test of the laboratory were examined by culture, sequencing or qPCR [42]. After discrepancy analysis, 100%, 97.8% and 97.8% concordance were reported for the detection of fungal pathogens, Gram-negative and Gram-positive bacteria, respectively. The BCID panels were also tested on polymicrobial cultures and able to detect polymicrobial infections of bacteria and/or fungi with 95.5% concordance compared to the control methods after discordant resolution [42,43].

### ***Molecular methods directly from blood (without previous incubation)***

Molecular techniques performed directly on whole blood samples allow rapid identification of the etiological agent of BSIs independently of a prior incubation phase. Direct diagnosis from whole blood circumvents the limitations of blood culturing in identifying slow-growing bacteria or uncultivable microorganisms, and in cases when the patient has already received antimicrobials. Also, processing of smaller volumes of blood samples compared to BCs makes these molecular methods especially useful for pediatric patients in whom it is difficult to obtain larger volumes of blood [11]. However, molecular approaches developed for the detection of microorganisms are mostly PCR-based assays, which introduce some challenges. The presence of high amounts of human DNA in patient blood, as well as other blood components can interfere with PCR and inhibit the reaction [44,45]. Although PCR assays are usually very sensitive, they may yield false positive results because of contaminating DNA (e.g. from the environment, or PCR reagents), or cell-free DNA from dead microorganisms. Also, DNA originating from pathogens and released after an infection controlled by the immune system or an efficient antimicrobial therapy can persist in the circulation for several days [11]. Here, we discuss recently used platforms and the promising approach of metagenomics used for diagnosis of BSIs directly from whole blood.

### ***Nucleic acid amplification-based methods***

#### ***LightCycler® SeptiFast***

The LightCycler® SeptiFast (Roche Molecular System, Basel, Switzerland) was developed based on a multiplex real time PCR assay. The assay is able to identify most commonly encountered pathogens of BSIs (19 bacterial and 6 fungal) directly from 1.5 mL of whole blood in about 6 hours. In a subsequent run, the presence of methicillin resistance in *S. aureus* positive samples is assessed. This assay uses fluorescence resonance energy transfer (FRET) probes targeting the internal transcribed spacer (ITS) regions between the 16S and 23S ribosomal DNA of bacteria and between the 18S and 5.8S ribosomal DNA of the fungal genomes for identification. After completion of the run data analysis is performed by manual editing of the PCR data using Light Cycler 2.0 software, and bacterial and fungal identification is accomplished through automated analyses with SeptiFast Identification Software (SIS) [46]. Although the test is not available in the USA, there have been several clinical studies conducted to evaluate its performance since it was commercialized in Europe. In a study for the diagnosis of CoNS in neutropenic hematological patients, the SeptiFast showed a sensitivity of 23.3% that was increased to 83.3% by neglecting the integrated cut off value [47]. In a recent study, 85.5% concordance between BC and SeptiFast was shown. Higher positivity rates in patients with sepsis and pneumonia, and immunocompromised patients with febrile episodes were reported than in a patients with endocarditis [48]. For the diagnosis of neonatal sepsis, reported sensitivity and specificity were 90.2% and 72.9%, respectively when using a modified multiplex PCR protocol [49]. As different patient populations were studied, variable sensitivities and specificities were reported, and variable concordances with conventional BCs were obtained. A drawback of the system is that identification is done without quantification of the pathogen. Indeed, reporting quantitative data might be useful in assessing the severity of the disease [50].

### ***Magicplex™ Sepsis Real-time Test***

Magicplex™ Sepsis Real-time Test (Seegene, Seoul, Korea) is a CE-approved automated assay which makes use of conventional PCR after DNA extraction and followed by real time PCR for the identification from 1 mL blood with a turnaround time of 3 to 6 hours. The assay has an extensive panel developed for the detection of more than 90 pathogens including a further identification of 27 pathogens at species level and 3 resistance genes (*mecA*, *vanA*, *vanB*). Clinical studies show reported sensitivities and specificities of 37%-65% and 66%-92%, respectively [51-53]. In a later study, test performance was improved by adjusting the cut off level of the quantification cycle (Cq) values [53]. In this study, setting a lower cut off for the Cq values resulted in a decrease in sensitivity to 38%, but an increase in PPV to 59% and in specificity to 95% together with less contaminant detection. Absence of quantification and lack of identification of low number of pathogens at species level are limitations of the assay [11].

Table 1. Commercially available methods for the identification of microorganisms and detection of the presence of antimicrobial resistant genes directly from positive blood cultures.

Assay	Gram staining need	Pathogens covered (Panel/Kit)	Resistance markers	Sensitivity /Specificity	Complexity -Personnel experience level -Equipment requirement	Hands-on Time	Times to result	References
<b>In situ hybridization-based methods</b>								
PNA-FISH* (OpGen®, USA)	Yes	4 Gram-positive bacteria ( <i>S. aureus</i> /CNS PNA-FISH* and <i>E. faecalis</i> /OE PNA-FISH*) 3 Gram-negative bacteria (Gram-Negative PNA-FISH*) 5 <i>Candida</i> spp. ( <i>Candida</i> PNA-FISH*, Yeast Traffic Light* PNA-FISH)	none	97-100% / 90-100%	Not automated -Trained personnel -Proprietary equipment	10 min	2.5 hours	[12,15,17] www.opgen.com
QuickFISH* (OpGen®, USA)	Yes	4 Gram-positive bacteria ( <i>Staphylococcus</i> QuickFISH*, <i>Enterococcus</i> QuickFISH*) 3 Gram-negative bacteria (Gram-Negative QuickFISH*) 3 <i>Candida</i> spp. ( <i>Candida</i> QuickFISH*)	none	97-100% / 90-100%	Not automated -Trained personnel -Generic equipment	5 min	30 min	[15-17] www.opgen.com
AccuProbe® (Hologic®, USA)	Yes	<i>S. pneumoniae</i> ( <i>S. pneumoniae</i> culture identification test) <i>S. aureus</i> ( <i>S. aureus</i> culture identification test)	none	>97% / 81-100%	Not automated -Trained personnel -Generic equipment	5 min	1 hour	[18] www.hologic.com
Accelerate PhenoTest™ BC (Accelerate Diagnostics™, USA)	No	15 Gram-positive bacteria 11 Gram-negative bacteria 2 <i>Candida</i> spp. (Accelerate PhenoTest™ BC Kit)	20 antimicrobial MICs determined	96% / 99%	Fully automated -No trained personnel -Proprietary equipment	2 min	1 hour ID 6 hours AST	[19-21] www.acceleratediagnostics.com
<b>DNA-Microarray based methods</b>								
Verigene® (Luminex® Corporation, USA)	Yes	12 Gram-positive bacteria (VERIGENE® Gram-Positive Blood Culture Test) 9 Gram-negative bacteria (VERIGENE® Gram-Negative Blood Culture Test)	<i>mecA</i> , <i>vanA</i> , <i>vanB</i> , <i>bla<sub>TEM</sub></i> , <i>bla<sub>VIM</sub></i> , <i>bla<sub>SHV</sub></i> , <i>bla<sub>OXA</sub></i> , <i>bla<sub>CTXM</sub></i>	81-100% / >98%	Fully automated -No trained personnel -Proprietary equipment	5 min	2.5 hours	[22,23] www.luminexcorp.com
<b>Nucleic-acid amplification-based methods</b>								
FilmArray® (bioMérieux, France)	No	8 Gram-positive bacteria 11 Gram-negative bacteria 5 <i>Candida</i> spp. (FILMARRAY® BCID Panel)	<i>mecA</i> , <i>vanA</i> , <i>vanB</i> , <i>bla<sub>SHV</sub></i>	>96% / 98-100%	Fully automated -No trained personnel -Proprietary equipment	2 min	1 hour	[29,30] www.biomerieux.com
Xpert® MRSA/SA BC (Cepheid®, USA)	Yes	5 <i>S. aureus</i> , MRSA	<i>mecA</i>	98-100% / 99-100%	Fully automated -No trained personnel -Proprietary equipment	1 min	1 hour	[31,32] www.cepheid.com
BD Max™ StaphSR Assay (BD Diagnostics, Canada)	Yes	<i>S. aureus</i> , MRSA, CoNS	<i>mecA</i> , and <i>mecC</i>	98-100% / 98-100%	Fully automated -Trained personnel -Proprietary equipment	1 min / sample preparation	2.5 hours	[34-36] www.bd.com
®Eazyplex® MRSA (Amplex Diagnostics GmbH, Germany)	Yes	<i>S. aureus</i> , <i>S. epidermidis</i> (eazyplex® MRSA)	<i>mecA</i> , and <i>mecC</i>	100% / 98%	Fully automated -Trained personnel -Proprietary equipment	2 min preparation	30 min	[37,38] www.eazyplex.com
<b>Combined methods (Multiplex PCR and hybridization based methods)</b>								
®Sepsis Flow Chip (Master Diagnostica, Spain)	No	>76 bacteria <i>Candida</i> spp (non-albicans). <i>Candida albicans</i> (Sepsis Flow Chip Kit)	20 antimicrobial resistance markers	93-94% / 100%	Fully automated -Trained personnel -Proprietary equipment	Not provided	3 hours	[39,40] www.masterdiagnostica.com
®ePlex® BCID (GenMarkDx®, USA)	Yes	20 Gram-positive bacteria (BCID-GP Panel) 21 Gram-negative bacteria (BCID-GN) 16 yeasts (including 10 <i>Candida</i> spp.) (BCID-FP Panel)	BCID-GP: <i>mecA</i> , <i>mecC</i> , <i>vanA</i> , <i>vanB</i> BCID-GN: <i>bla<sub>CTXM</sub></i> , <i>bla<sub>SHV</sub></i> , <i>bla<sub>OXA</sub></i> , <i>bla<sub>TEM</sub></i> , <i>bla<sub>SHV</sub></i>	Not provided	Fully automated -No trained personnel -Proprietary equipment	2 min	1.5 hours	[41,42] www.genmarkdx.com

Table 1: Specifications of the assays have been collected from the literature and manufacturer's information.  
\*Not available in the USA.

## SepsiTest™

The SepsiTest™ (Molzzy, Bremen, Germany) is a CE-marked assay combining broad range PCR and sequencing. The test performs a broad range PCR by using universal primers targeting the bacterial 16S and the fungal 18S rDNA regions, followed by sequencing of the amplicon product from positive samples and subsequent BLAST analysis for the identification of the pathogens at the genus and species level. The assay processes blood volumes as low as 1 mL and can also be used on samples from other primary sterile body fluids with turnaround times of 8 to 12 hours [11,54,55] including a hands on time of 75 min for PCR and 70 min for sequence analysis [56]. In a multicenter study of blood samples taken from patients with different pathologies, reported sensitivity

and specificity of the test was 87% and 85.8%, respectively [54]. Later, the assay was evaluated for the monitoring of patients for systemic infections during extracorporeal membrane oxygenation [57]. The assay was able to identify 45% of positive samples, 13 to 75 hours earlier than BC and detected blood culture-negative infections in 25% of patients [57]. Yet, detection of any resistance genes is not available in the assay.

## ***T<sub>2</sub> Magnetic resonance-based methods***

T<sub>2</sub> Magnetic resonance is a magnetic resonance-based diagnostic method that measures changes in the water proton T<sub>2</sub> relaxation signal in the presence of magnetic fields [14]. It usually includes a pathogen-specific PCR amplification step, followed by hybridization of the amplicons to probe-enriched superparamagnetic nanoparticles [14].

### ***T<sub>2</sub>Candida® Panel***

T<sub>2</sub>Candida® Panel (T<sub>2</sub>Biosystems, Lexington, Massachusetts, USA) is an FDA approved diagnostic assay based on T<sub>2</sub> magnetic resonance (T<sub>2</sub>MR) detection technology which measures reacting water molecules in the presence of magnetic fields. T<sub>2</sub>Candida enables identification of five *Candida* species directly in whole blood without the need for extraction of DNA with a time to result of 3-5 hours and a LOD of 1 CFU/mL [58,59]. In a multicenter study, sensitivity and specificity of the assay were 99.4% and 91.1%, respectively [60]. In a further study, potential use of T<sub>2</sub>MR technology as a monitoring assay for candidemia has been demonstrated [61]. Most recently, T<sub>2</sub>Biosystems has launched a CE-marked T<sub>2</sub>Bacteria® Panel, which identifies the six most common BSI-causing bacterial species (*E. coli*, *K. pneumoniae*, *P. aeruginosa*, *Acinetobacter baumannii*, *S. aureus* and *E. faecium*) which may not respond to empiric therapy depending on the regional epidemiology.

### ***Metagenomics***

Metagenomics is the genomic analysis of specific or all genetic material present in and directly from a (clinical) sample. Sequencing of nucleic acids within a sample can be based on a specific amplicon (amplicon-based metagenomics) or on the entire genomes (shotgun metagenomics). The 16S and 16S-23S rDNA amplicon-based metagenomics methods are based on parallel deep sequencing of the prokaryote-specific 16S ribosomal RNA gene [62-65] or the entire 16S – intergenic spacer region – 23S rDNA region, respectively [66]. These methods have been applied successfully for the diagnosis of bacterial BSIs from BCs [62,65,66] or directly from blood [63-65], with equal or superior sensitivity compared to blood culture. However, no commercially available method has

been developed so far. Amplicon-based metagenomics has the disadvantage of only identifying a specific set of pathogens, in this case, bacteria. Shotgun metagenomics, on the other hand, has the advantage of not being limited to certain pathogens, and provides information on all microorganisms and viruses present in a sample, even if the sample contains a complex (polymicrobial) community [67]. In addition, shotgun metagenomics can detect antimicrobial resistance and virulence genes and provide important information on molecular epidemiology [68]. Therefore, shotgun metagenomics represents an essential add-on in pathogen detection compared to conventional blood culture or amplicon-based sequencing of suspected BSI with an even higher sensitivity and reliability. So far shotgun metagenomics has only been applied to blood or blood components [67,69,70] and only one commercially available tool has been developed.

### ***iDTECT<sup>TM</sup> Dx Blood***

The *iDTECT<sup>TM</sup> Dx Blood* test (PathoQuest SAS, France) is a new sequencing approach for the identification of pathogens and was originally developed at the Institute Pasteur ([www.pathoquest.com](http://www.pathoquest.com)). To this end, blood samples are prepared (DNA/RNA extraction, library preparation) and then analyzed using next-generation sequencing technology to identify genomic sequences of pathogens. In a multicenter, blinded, prospective, proof-of-concept study, the investigators compared untargeted next-generation sequencing with conventional microbiological methods for first-line diagnosis of infection in 101 immunocompromised adults [69]. The method showed a high negative predictive value (NPV) compared to conventional methods (64/65, 95% CI 0.95-1), and also detected more clinically relevant viruses and bacteria than conventional microbiological methods [69]. However, there were two discordant cases that were positive by conventional microbiological methods but negative by sequencing. In one case, this was due to contamination of the negative control with *E. coli* DNA, which led to the *E. coli* results not being reported in the NGS analysis. Currently, the detection of yeasts/fungi is hampered by limitation of the database and extraction procedures. In addition, introduction within a clinical lab may be hampered by the high costs per sample, the batch-wise approach of testing and the highly sophisticated infrastructure needed to perform the sequencing [69,71].

Table 2. Commercially available methods for the identification of microorganisms and detection of the presence of antimicrobial resistant genes directly from whole blood.

Assay	Pathogens covered	Resistance markers	Sensitivity / Specificity	Complexity -Personnel experience level -Equipment requirement	Hands-on Time	Times to result	References
<b>Nucleic-acid amplification-based methods</b>							
The LightCycler®	19 Bacteria	<i>mecA</i>	83-90% / 73%	Automated	3 hours	6 hours	[47,49,50]
SeptiFast (Roche Molecular System, Switzerland)	5 <i>Candida</i> spp. and <i>Aspergillus fumigatus</i>			-Trained personnel -Proprietary equipment			www.molecular.roche.com
*Magicplex™ Sepsis Real-time Test (SeeGene, Korea)	>90 pathogens at species level	<i>mecA</i> , <i>vanA</i> , <i>vanB</i>	37-65 / 66-92%	Multi-step automated -Trained personnel -Proprietary equipment	not provided	3-6 hours	[51-53] http://www.see.gene.com
*SepsiTest™ (Molzym, Germany)	345 bacteria, <i>Candida</i> , <i>Cryptococcus</i> and <i>Aspergillus</i> species	none	87% / 86%	Partially automated -Trained personnel -Generic equipment	75 min (PCR) 70 min (sequence analysis)	8-12 hours	[54-57] www.sepsitest.com
<b>T2 Magnetic resonance-based methods</b>							
T2Candida Panel (T2Biosystems, USA)	5 <i>Candida</i> species	none	99% / 91%	Fully automated -No trained personnel -Proprietary equipment	5 min	3-5 hours	[58-61] www.t2biosystems.com
<b>Metagenomics</b>							
iDTECT™ Dx Blood (PathoQuest SAS, France)	> 1200 bacteria and viruses	none	not provided	Not provided -Trained personnel -Generic equipment but proprietary software	not provided	not provided	[69] (www.pathoquest.com)

Table 2: Specifications of the assays have been collected from the literature and manufacturer's information.

\*Not available in the USA

## Platforms currently not available

Marketing of some molecular assays or platforms has been discontinued by the manufacturer. Since their performance in diagnosis of BSIs was evaluated in recent studies and, moreover, assays a, e.g., IRIDICA has been reintroduced in the past, they are briefly discussed (Table 3).

### BD GeneOhm™ StaphSR

The BD GeneOhm™ StaphSR assay (Becton Dickinson Diagnostics, Franklin Lakes, USA) was a multiplexed real-time PCR based assay, allowed the detection of *S. aureus* and MRSA directly from positive BCs. This assay differentiated MSSA from MRSA by detecting two targets; one specific for *S. aureus* and another specific for the *SCCmec-orfX* junction site in MRSA [33,72]. The BD GeneOhm™ StaphSR assay displayed a turnaround time of approximately 2.5 h [72]. Although the sensitivity and specificity of the test were 99.2% and 96.5% for MSSA and 94.3% and 97.8%, for MRSA, a poorer performance was reported in a study conducted in the USA [33]. This test has been discontinued and replaced by the newer version BD Max™ StaphSR Assay.

### Prove-it™ Sepsis

Prove-it™ Sepsis (Mobidiag, Espoo, Finland) was developed as a combination of a broad range PCR and a DNA-microarray. This assay identified the pathogens through the amplification of the topoisomerase gene targets (*parE* and *gyrB*) and further detection was done by microarray hybridization within

3 hours directly from BCs [8]. The first version of the test was able to identify 50 bacterial species as well as the *mecA* gene with a sensitivity and specificity of 94.7% and 98.8%, respectively [73]. In the latter version of the Prove-it Sepsis assay (Strip Array 2.0), pathogen coverage of the DNA-Microarray panel had been extended to include yeast targets as well. It was able to detect 13 yeasts (twelve *Candida* species and *Saccharomyces cerevisiae*) and more than 60 bacteria, covering most of the sepsis-associated pathogens. In a recent clinical evaluation study, sensitivity and specificity of the test for the detection of yeast were 99% and 97%, respectively [74]. Although the turnaround time for the assay was said to be 3 hours, an extra hour was required for the sample preparation [74]. While the results were promising, the product has now been discontinued.

### The IRIDICA™ BAC BSI Assay

The IRIDICA™ BAC BSI Assay (Ibis Biosciences, (Ibis Biosciences, Abbott, Carlsbad, USA) was developed based on a PCR/electrospray ionization-mass spectrometry (PCR/ESI-MS) technology. Principally, following PCR, the mass of the PCR amplicons was determined by ESI-MS, then the nucleotide base composition was calculated based on the mass spectra and compared to a database to identify the microorganism. This assay, marketed in Europe for use in clinical research, could identify more than 600 bacteria and *Candida* species and also detect resistance genes *mecA*, *vanA*, *vanB* and *bla<sub>KPC</sub>* from 5 mL of whole blood within 6 hours [75]. An evaluation of four studies comparing the IRIDICA BAC BSI assay and blood culture found the estimated summary specificity and sensitivity of the assay to be 84% and 81%, respectively [76]. Although the results were favorable to implement the assay in routine diagnosis, the manufacturer decided to cease producing the IRIDICA assay in April 2017 [77].

Table 3. Methods currently not available on the market.

Assay	Pathogens covered	Resistance markers	Sensitivity / Specificity	(Method) Complexity -Personnel experience level -Equipment requirement	Hands-on Time	Times to result	References
BD GeneOhm™ StaphSR (BD Diagnostics, USA)	<i>S. aureus</i> , MRSA	none	94-99% / 97-98%	(Multiplex Real-time PCR) Not automated -Trained personnel -Proprietary equipment	Not provided	2.5 hours	[33,72] www.bd.com
Prove-it™ Sepsis: StripArray v2.0 (Mobidiag, Finland)	>60 Bacteria 12 <i>Candida</i> spp. and <i>Saccharomyces cerevisiae</i>	<i>mecA</i>	95-99% / 97-99%	(Multiplex PCR and DNA-Microarray) Partially automated* -Trained personnel -Proprietary equipment	1.5 hour	3-4 hours	[73,74] www.mobidiag.com
IRIDICA BAC BSI Assay (Ibis Biosciences, Abbott, USA)	>600 bacteria and <i>Candida</i> species	<i>mecA</i> , <i>vanA</i> , <i>vanB</i> , <i>bla<sub>KPC</sub></i>	83-99% / 87-94%	(PCR/ESI-MS) Fully automated -Trained personnel -Proprietary equipment	30 min	6 hours	[75-77] www.molecular.abbott

\*Table 3: Specifications of the assays have been collected from the literature and manufacturer's information.  
\*Automated software for detection and analysis.

## *Clinical Outcome and Impact of Molecular Diagnostic Tests*

The specific requirements, performance and limitations of the molecular techniques described above vary to a large extent (summarized in Tables 4 and 5). None of the molecular methods is yet able to completely replace conventional culture-based methods. This is mainly due to the importance of phenotypic antimicrobial susceptibility determination (e.g. identification of minimum inhibitory concentrations [MIC]) for tailored, directed treatment, but also due to a rather limited spectrum and sensitivity of most available assays. Thus, they need to be performed as on-top assays to standard methods for identification and AST. Furthermore, identification by these methods has now to compete with identification protocols by MALDI-TOF MS directly from positive BCs – which is ideally suited for on-demand diagnostics and has a time-to-result of 30 to 60 min (from positive blood culture bottles). Nevertheless, the molecular methods reviewed here may result in a significant reduction in the turnaround time needed for microbial identification and susceptibility testing. It is difficult to make general statements with regard to costs of each specific method, because each method has different properties and requirements. Cost-effectiveness calculations would depend to a large extent on several factors, such as the patient population served, logistics, service hours of the laboratory, etc. The globally rather high unit cost price and often additionally required infrastructure (specific assay platforms, etc.), as well as the varying on-demand availability would require a thorough analysis with regard to the specific setting (patient population served, available infrastructure, 24/7 service implemented, laboratory structure and requirements, reimbursement, epidemiology, empiric regimens and switch policies, etc.) in order to make useful statements. Ultimately, the use of such methods may be cost-effective if considering other costs allocated to antimicrobial use and hospital stay, as rapid identification results allow early start of (or change to) appropriate antimicrobial therapy [78]. Indeed, some recent studies have evaluated the impact of the rapid molecular tests on improved clinical outcome by means of time to optimal therapy and also association with length of stay, mortality and costs. However, the impact of rapid BC diagnostics on clinical and economical outcomes depends on multiple aspects specific to each institution, namely the Antimicrobial Stewardship Program (ASP) [78]. Indeed, most of the observational studies suggest that rapid testing implemented with an ASP provides more favorable clinical outcomes than rapid testing alone [78]. A recent systematic review and meta-analysis investigated the clinical outcome of using rapid molecular tests for diagnosis of BSIs by analyzing thirty-one studies [79]. The impact of these assays was also evaluated in the presence of an ASP intervention. Rapid molecular testing showed significant decrease in mortality risk, decrease in time to effective therapy and shortened in length of stay, when implemented

with an ASP intervention [79]. In light of the evidence provided by these studies, molecular tests should be integrated as part of the diagnostic workflow in patients with BSIs.

Table 4. Requirements, benefits, and limitations of molecular techniques from positive blood cultures.

	Blood subculture	<i>In situ</i> hybridization-based methods	DNA-Microarray based methods	Nucleic-acid amplification based methods	Combined methods
<b>Duration*</b>	Very slow	Fast or Very fast	Fast	Fast or Very fast	Fast
<b>Complexity</b>	Laborious	Laborious	Simple	Simple	Simple
<b>Role</b>	Standard diagnostics	On-top assay	On-top assay	On-top assay	On-top assay
<b>Limit of detection</b>	1 – 10 CFU/ml	10 <sup>6</sup> CFU/ml	10 <sup>6</sup> – 10 <sup>7</sup> CFU/ml	10 <sup>2</sup> – 10 <sup>6</sup> CFU/ml	10 <sup>2</sup> – 10 <sup>6</sup> CFU/ml <sup>f</sup>
<b>Suitability for therapy monitoring</b>	Yes (bacteremia; conditionally: time to positivity)	Conditionally (with live/dead staining)	No (only conditionally indirectly via: time to positivity of culture)	No (only conditionally indirectly via: time to positivity of culture)	No (only conditionally indirectly via: time to positivity of culture)
<b>Costs<sup>b</sup></b>	< 10€ (if negative), 10-50€ (if positive)	10-50€	50-250€	30-150€	100-150€
<b>Laboratory type (setting)</b>	Satellite	Satellite	Centralized or academic	Satellite or standard (assay-dependent)	Centralized or academic
<b>Limitations</b>	Very specialized technicians	Very specialized technicians Most methods do not provide information about antimicrobial resistance	Variable sensitivity	False-positive results due to artifacts	Limited clinical evaluation studies

Table 4: \*Four categories were considered: Very slow (6-48h); Slow (3-6h); Fast (1-3h); Very fast (< 1h).

<sup>b</sup>Cost per test was considered, not the system cost. N.B. **Cost-effectiveness analyses require more complex calculations, considering total cost and clinical impact (c.f. to ref [76], chapter 3 for an example).**

<sup>f</sup>LOD was given for ePlex BCID-FP panel. There are limited number of studies evaluating the analytic sensitivity of ePlex<sup>®</sup> BCID and Sepsis Flow Chip. Quantification is not addressed since these assays all start from highly enriched cultures.

Table 5. Requirements, benefits, and limitations of molecular techniques from whole blood.

	Blood culture	Nucleic-acid amplification based methods	T2 Magnetic resonance based methods	Metagenomics
<b>Duration*</b>	Very slow	Fast or Very fast	Slow	Very slow
<b>Complexity</b>	Laborious	Simple	Simple	Laborious
<b>Role</b>	Not applicable	On-top assay	On-top assay	On-top assay
<b>Limits of detection</b>	1 – 10 CFU/ml	3 – 30 CFU/ml	1 CFU/ml	Possibly similar to nucleic acid amplification based methods, depending on the depth of sequencing
<b>Quantification possibilities</b>	Yes (conditionally: time to positivity)	Possible	Not possible	Semi-quantitative
<b>Suitability for therapy monitoring</b>	Yes (bacteremia)	Potentially (conditionally; indirectly via quantification)	No	Potentially (conditionally; indirectly via quantification)
<b>Costs<sup>b</sup></b>	< 10€ (if negative), 10-50€ (if positive)	50-250€	150-200€	150-300€
<b>Laboratory type (setting)</b>	Satellite	Standard	Centralized or academic	Centralized or academic
<b>Limitations</b>	Very specialized technicians	False-positive results due to artifacts	Does not provide information on antifungal resistance	Very specialized technicians

Table 5: \*Four categories were considered: Very slow (6-48h); Slow (3-6h); Fast (1-3h); Very fast (< 1h).

<sup>b</sup>Cost per test was considered, not the system cost. N.B. **Cost-effectiveness analyses require more complex calculations, considering total cost and clinical impact (c.f. to ref [76] chapter 3 for an example).**

## Conclusions

In general, molecular methods available to date aim at cutting down on substantial incubation time that is still required for culture-based methods. Mainly due to sensitivity issues, most approaches and assays have been developed to start with positive BCs, still leaving a detection gap for non-cultivable some fastidious microorganisms. Assays starting directly from blood of affected patients have probably the highest potential with this regard, provided they perform with a sensitivity that equals at least that of current BC diagnostics. In addition, these (metagenomics) approaches have the potential

to compensate for inherent shortcomings of culture-based diagnostics. Provided, aspects such as on-demand availability, platform commonality, data analysis and clinical interpretation, as well as robustness are adequately addressed, this approach could represent a quantum leap comparable to the introduction of MALDI-TOF MS in microbiological diagnostics, and thus substantially improve quality of patient care.

### **Transparency Declaration**

NP has received funding from the European Union's Horizon 2020 research and innovation program under the Marie Skłodowska-Curie grant agreement 713660 (MSCA-COFUND-2015-DP "Pronkjewail"), which includes in kind contributions by commercial partners. None of the commercial partners had any influence on interpretation of reviewed data and conclusions drawn, nor on drafting of the manuscript. No support was obtained from the companies of which the products were discussed in this review. In addition, NP, NC, BS and JR have no conflicts of interest to declare.

### **Acknowledgments**

Part of the content of this article was presented at the ESCMID postgraduate education course (PGEC) entitled "Improving the Diagnosis of Bloodstream Infections - Advancing Technology and Quality for Better Care" held in Nice on 28 March 2017. The course was organized by the following ESCMID study groups: ESGBIS, ESGB, ESGEM, ESGMD, ESGS and EFISG.

## References

1. Rutanga JP, Nyirahabimana T. Clinical Significance of Molecular Diagnostic Tools for Bacterial Bloodstream Infections: A Systematic Review. *Interdiscip Perspect Infect Dis* 2016;2016. doi:10.1155/2016/6412085.
2. Opota O, Croxatto A, Prod'hom G, Greub G. Blood culture-based diagnosis of bacteraemia: state of the art. *Eur Soc Clin Microbiol Infect Dis* 2015;21:313–22. doi:10.1016/j.cmi.2015.01.003.
3. Loonen AJM, Wolffs PFG, Bruggeman CA, van den Brule AJC. Developments for improved diagnosis of bacterial bloodstream infections. *Eur J Clin Microbiol Infect Dis* 2014;33:1687–702. doi:10.1007/s10096-014-2153-4.
4. Heron M. Deaths: Leading Causes for 2014. *Natl Vital Stat Rep* 2016;65:1–96.
5. Naghavi M, Wang H, Lozano R, Davis A, Liang X, Zhou M, et al. Global, regional, and national age-sex specific all-cause and cause-specific mortality for 240 causes of death, 1990–2013: a systematic analysis for the Global Burden of Disease Study 2013. *Lancet* 2015;385:117–71. doi:10.1016/S0140-6736(14)61682-2.
6. Bearman GML, Wenzel RP. Bacteremias: a leading cause of death. *Arch Med Res* 2005;36:646–59. doi:10.1016/j.arcmed.2005.02.005.
7. Towns ML, Jarvis WR, Hsueh PR. Guidelines on Blood Cultures. *J Microbiol Immunol Infect* 2010;43:347–9. doi:10.1016/S1684-1182(10)60054-0.
8. Dubourg G, Raoult D. Emerging methodologies for pathogen identification in positive blood culture testing. *Expert Rev Mol Diagn* 2016;16:97–111. doi:10.1586/14737159.2016.112274.
9. Fenollar F, Raoult D. Molecular diagnosis of bloodstream infections caused by non-cultivable bacteria. *Int J Antimicrob Agents* 2007;30:7–15. doi:10.1016/j.ijantimicag.2007.06.024.
10. Rhodes A, Evans LE, Alhazzani W, Levy MM, Antonelli M, Ferrer R, et al. Surviving Sepsis Campaign: International Guidelines for Management of Sepsis and Septic Shock: 2016. *Intensive Care Med* 2017;43:304–77. doi:10.1007/s00134-017-4683-6.
11. Opota O, Jatou K, Greub G. Microbial diagnosis of bloodstream infection: towards molecular diagnosis directly from blood. *Eur Soc Clin Microbiol Infect Dis* 2015;21:323–31. doi:10.1016/j.cmi.2015.02.005.
12. Florio W, Morici P, Ghelardi E, Barnini S, Lupetti A. Recent advances in the microbiological diagnosis of bloodstream infections. *Crit Rev Microbiol* 2017;0:1–20. doi:10.1080/1040841X.2017.1407745.
13. Tille PM. Nucleic Acid–Based Analytic Methods for Microbial Identification and Characterization. *Bailey Scott's Diagnostic Microbiol*. 13th ed., Elsevier Health Sciences; 2013, p. 1056.
14. Paolucci M, Landini MP, Sambri V. Conventional and molecular techniques for the early diagnosis of bacteraemia. *Int J Antimicrob Agents* 2010;36 Suppl 2:S6–16. doi:10.1016/j.ijantimicag.2010.11.010.
15. Deck MK, Anderson ES, Buckner RJ, Colasante G, Coull JM, Crystal B, et al. Multicenter Evaluation of the *Staphylococcus QuickFISH* Method for Simultaneous Identification of *Staphylococcus aureus* and Coagulase-Negative Staphylococci Directly from Blood Culture Bottles in Less than 30 Minutes. *J Clin Microbiol* 2012;50:1994–8. doi:10.1128/JCM.00225-12.
16. Martinez RM, Bauerle ER, Fang FC, Butler-Wu SM. Evaluation of three rapid diagnostic methods for direct identification of microorganisms in positive blood cultures. *J Clin Microbiol* 2014;52:2521–9. doi:10.1128/JCM.00529-14.
17. Oliveira K, Procop GW, Wilson D, Coull J, Stender H. Rapid Identification of *Staphylococcus aureus* Directly from Blood Cultures by Fluorescence *In Situ* Hybridization with Peptide Nucleic Acid Probes. *J Clin Microbiol* 2002;40:247–51. doi:10.1128/JCM.40.1.247-251.2002.

18. Lindholm L, Sarkkinen H. Direct Identification of Gram-Positive Cocci from Routine Blood Cultures by Using AccuProbe Tests. *J Clin Microbiol* 2004;42:5609-13. doi:10.1128/JCM.42.12.5609-5613.2004.
19. Marschal M, Bachmaier J, Autenrieth I, Oberhettinger P, Willmann M, Peter S. Evaluation of the Accelerate Pheno<sup>TM</sup> system for fast identification and antimicrobial susceptibility testing from positive blood culture in Gram-negative bloodstream infection. *J Clin Microbiol* 2017;JCM.00181-17. doi:10.1128/JCM.00181-17.
20. Chantell C. Multiplexed Automated Digital Microscopy for Rapid Identification and Antimicrobial Susceptibility Testing of Bacteria and Yeast Directly from Clinical Samples. *Clin Microbiol Newsl* 2015;37:161-7. doi:10.1016/j.clinmicnews.2015.10.001.
21. Charnot-Katsikas A, Tesic V, Love N, Hill B, Bethel C, Boonlayangoor S, et al. Use of Accelerate Pheno<sup>TM</sup> System for Identification and Antimicrobial Susceptibility Testing (ID/AST) of Pathogens in Positive Blood Cultures and Impact on Time to Results and Workflow. *J Clin Microbiol* 2017. doi:10.1128/JCM.01166-17.
22. Mancini N, Infurnari L, Ghidoli N, Valzano G, Clementi N, Burioni R, et al. Potential Impact of a Microarray-Based Nucleic Acid Assay for Rapid Detection of Gram-Negative Bacteria and Resistance Markers in Positive Blood Cultures. *J Clin Microbiol* 2014;52:1242-5. doi:10.1128/JCM.00142-14.
23. Wojewoda CM, Sercia L, Navas M, Tuohy M, Wilson D, Hall GS, et al. Evaluation of the Verigene Gram-Positive Blood Culture Nucleic Acid Test for Rapid Detection of Bacteria and Resistance Determinants. *J Clin Microbiol* 2013;51:2072-6. doi:10.1128/JCM.00831-13.
24. Qian Q, Tang Y-W, Kolbert CP, Torgerson CA, Hughes JG, Vetter EA, et al. Direct Identification of Bacteria from Positive Blood Cultures by Amplification and Sequencing of the 16S rRNA Gene: Evaluation of BACTEC 9240 Instrument True- Positive and False-Positive Results. *J Clin Microbiol* 2001;39:3578-82. doi:10.1128/JCM.39.10.3578-3582.2001.
25. Gebert S, Siegel D, Wellinghausen N. Rapid detection of pathogens in blood culture bottles by real-time PCR in conjunction with the pre-analytic tool MoLYsis. *J Infect* 2008;57:307-16. doi:10.1016/j.jinf.2008.07.013.
26. Misawa Y, Yoshida A, Saito R, Yoshida H, Okuzumi K, Ito N, et al. Application of loop-mediated isothermal amplification technique to rapid and direct detection of methicillin-resistant *Staphylococcus aureus* (MRSA) in blood cultures. *J Infect Chemother Off J Japan Soc Chemother* 2007;13:134-40. doi:10.1007/s10156-007-0508-9.
27. Notomi T, Okayama H, Masubuchi H, Yonekawa T, Watanabe K, Amino N, et al. Loop-mediated isothermal amplification of DNA. *Nucleic Acids Res* 2000;28:E63.
28. Perera RS, Ding XC, Tully F, Oliver J, Bright N, Bell D, et al. Development and clinical performance of high throughput loop-mediated isothermal amplification for detection of malaria. *PLoS One* 2017;12:e0171126. doi:10.1371/journal.pone.0171126.
29. Altun O, Almuhayawi M, Ullberg M, Özenci V. Clinical Evaluation of the FilmArray Blood Culture Identification Panel in Identification of Bacteria and Yeasts from Positive Blood Culture Bottles. *J Clin Microbiol* 2013;51:4130-6. doi:10.1128/JCM.01835-13.
30. Salimnia H, Fairfax MR, Lephart PR, Schreckenberger P, DesJarlais SM, Johnson JK, et al. Evaluation of the FilmArray Blood Culture Identification Panel: Results of a Multicenter Controlled Trial. *J Clin Microbiol* 2016;54:687-98. doi:10.1128/JCM.01679-15.
31. Pardo J, Klinker KP, Borgert SJ, Butler BM, Giglio PG, Rand KH. Clinical and economic impact of antimicrobial stewardship interventions with the FilmArray blood culture identification panel. *Diagn Microbiol Infect Dis* 2016;84:159-64. doi:10.1016/j.diagmicrobio.2015.10.023.
32. Spencer DH, Sellenriek P, Burnham C-AD. Validation and Implementation of the GeneXpert MRSA/SA Blood Culture Assay in a Pediatric Setting. *Am J Clin Pathol* 2011;136:690-4. doi:10.1309/AJCP07UGYOKBVVNC.

33. Buchan BW, Allen S, Burnham C-AD, TeKippe EM, Davis T, Levi M, et al. Comparison of the Next-Generation Xpert MRSA/SA BC Assay and the GeneOhm StaphSR Assay to Routine Culture for Identification of *Staphylococcus aureus* and Methicillin-Resistant *S. aureus* in Positive-Blood-Culture Broths. *J Clin Microbiol* 2015;53:804–9. doi:10.1128/JCM.03108-14.
34. Dalpke AH, Hofko M, Hamilton F, Mackenzie L, Zimmermann S, Templeton K. Evaluation of the BD Max StaphSR Assay for Rapid Identification of *Staphylococcus aureus* and Methicillin-Resistant *S. aureus* in Positive Blood Culture Broths. *J Clin Microbiol* 2015;53:3630–2. doi:10.1128/JCM.01922-15.
35. Ellem JA, Olma T, O'Sullivan MVN. Rapid Detection of Methicillin-Resistant *Staphylococcus aureus* and Methicillin-Susceptible *S. aureus* Directly from Positive Blood Cultures by Use of the BD Max StaphSR Assay. *J Clin Microbiol* 2015;53:3900–4. doi:10.1128/JCM.02155-15.
36. Lee J, Park YJ, Park DJ, Park KG, Lee HK. Evaluation of BD MAX Staph SR Assay for differentiating between *Staphylococcus aureus* and coagulase-negative Staphylococci and determining methicillin resistance directly from positive blood cultures. *Ann Lab Med* 2017;37:39–44. doi:10.3343/alm.2017.37.1.39.
37. Zboromyrska Y, Vergara A, Cosgaya C, Verger G, Mosqueda N, Almela M, et al. Rapid detection of  $\beta$ -lactamases directly from positive blood cultures using a loop-mediated isothermal amplification (LAMP)-based assay. *Int J Antimicrob Agents* 2015;46:355–6. doi:10.1016/j.ijantimicag.2015.06.002.
38. Rödel J, Bohnert JA, Stoll S, Wassill L, Edel B, Karrasch M, et al. Evaluation of loop-mediated isothermal amplification for the rapid identification of bacteria and resistance determinants in positive blood cultures. *Eur J Clin Microbiol Infect Dis* 2017;36:1033–40. doi:10.1007/s10096-016-2888-1.
39. Galiana A, Coy J, Gimeno A, Guzman NM, Rosales F, Merino E, et al. Evaluation of the Sepsis Flow Chip assay for the diagnosis of blood infections. *PLoS One* 2017;12:1–12. doi:10.1371/journal.pone.0177627.
40. Avda L, Ciencias T. Sepsis Flow Chip Kit. *Máster Diagnóstica SL* 2016:1–24.
41. Martinez RM, Wolk DM. Bloodstream Infections. *Microbiolspectrum* 2016:1–34. doi:10.1128/microbiolspec.DMIH2-0031-2016.
42. Hogan D, Brown B, Shaw C, Thornberg A, Martinez M, Al-khouri A, et al. Characterization of clinical specimens using GenMark's ePlex blood culture identification (BCID) panels. Poster P1018 presented at: 25th European Congress of Clinical Microbiology and Infectious Diseases (ECCMID), Vienna, Austria: 2017.
43. Maubon D, Dard C, Garnaud C, Cornet M. Profile of GenMark's ePlex® blood culture identification fungal pathogen panel. *Expert Rev Mol Diagn* 2017;0:1–14. doi:10.1080/14737159.2018.1420476.
44. Handschur M, Karlic H, Hertel C, Pfeilstöcker M, Haslberger AG. Preanalytic removal of human DNA eliminates false signals in general 16S rDNA PCR monitoring of bacterial pathogens in blood. *Comp Immunol Microbiol Infect Dis* 2009;32:207–19. doi:10.1016/j.cimid.2007.10.005.
45. Al-Soud WA, Rådström P. Purification and characterization of PCR-inhibitory components in blood cells. *J Clin Microbiol* 2001;39:485–93. doi:10.1128/JCM.39.2.485-493.2001.
46. Warhurst G, Dunn G, Chadwick P, Blackwood B, McAuley D, Perkins GD, et al. Rapid detection of health-care-associated bloodstream infection in critical care using Multipathogen real-time polymerase chain reaction technology: A diagnostic accuracy study and systematic review. *Health Technol Assess (Rockv)* 2015;19:1–141. doi:10.3310/hta19350.
47. Reers Y, Idelevich EA, Pätkau H, Sauerland MC, Tafelski S, Nachtigall I, et al. Multiplex PCR assay underreports true bloodstream infections with coagulase-negative staphylococci in

- hematological patients with febrile neutropenia. *Diagn Microbiol Infect Dis* 2016;85:413–5. doi:10.1016/j.diagmicrobio.2016.04.024.
48. Korber F, Zeller I, Grünstäudl M, Willinger B, Apfalter P, Hirschl AM, et al. SeptiFast versus blood culture in clinical routine – A report on 3 years experience. *Wien Klin Wochenschr* 2017;129:427–34. doi:10.1007/s00508-017-1181-3.
  49. Straub J, Paula H, Mayr M, Kasper D, Assadian O, Berger A, et al. Diagnostic accuracy of the ROCHE Septifast PCR system for the rapid detection of blood pathogens in neonatal sepsis— A prospective clinical trial. *PLoS One* 2017;12:e0187688. doi:10.1371/journal.pone.0187688.
  50. Ziegler I, Josefson P, Olcen P, Molling P, Stralin K. Quantitative data from the SeptiFast real-time PCR is associated with disease severity in patients with sepsis. *BMC Infect Dis* 2014;14:155. doi:10.1186/1471-2334-14-155.
  51. Carrara L, Navarro F, Turbau M, Seres M, Morán I, Quintana I, et al. Molecular diagnosis of bloodstream infections with a new dual-priming oligonucleotide-based multiplex PCR assay. *J Med Microbiol* 2013;62:1673–9. doi:10.1099/jmm.0.064758-0.
  52. Loonen AJM, De Jager CPC, Tosserams J, Kusters R, Hilbink M, Wever PC, et al. Biomarkers and molecular analysis to improve bloodstream infection diagnostics in an emergency care unit. *PLoS One* 2014;9. doi:10.1371/journal.pone.0087315.
  53. Ziegler I, Fagerström A, Strålin K, Mölling P. Evaluation of a Commercial Multiplex PCR Assay for Detection of Pathogen DNA in Blood from Patients with Suspected Sepsis. *PLoS One* 2016;11:e0167883. doi:10.1371/journal.pone.0167883.
  54. Wellinghausen N, Kochem AJ, Disqué C, Mühl H, Gebert S, Winter J, et al. Diagnosis of bacteremia in whole-blood samples by use of a commercial universal 16S rRNA gene-based PCR and sequence analysis. *J Clin Microbiol* 2009;47:2759–65. doi:10.1128/JCM.00567-09.
  55. Grif K, Heller I, Prodinger WM, Lechleitner K, Lass-Flörl C, Orth D. Improvement of detection of bacterial pathogens in normally sterile body sites with a focus on orthopedic samples by use of a commercial 16S rRNA broad-range PCR and sequence analysis. *J Clin Microbiol* 2012;50:2250–4. doi:10.1128/JCM.00362-12.
  56. Lebovitz EE, Burbelo PD. Commercial multiplex technologies for the microbiological diagnosis of sepsis. *Mol Diagnosis Ther* 2013;17:221–31. doi:10.1007/s40291-013-0037-4.
  57. Orszag P, Disqué C, Keim S, Lorenz MG, Wiesner O, Hadem J, et al. Monitoring of patients supported by extracorporeal membrane oxygenation for systemic infections by broad-range rRNA gene PCR amplification and sequence analysis. *J Clin Microbiol* 2014;52:307–11. doi:10.1128/JCM.02493-13.
  58. Neely LA, Audeh M, Phung NA, Min M, Suchocki A, Plourde D, et al. T<sub>2</sub> Magnetic Resonance Enables Nanoparticle-Mediated Rapid Detection of Candidemia in Whole Blood. *Sci Transl Med* 2013;5:182ra54-182ra54. doi:10.1126/scitranslmed.3005377.
  59. Zervou FN, Zacharioudakis IM, Kurpewski J, Mylonakis E. T<sub>2</sub> magnetic resonance for fungal diagnosis. *Methods Mol. Biol.*, vol. 1508, 2017, p. 305–19. doi:10.1007/978-1-4939-6515-1\_18.
  60. Mylonakis E, Clancy CJ, Ostrosky-Zeichner L, Garey KW, Alangaden GJ, Vazquez JA, et al. T<sub>2</sub> magnetic resonance assay for the rapid diagnosis of candidemia in whole blood: A clinical trial. *Clin Infect Dis* 2015;60:892–9. doi:10.1093/cid/ciu959.
  61. Mylonakis E, Zacharioudakis IM, Clancy CJ, Nguyen MH, Pappas PG. The Efficacy of T<sub>2</sub> Magnetic Resonance Assay in Monitoring Candidemia after the Initiation of Antifungal Therapy: The Serial Therapeutic and Antifungal Monitoring Protocol (STAMP) trial. *J Clin Microbiol* 2018;JCM.01756-17. doi:10.1128/JCM.01756-17.
  62. Frickmann H, Dekker D, Schwarz NG, Hahn A, Boahen K, Sarpong N, et al. 16S rRNA Gene Sequence-Based Identification of Bacteria in Automatically Incubated Blood Culture Materials from Tropical Sub-Saharan Africa. *PLoS One* 2015;10. doi:10.1371/journal.pone.0135923.

63. Faria M, Conly J, Surette M. The development and application of a molecular community profiling strategy to identify polymicrobial bacterial DNA in the whole blood of septic patients. *BMC Microbiol* 2015;15:215. doi:10.1186/s12866-015-0557-7.
64. Gyarmati P, Kjellander C, Aust C, Kalin M, Öhrmalm L, Giske CG. Bacterial Landscape of Bloodstream Infections in Neutropenic Patients via High Throughput Sequencing. *PLoS One* 2015;10. doi:10.1371/journal.pone.0135756.
65. Decuyper S, Meehan CJ, Van Puyvelde S, De Block T, Maltha J, Palpouguini L, et al. Diagnosis of Bacterial Bloodstream Infections: A 16S Metagenomics Approach. *PLoS Negl Trop Dis* 2016;10:1–12. doi:10.1371/journal.pntd.0004470.
66. Sabat AJ, van Zanten E, Akkerboom V, Wisselink G, van Slochteren K, de Boer RF, et al. Targeted next-generation sequencing of the 16S-23S rRNA region for culture-independent bacterial identification - increased discrimination of closely related species. *Sci Rep* 2017;7:3434. doi:10.1038/s41598-017-03458-6.
67. Gyarmati P, Kjellander C, Aust C, Song Y, Öhrmalm L, Giske CG. Metagenomic analysis of bloodstream infections in patients with acute leukemia and therapy-induced neutropenia. *Sci Rep* 2016;6. doi:10.1038/srep23532.
68. Couto N, Schuele L, Machado M, Raangs EC, Machado M, Mendes CI et al. Clinical shotgun metagenomics for the concomitant identification and typing of microorganisms. Submitted Manuscript 2018.
69. Parize P, Pilms B, Lanternier F, Lortholary O, Lecuit M, Muth E, et al. Untargeted next-generation sequencing-based first-line diagnosis of infection in immunocompromised adults: a multicentre, blinded, prospective study. *Clin Microbiol Infect* 2017;23:574.e1-574.e6. doi:10.1016/j.cmi.2017.02.006.
70. Lizarazo EF, Couto N, Raangs E, Vincenti-Gonzalez MF, Tovar R, Bethencourt S et al. Shotgun metagenomics as a tool for the rapid diagnosis and genotyping of Dengue. *Trop Med Int Heal* 2017;22:125.
71. Rossen JWA, Friedrich AW, Moran-Gilad J. Practical issues in implementing whole-genome-sequencing in routine diagnostic microbiology. *Clin Microbiol Infect* 2017. doi:10.1016/j.cmi.2017.11.001.
72. Stamper PD, Cai M, Howard T, Speser S, Carroll KC. Clinical Validation of the Molecular BD GeneOhm StaphSR Assay for Direct Detection of *Staphylococcus aureus* and Methicillin-Resistant *Staphylococcus aureus* in Positive Blood Cultures. *J Clin Microbiol* 2007;45:2191–6. doi:10.1128/JCM.00552-07.
73. Tissari P, Zumla A, Tarkka E, Mero S, Savolainen L, Vaara M, et al. Accurate and rapid identification of bacterial species from positive blood cultures with a DNA-based microarray platform: an observational study. *Lancet* 2010;375:224–30. doi:10.1016/S0140-6736(09)61569-5.
74. Aittakorpi A, Kuusela P, Koukila-Kähkölä P, Vaara M, Petrou M, Gant V, et al. Accurate and Rapid Identification of *Candida* spp. Frequently Associated with Fungemia by Using PCR and the Microarray-Based Prove-it Sepsis Assay. *J Clin Microbiol* 2012;50:3635–40. doi:10.1128/JCM.01461-12.
75. Jordana-Lluch E, Giménez M, Dolores Quesada M, Rivaya B, Marcó C, Jesus Domínguez M, et al. Evaluation of the Broad-Range PCR/ESI-MS Technology in Blood Specimens for the Molecular Diagnosis of Bloodstream Infections. *PLoS One* 2015;10:1–13. doi:10.1371/journal.pone.0140865.
76. Stevenson M, Pandor A, James MMS, Rafia R, Uttley L, Stevens J, et al. Sepsis: the LightCycler SeptiFast Test MGRADE®, SepsisTest™ and IRIDICA BAC BSI assay for rapidly identifying bloodstream bacteria and fungi – a systematic review and economic evaluation. *Health Technol Assess (Rockv)* 2016;20. doi:10.3310/hta20460.

77. Özenci V, Patel R, Ullberg M, Strålin K. Demise of Polymerase Chain Reaction/Electrospray Ionization-Mass Spectrometry as an Infectious Diseases Diagnostic Tool. *Clin Infect Dis* 2018;66:452–5. doi:10.1093/cid/cix743.
78. Banerjee R, Özenci V, Patel R. Individualized Approaches Are Needed for Optimized Blood Cultures. *Clin Infect Dis* 2016;63:1332–9. doi:10.1093/cid/ciw573.
79. Timbrook TT, Morton JB, McConeghy KW, Caffrey AR, Mylonakis E, LaPlante KL. The Effect of Molecular Rapid Diagnostic Testing on Clinical Outcomes in Bloodstream Infections: A Systematic Review and Meta-analysis. *Clin Infect Dis* 2017;64:15–23. doi:10.1093/cid/ciw649.

## **Appendix**

### ***Supplementary Information***

Search terms on PubMed: (bloodstream\*[tiab] OR blood circul\* [tiab]) AND ("Sepsis"[Mesh] OR sep-sis [tiab] OR infection\*[tiab] OR bBSI\* [tiab]) AND ("Molecular Diagnostic Techniques"[Mesh] OR mo-lecular diagnos\*[tiab] OR rapid molecular identification\* [tiab] OR molecular detect\* [tiab])

More specific information about each of the “tests” used was subsequently searched through Pub-Med and through the companies’ websites.





**Part II.**  
Whole Genome  
Sequencing

