

## University of Groningen

### The influence of the sample matrix on LC-MS/MS method development and analytical performance

Koster, Remco Arjan

**IMPORTANT NOTE: You are advised to consult the publisher's version (publisher's PDF) if you wish to cite from it. Please check the document version below.**

*Document Version*

Publisher's PDF, also known as Version of record

*Publication date:*

2015

[Link to publication in University of Groningen/UMCG research database](#)

*Citation for published version (APA):*

Koster, R. A. (2015). *The influence of the sample matrix on LC-MS/MS method development and analytical performance*. [Thesis fully internal (DIV), University of Groningen]. University of Groningen.

**Copyright**

Other than for strictly personal use, it is not permitted to download or to forward/distribute the text or part of it without the consent of the author(s) and/or copyright holder(s), unless the work is under an open content license (like Creative Commons).

The publication may also be distributed here under the terms of Article 25fa of the Dutch Copyright Act, indicated by the "Taverne" license. More information can be found on the University of Groningen website: <https://www.rug.nl/library/open-access/self-archiving-pure/taverne-amendment>.

**Take-down policy**

If you believe that this document breaches copyright please contact us providing details, and we will remove access to the work immediately and investigate your claim.

*Downloaded from the University of Groningen/UMCG research database (Pure): <http://www.rug.nl/research/portal>. For technical reasons the number of authors shown on this cover page is limited to 10 maximum.*

# The influence of the sample matrix on LC-MS/MS method development and analytical performance

Remco A. Koster

Publication of this thesis was financially supported by Rijks Universiteit Groningen (RUG), the Stichting ter bevordering van Onderzoek in de Ziekenhuisfarmacie te Groningen (Stichting O.Z.G.), Groningen University Institute for Drug Exploration (GUIDE), Agilent Technologies, Thermo Scientific, DBSL, Spark Holland, Astellas, Pfizer, Dutch Kidney Foundation, KNCV Tuberculosis Foundation, and Stichting Beatrixoord Noord Nederland.

<b>Cover</b>	Thermo Scientific
<b>Layout</b>	Renate Siebes, Proefschrift.nu
<b>Printed by</b>	Ridderprint, Ridderkerk
<b>ISBN</b>	978-90-367-8022-3

© Remco A. Koster, 2015

Copyright of the published articles is with the corresponding journal or otherwise with the author. No part of this publication may be reproduced, stored in a retrieval system, or transmitted, in any form or by any means, without the prior permission in writing from the author or the copyright-owning journal.



rijksuniversiteit  
groningen

# **The influence of the sample matrix on LC-MS/MS method development and analytical performance**

## **Proefschrift**

ter verkrijging van de graad van doctor aan de  
Rijksuniversiteit Groningen  
op gezag van de  
rector magnificus prof. dr. E. Sterken  
en volgens besluit van het College voor Promoties.

De openbare verdediging zal plaatsvinden op

vrijdag 25 september 2015 om 16:15 uur

door

**Remco Arjan Koster**

geboren op 9 juli 1979  
te Utrecht

**Promotores**

Prof. dr. D.R.A. Uges

Prof. dr. J.G.W. Kosterink

**Copromotor**

Dr. J.W.C. Alffenaar

**Beoordelingscommissie**

Prof. dr. I.P. Kema

Prof. dr. N.C. van de Merbel

Prof. dr. G.W. Somsen

# Contents

<b>Chapter 1</b>	General introduction, scope and outline of the thesis	<b>7</b>
<b>Chapter 2</b>	Robust, high-throughput LC-MS/MS method for Therapeutic Drug Monitoring of cyclosporin A, tacrolimus, everolimus and sirolimus in whole blood	<b>15</b>
<b>Chapter 3.1</b>	What is the right blood hematocrit preparation procedure for standards and quality control samples for dried blood spot analysis?	<b>37</b>
<b>Chapter 3.2</b>	Determination of moxifloxacin in dried blood spot using LC-MS/MS and the impact of hematocrit and blood volume	<b>51</b>
<b>Chapter 3.3</b>	Simultaneous determination of rifampicin, clarithromycin and their metabolites in dried blood spots using LC-MS/MS	<b>73</b>
<b>Chapter 3.4</b>	Fast LC-MS/MS analysis of tacrolimus, sirolimus, everolimus and cyclosporin A in dried blood spots and the influence of the hematocrit and immunosuppressant concentration on recovery	<b>97</b>
<b>Chapter 3.5</b>	Dried blood spot analysis of creatinine with LC-MS/MS in addition to immunosuppressants analysis	<b>117</b>
<b>Chapter 3.6</b>	The relation of the number of hydrogen bond acceptors with recoveries of immunosuppressants in dried blood spot analysis	<b>139</b>
<b>Chapter 3.7</b>	The influence of the dried blood spot drying time on the recoveries of six immunosuppressants	<b>153</b>
<b>Chapter 3.8</b>	The performance of five different dried blood spot cards for the analysis of six immunosuppressants	<b>165</b>
<b>Chapter 4</b>	Fast analysis of remifentanyl with the use of LC-MS/MS and an extensive stability investigation in whole blood and acidified EDTA plasma	<b>183</b>
<b>Chapter 5.1</b>	Fast and highly selective LC-MS/MS screening for THC and 16 other abused drugs and metabolites in human hair to monitor patients for drug abuse	<b>201</b>
<b>Chapter 5.2</b>	Application of sweat patch screening for 16 drugs and metabolites using a fast and highly selective LC-MS/MS method	<b>229</b>
<b>Chapter 6</b>	General discussion and future perspectives	<b>255</b>
<b>Chapter 7</b>	Summary	<b>269</b>
	Samenvatting	<b>277</b>
	Publication list	<b>285</b>
	Dankwoord	<b>289</b>
	About the author	<b>295</b>

